

# eCVP

## Electronic Comprehensive Validation Package



## Air Toxics Ltd.

180 Blue Ravine Road Ste. B  
Folsom, CA 95630  
Phone: 916/985-1000  
Fax: 916/985-1020  
eMail: [atl@airtoxics.com](mailto:atl@airtoxics.com)  
[www.airtoxics.com](http://www.airtoxics.com)



AN ENVIRONMENTAL ANALYTICAL LABORATORY

### COMPREHENSIVE VALIDATION PACKAGE

Modified TO-15

### INVENTORY SHEET

Work Order #: 0704143

	Page Nos.	
	From	To
1. Work Order Cover Page & Laboratory Narrative	1	4
a. <u>Lumen Validation Report</u>	--	--
2. Sample Results and Raw Data (Organized by Sample)	5	25
a. ATL Sample Results Form		
b. Target Compound Raw Data		
-Internal Standard Area and Retention Time Summary		
-Surrogate Recovery Summary (If Applicable)		
-Chromatogram(s) and Ion Profiles (If Applicable)		
3. QC Results and Raw Data		
a. Method Blank (Results+ Raw Data)	26	32
b. Surrogate Recover Summary Form (If Applicable)	33	33
c. Internal Standard Summary Form (If Applicable)	34	34
d. Duplicate Results Summary Sheet	--	--
e. Matrix Spike/Matrix Spike Duplicate (Results + Raw Data)	--	--
f. Initial Calibration Data (Summary Sheet + Raw Data)	35	146
g. MDL Study (If Applicable)	--	--
h. Continuing Calibration Verification Data (Summary Sheet	147	160
i. Second Source LCS(Summary + Raw Data)	161	174
j. Extraction Logs	--	--
k. Instrument Run Logs/Software Verification	175	175
l. GC/MS Tune (Results + Raw Data)	176	190
4. Shipping/Receiving Documents		
a. Login Receipt Summary Sheet	191	192
b. Chain-of-Custody Records	193	193
c. Sample Log-In Sheet	194	194
d. Misc Shipping/Receiving Records (list of individual records)		
<u>Sample Receipt Discrepancy Report</u>	195	195
5. Other Records (describe or list)		
a. <u>Manual Spectral Defense</u>	196	197
b. <u>Manual Integrations</u>	--	--
c. <u>Manual Calculations</u>	--	--
d. <u>Canister Dilution Factors</u>	198	199
e. <u>Laboratory Corrective Action Request</u>	--	--
f. <u>CAS Number Reference</u>	200	201
g. <u>Variance Table</u>	--	--
h. <u>Canister Certification</u>	--	--
i. <u>Data Review Check Sheet</u>	202	202

Comments:

---

Completed by:

*Judy Lee*

Judy Lee / Document Control

4/24/07

(Signature)

( Print Name & Title)

(Date)




AN ENVIRONMENTAL ANALYTICAL LABORATORY

**WORK ORDER #: 0704143**

Work Order Summary

<b>CLIENT:</b>	Ms. Sarah Aldridge GEI Consultants, Inc. 455 Winding Brook Dr. Suite 201 Glastonbury, CT 06033	<b>BILL TO:</b>	Ms. Sarah Aldridge GEI Consultants, Inc. 455 Winding Brook Dr. Suite 201 Glastonbury, CT 06033
<b>PHONE:</b>	860-368-5300	<b>P.O. #</b>	
<b>FAX:</b>	860-368-5307	<b>PROJECT #</b>	061140-8-1703 Bayshore OUI Southern
<b>DATE RECEIVED:</b>	04/06/2007	<b>CONTACT:</b>	cell Air Monitor Kelly Buettner
<b>DATE COMPLETED:</b>	04/19/2007		

<u>FRACTION #</u>	<u>NAME</u>	<u>TEST</u>	<u>RECEIPT VAC./PRES.</u>
01A	000001016 Downwind	Modified TO-15	5.0 "Hg
02A	12007 Upwind	Modified TO-15	5.0 "Hg
03A	Lab Blank	Modified TO-15	NA
04A	CCV	Modified TO-15	NA
05A	LCS	Modified TO-15	NA

CERTIFIED BY:  DATE: 04/19/07

Laboratory Director

Certification numbers: CA NELAP - 02110CA, LA NELAP/LELAP- AI 30763, NJ NELAP - CA004  
 NY NELAP - 11291, UT NELAP - 9166389892  
 Name of Accrediting Agency: NELAP/Florida Department of Health, Scope of Application: Clean Air Act,  
 Accreditation number: E87680, Effective date: 07/01/06, Expiration date: 06/30/07  
 Air Toxics Ltd. certifies that the test results contained in this report meet all requirements of the NELAC standards  
 This report shall not be reproduced, except in full, without the written approval of Air Toxics Ltd.

180 BLUE RAVINE ROAD, SUITE B FOLSOM, CA - 95630  
 (916) 985-1000 . (800) 985-5955 . FAX (916) 985-1020

**LABORATORY NARRATIVE**  
**Modified TO-15**  
**GEI Consultants, Inc.**  
**Workorder# 0704143**

Two 6 Liter Summa Canister samples were received on April 06, 2007. The laboratory performed analysis via modified EPA Method TO-15 using GC/MS in the full scan mode. The method involves concentrating up to 0.2 liters of air. The concentrated aliquot is then flash vaporized and swept through a water management system to remove water vapor. Following dehumidification, the sample passes directly into the GC/MS for analysis.

Method modifications taken to run these samples are summarized in the below table. Specific project requirements may over-ride the ATL modifications.

<i>Requirement</i>	<i>TO-15</i>	<i>ATL Modifications</i>
Daily CCV	+/- 30% Difference	<= 30% Difference with two allowed out up to <=40%.; flag and narrate outliers
Sample collection media	Summa canister	ATL recommends use of summa canisters to insure data defensibility, but will report results from Tedlar bags at client request
Method Detection Limit	Follow 40CFR Pt.136 App. B	The MDL met all relevant requirements in Method TO-15 (statistical MDL less than the LOQ). The concentration of the spiked replicate may have exceeded 10X the calculated MDL in some cases

**Receiving Notes**

Sample identifications for samples 000001016 Downwind and 12007 Upwind were not provided on the sample tags. The discrepancy was noted in the Sample Receipt Confirmation email/fax and the information on the Chain of Custody was used to process and report the samples.

**Analytical Notes**

There were no analytical discrepancies.

**Definition of Data Qualifying Flags**

Eight qualifiers may have been used on the data analysis sheets and indicates as follows:

- B - Compound present in laboratory blank greater than reporting limit (background subtraction no performed).
- J - Estimated value.
- E - Exceeds instrument calibration range.
- S - Saturated peak.
- Q - Exceeds quality control limits.
- U - Compound analyzed for but not detected above the reporting limit.
- UJ- Non-detected compound associated with low bias in the CCV
- N - The identification is based on presumptive evidence.

File extensions may have been used on the data analysis sheets and indicates

as follows:

a-File was requantified

b-File was quantified by a second column and detector

r1-File was requantified for the purpose of reissue

**Table 1**

Client Sample ID	Lab Sample ID	Date Collected	Date Received	Date Extracted	Sample	Sample Extract		Sample Condition
					Holding Time (Days)	Date Analyzed	Holding Time (Days)	
000001016 Downwind	0704143-01A	4/ 5/2007	4/ 6/2007	NA	7	4/12/2007	NA	Good
12007 Upwind	0704143-02A	4/ 5/2007	4/ 6/2007	NA	7	4/12/2007	NA	Good
Lab Blank	0704143-03A	NA	NA	NA	NA	4/12/2007	NA	Good
CCV	0704143-04A	NA	NA	NA	NA	4/12/2007	NA	Good
LCS	0704143-05A	NA	NA	NA	NA	4/12/2007	NA	Good

## **Sample Results and Raw Data**



AN ENVIRONMENTAL ANALYTICAL LABORATORY

---

## Summary of Detected Compounds MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

Client Sample ID: 000001016 Downwind

Lab ID#: 0704143-01A

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (uG/m3)	Amount (uG/m3)
1,2,4-Trichlorobenzene	3.2	3.7	24	27
Acetone	3.2	3.6	7.6	8.7
Carbon Disulfide	0.80	11	2.5	36
Tetrahydrofuran	0.80	0.87	2.4	2.6





AN ENVIRONMENTAL ANALYTICAL LABORATORY

Client Sample ID: 00001016 Downwind

Lab ID#: 0704143-01A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	8041206	Date of Collection:	4/5/07
Dil. Factor:	1.61	Date of Analysis:	4/12/07 04:40 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (uG/m3)	Amount (uG/m3)
Freon 12	0.80	Not Detected	4.0	Not Detected
Freon 114	0.80	Not Detected	5.6	Not Detected
Vinyl Chloride	0.80	Not Detected	2.0	Not Detected
Bromomethane	0.80	Not Detected	3.1	Not Detected
Chloroethane	0.80	Not Detected	2.1	Not Detected
Freon 11	0.80	Not Detected	4.5	Not Detected
1,1-Dichloroethene	0.80	Not Detected	3.2	Not Detected
Freon 113	0.80	Not Detected	6.2	Not Detected
Methylene Chloride	0.80	Not Detected	2.8	Not Detected
1,1-Dichloroethane	0.80	Not Detected	3.2	Not Detected
cis-1,2-Dichloroethene	0.80	Not Detected	3.2	Not Detected
Chloroform	0.80	Not Detected	3.9	Not Detected
1,1,1-Trichloroethane	0.80	Not Detected	4.4	Not Detected
Carbon Tetrachloride	0.80	Not Detected	5.1	Not Detected
Benzene	0.80	Not Detected	2.6	Not Detected
1,2-Dichloroethane	0.80	Not Detected	3.2	Not Detected
Trichloroethene	0.80	Not Detected	4.3	Not Detected
1,2-Dichloropropane	0.80	Not Detected	3.7	Not Detected
cis-1,3-Dichloropropene	0.80	Not Detected	3.6	Not Detected
Toluene	0.80	Not Detected	3.0	Not Detected
trans-1,3-Dichloropropene	0.80	Not Detected	3.6	Not Detected
1,1,2-Trichloroethane	0.80	Not Detected	4.4	Not Detected
Tetrachloroethene	0.80	Not Detected	5.5	Not Detected
1,2-Dibromoethane (EDB)	0.80	Not Detected	6.2	Not Detected
Chlorobenzene	0.80	Not Detected	3.7	Not Detected
Ethyl Benzene	0.80	Not Detected	3.5	Not Detected
m,p-Xylene	0.80	Not Detected	3.5	Not Detected
o-Xylene	0.80	Not Detected	3.5	Not Detected
Styrene	0.80	Not Detected	3.4	Not Detected
1,1,2,2-Tetrachloroethane	0.80	Not Detected	5.5	Not Detected
1,3,5-Trimethylbenzene	0.80	Not Detected	4.0	Not Detected
1,2,4-Trimethylbenzene	0.80	Not Detected	4.0	Not Detected
1,3-Dichlorobenzene	0.80	Not Detected	4.8	Not Detected
1,4-Dichlorobenzene	0.80	Not Detected	4.8	Not Detected
alpha-Chlorotoluene	0.80	Not Detected	4.2	Not Detected
1,2-Dichlorobenzene	0.80	Not Detected	4.8	Not Detected
1,3-Butadiene	0.80	Not Detected	1.8	Not Detected
Hexane	0.80	Not Detected	2.8	Not Detected
Cyclohexane	0.80	Not Detected	2.8	Not Detected



AN ENVIRONMENTAL ANALYTICAL LABORATORY

Client Sample ID: 00001016 Downwind

Lab ID#: 0704143-01A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	8041206	Date of Collection:	4/5/07
Dil. Factor:	1.61	Date of Analysis:	4/12/07 04:40 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (uG/m3)	Amount (uG/m3)
Heptane	0.80	Not Detected	3.3	Not Detected
Bromodichloromethane	0.80	Not Detected	5.4	Not Detected
Dibromochloromethane	0.80	Not Detected	6.8	Not Detected
Cumene	0.80	Not Detected	4.0	Not Detected
Propylbenzene	0.80	Not Detected	4.0	Not Detected
Chloromethane	3.2	Not Detected	6.6	Not Detected
1,2,4-Trichlorobenzene	3.2	3.7	24	27
Hexachlorobutadiene	3.2	Not Detected	34	Not Detected
Acetone	3.2	3.6	7.6	8.7
Carbon Disulfide	0.80	11	2.5	36
2-Propanol	3.2	Not Detected	7.9	Not Detected
trans-1,2-Dichloroethene	0.80	Not Detected	3.2	Not Detected
2-Butanone (Methyl Ethyl Ketone)	0.80	Not Detected	2.4	Not Detected
Tetrahydrofuran	0.80	0.87	2.4	2.6
1,4-Dioxane	3.2	Not Detected	12	Not Detected
4-Methyl-2-pentanone	0.80	Not Detected	3.3	Not Detected
2-Hexanone	3.2	Not Detected	13	Not Detected
Bromoform	0.80	Not Detected	8.3	Not Detected
4-Ethyltoluene	0.80	Not Detected	4.0	Not Detected
Ethanol	3.2	Not Detected	6.1	Not Detected
Methyl tert-butyl ether	0.80	Not Detected	2.9	Not Detected
3-Chloropropene	3.2	Not Detected	10	Not Detected
2,2,4-Trimethylpentane	0.80	Not Detected	3.8	Not Detected
Naphthalene	3.2	Not Detected	17	Not Detected

Container Type: 6 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	94	70-130
1,2-Dichloroethane-d4	97	70-130
4-Bromofluorobenzene	93	70-130

Report Date: 19-Apr-2007 15:59

## Air Toxics Ltd.

## AMBIENT AIR METHOD TO14A/TO15

Data file : /chem/msd8.i/8-12apr.b/8041206.d  
 Lab Smp Id: 0704143-01A  
 Inj Date : 12-APR-2007 16:40  
 Operator : JG Inst ID: msd8.i  
 Smp Info : 200mL #24231  
 Misc Info : 5.0"Hg-5.0psi  
 Comment :  
 Method : /chem/msd8.i/8-12apr.b/t14q322b.m  
 Meth Date : 12-Apr-2007 23:51 jgray Quant Type: ISTD  
 Cal Date : 26-MAR-2007 13:11 Cal File: 8032608.d  
 Als bottle: 1  
 Dil Factor: 1.61000  
 Integrator: HP RTE Compound Sublist: AT04+Na.sub  
 Target Version: 3.50 Sample Matrix: AIR  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* CpndVariable

Cpnd Variable Local Compound Variable

CONCENTRATIONS								
RT	EXP RT	(REL RT)	MASS	RESPONSE		TARGET RANGE	RATIO	
				( PPBV)	( PPBV)			
==	=====	=====	====	=====	=====	=====	=====	=====
-----								
* 67	Bromochloromethane					CAS #: 74-97-5		
8.395	8.395	(1.000)	130	209705	25.0000	80.00- 120.00	100.00	
8.395	8.395	(1.000)	128	164733		45.68- 105.68	78.55	
8.395	8.395	(1.000)	49	631003		268.48- 328.48	300.90	
-----								
* 86	1,4-Difluorobenzene					CAS #: 540-36-3		
10.248	10.275	(1.000)	114	923293	25.0000	80.00- 120.00	100.00	
10.248	10.275	(1.000)	88	162403		0.00- 48.01	17.59	
-----								
* 123	Chlorobenzene-d5					CAS #: 3114-55-4		
15.225	15.225	(1.000)	117	673199	25.0000	80.00- 120.00	100.00	
15.197	15.225	(1.000)	82	465180		35.49- 95.49	69.10	
-----								
\$ 80	1,2-Dichloroethane-d4					CAS #: 17060-07-0		
9.473	9.474	(1.128)	65	441498	24.3561	24.356 80.00- 120.00	100.00	
9.473	9.474	(1.128)	67	210129		27.92- 87.92	47.59	
-----								
\$ 102	Toluene-d8					CAS #: 2037-26-5		
12.985	12.985	(1.267)	98	858358	23.6401	23.640 80.00- 120.00	100.00	
12.985	12.985	(1.267)	70	106350		0.00- 42.61	12.39	

CONCENTRATIONS

ON-COL FINAL

RT EXP RT (REL RT) MASS RESPONSE ( PPEV) ( PPBV) TARGET RANGE RATIO  
== =====

\$ 102 Toluene-d8 (continued)

12.985 12.985 (1.267) 100 570095 40.27- 100.27 66.42

\$ 138 Bromofluorobenzene

CAS #: 460-00-4

16.773 16.773 (1.102) 174 321227 23.2838 23.284 80.00- 120.00 100.00

16.745 16.773 (1.100) 95 565359 138.09- 198.09 176.00

16.773 16.773 (1.102) 176 329898 69.53- 129.53 102.70

30 Acetone

CAS #: 67-64-1

4.994 4.967 (0.595) 58 27449 2.26654 3.649 80.00- 120.00 100.00

4.994 4.967 (0.595) 43 97419 337.35- 397.35 354.90

33 Carbon Disulfide

CAS #: 75-15-0

5.160 5.160 (0.615) 76 372461 7.14098 11.497 80.00- 120.00 100.00

66 Tetrahydrofuran

CAS #: 109-99-9

8.367 8.368 (0.997) 42 19776 0.54086 0.8708 80.00- 120.00 100.00

8.367 8.368 (0.997) 71 5059 0.00- 55.51 25.59

8.367 8.368 (0.997) 72 5584 0.00- 56.52 28.24

163 1,2,4-Trichlorobenzene

CAS #: 120-82-1

19.400 19.593 (1.274) 180 38640 2.30043 3.704 80.00- 120.00 100.00

19.289 19.593 (1.267) 182 36590 65.68- 125.68 94.69

Report Date: 19-Apr-2007 15:59

Air Toxics Ltd.

INTERNAL STANDARD COMPOUNDS  
AREA AND RT SUMMARY

Instrument ID: msd8.i  
 Lab File ID: 8041206.d  
 Lab Smp Id: 0704143-01A  
 Analysis Type: VOA  
 Quant Type: ISTD  
 Operator: JG  
 Method File: /chem/msd8.i/8-12apr.b/t14q322b.m  
 Misc Info: 5.0"Hg-5.0psi

Calibration Date: 12-APR-2007  
 Calibration Time: 09:53  
 Level: LOW  
 Sample Type: AIR

COMPOUND	STANDARD	AREA LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	269481	161689	377273	209705	-22.18
86 1,4-Difluorobenze	1204367	722620	1686114	923293	-23.34
123 Chlorobenzene-d5	906226	543736	1268716	673199	-25.71

COMPOUND	STANDARD	RT LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	8.40	8.07	8.73	8.40	0.00
86 1,4-Difluorobenze	10.28	9.95	10.61	10.25	-0.27
123 Chlorobenzene-d5	15.22	14.89	15.55	15.22	0.00

AREA UPPER LIMIT = + 40% of internal standard area.

AREA LOWER LIMIT = - 40% of internal standard area.

RT UPPER LIMIT = + 0.33 minutes of internal standard RT.

RT LOWER LIMIT = - 0.33 minutes of internal standard RT.

Air Toxics Ltd.

RECOVERY REPORT

Client Name: Client SDG: 8-12apr  
Sample Matrix: GAS Fraction: VOA  
Lab Smp Id: 0704143-01A  
Level: LOW Operator: JG  
Data Type: MS DATA SampleType: SAMPLE  
SpikeList File: AT04+NA-2.spk Quant Type: ISTD  
Sublist File: AT04+Na.sub  
Method File: /chem/msd8.i/8-12apr.b/t14q322b.m  
Misc Info: 5.0"Hg-5.0psi

SURROGATE COMPOUND	CONC ADDED PPBV	CONC RECOVERED PPBV	% RECOVERED	LIMITS
\$ 80 1,2-Dichloroethane	25.000	24.356	97.42	70-130
\$ 102 Toluene-d8	25.000	23.640	94.56	70-130
\$ 138 Bromofluorobenzene	25.000	23.284	93.14	70-130

Data File: /chem/msd8.1/8-12apr.b/8041206.d

Date: 12-APR-2007 16:40

Client ID:


Sample Info: 200ML #24231

Column phase: RTX-624

Instrument: msd8.1

Operator: JG

Column diameter: 0.53



Date : 12-APR-2007 16:40

Client ID:

Instrument: msd8.i

Sample Info: 200mL #24231







Operator: JG

Column phase: RTX-624

Column diameter: 0.53

30 Acetone

Concentration: 3,649 PPBV





Date : 12-APR-2007 16:40

Client ID:

Instrument: msd8.i

Sample Info: 200mL #24231






Operator: JG

Column phase: RTX-624

Column diameter: 0.53

33 Carbon Disulfide

Concentration: 11.497 PPBV



Date : 12-APR-2007 16:40

Client ID:

Instrument: msd8.i

Sample Info: 200mL #24231








Operator: JG

Column phase: RTX-624

Column diameter: 0.53

66 Tetrahydrofuran

Concentration: 0.8708 PPBV



Date : 12-APR-2007 16:40

Client ID:

Instrument: msd8,i

Sample Info: 200mL #24231







Operator: JG

Column phase: RTX-624

Column diameter: 0.53

163 1,2,4-Trichlorobenzene

Concentration: 3,704 PPBV





AN ENVIRONMENTAL ANALYTICAL LABORATORY

---

**Summary of Detected Compounds**  
**MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN**

**Client Sample ID: 12007 Upwind**

**Lab ID#: 0704143-02A**

No Detections Were Found.



AN ENVIRONMENTAL ANALYTICAL LABORATORY

Client Sample ID: 12007 Upwind

Lab ID#: 0704143-02A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	8041207	Date of Collection:	4/5/07
Dil. Factor:	1.61	Date of Analysis:	4/12/07 05:23 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (uG/m3)	Amount (uG/m3)
Freon 12	0.80	Not Detected	4.0	Not Detected
Freon 114	0.80	Not Detected	5.6	Not Detected
Vinyl Chloride	0.80	Not Detected	2.0	Not Detected
Bromomethane	0.80	Not Detected	3.1	Not Detected
Chloroethane	0.80	Not Detected	2.1	Not Detected
Freon 11	0.80	Not Detected	4.5	Not Detected
1,1-Dichloroethene	0.80	Not Detected	3.2	Not Detected
Freon 113	0.80	Not Detected	6.2	Not Detected
Methylene Chloride	0.80	Not Detected	2.8	Not Detected
1,1-Dichloroethane	0.80	Not Detected	3.2	Not Detected
cis-1,2-Dichloroethene	0.80	Not Detected	3.2	Not Detected
Chloroform	0.80	Not Detected	3.9	Not Detected
1,1,1-Trichloroethane	0.80	Not Detected	4.4	Not Detected
Carbon Tetrachloride	0.80	Not Detected	5.1	Not Detected
Benzene	0.80	Not Detected	2.6	Not Detected
1,2-Dichloroethane	0.80	Not Detected	3.2	Not Detected
Trichloroethene	0.80	Not Detected	4.3	Not Detected
1,2-Dichloropropane	0.80	Not Detected	3.7	Not Detected
cis-1,3-Dichloropropene	0.80	Not Detected	3.6	Not Detected
Toluene	0.80	Not Detected	3.0	Not Detected
trans-1,3-Dichloropropene	0.80	Not Detected	3.6	Not Detected
1,1,2-Trichloroethane	0.80	Not Detected	4.4	Not Detected
Tetrachloroethene	0.80	Not Detected	5.5	Not Detected
1,2-Dibromoethane (EDB)	0.80	Not Detected	6.2	Not Detected
Chlorobenzene	0.80	Not Detected	3.7	Not Detected
Ethyl Benzene	0.80	Not Detected	3.5	Not Detected
m,p-Xylene	0.80	Not Detected	3.5	Not Detected
o-Xylene	0.80	Not Detected	3.5	Not Detected
Styrene	0.80	Not Detected	3.4	Not Detected
1,1,2,2-Tetrachloroethane	0.80	Not Detected	5.5	Not Detected
1,3,5-Trimethylbenzene	0.80	Not Detected	4.0	Not Detected
1,2,4-Trimethylbenzene	0.80	Not Detected	4.0	Not Detected
1,3-Dichlorobenzene	0.80	Not Detected	4.8	Not Detected
1,4-Dichlorobenzene	0.80	Not Detected	4.8	Not Detected
alpha-Chlorotoluene	0.80	Not Detected	4.2	Not Detected
1,2-Dichlorobenzene	0.80	Not Detected	4.8	Not Detected
1,3-Butadiene	0.80	Not Detected	1.8	Not Detected
Hexane	0.80	Not Detected	2.8	Not Detected
Cyclohexane	0.80	Not Detected	2.8	Not Detected



AN ENVIRONMENTAL ANALYTICAL LABORATORY

Client Sample ID: 12007 Upwind

Lab ID#: 0704143-02A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	8041207	Date of Collection:	4/5/07
Dil. Factor:	1.61	Date of Analysis:	4/12/07 05:23 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (uG/m3)	Amount (uG/m3)
Heptane	0.80	Not Detected	3.3	Not Detected
Bromodichloromethane	0.80	Not Detected	5.4	Not Detected
Dibromochloromethane	0.80	Not Detected	6.8	Not Detected
Cumene	0.80	Not Detected	4.0	Not Detected
Propylbenzene	0.80	Not Detected	4.0	Not Detected
Chloromethane	3.2	Not Detected	6.6	Not Detected
1,2,4-Trichlorobenzene	3.2	Not Detected	24	Not Detected
Hexachlorobutadiene	3.2	Not Detected	34	Not Detected
Acetone	3.2	Not Detected	7.6	Not Detected
Carbon Disulfide	0.80	Not Detected	2.5	Not Detected
2-Propanol	3.2	Not Detected	7.9	Not Detected
trans-1,2-Dichloroethene	0.80	Not Detected	3.2	Not Detected
2-Butanone (Methyl Ethyl Ketone)	0.80	Not Detected	2.4	Not Detected
Tetrahydrofuran	0.80	Not Detected	2.4	Not Detected
1,4-Dioxane	3.2	Not Detected	12	Not Detected
4-Methyl-2-pentanone	0.80	Not Detected	3.3	Not Detected
2-Hexanone	3.2	Not Detected	13	Not Detected
Bromoform	0.80	Not Detected	8.3	Not Detected
4-Ethyltoluene	0.80	Not Detected	4.0	Not Detected
Ethanol	3.2	Not Detected	6.1	Not Detected
Methyl tert-butyl ether	0.80	Not Detected	2.9	Not Detected
3-Chloropropene	3.2	Not Detected	10	Not Detected
2,2,4-Trimethylpentane	0.80	Not Detected	3.8	Not Detected
Naphthalene	3.2	Not Detected	17	Not Detected

Container Type: 6 Liter Summa Canister

Surrogates	%Recovery	Method Limits
Toluene-d8	93	70-130
1,2-Dichloroethane-d4	96	70-130
4-Bromofluorobenzene	93	70-130

Report Date: 19-Apr-2007 16:00

## Air Toxics Ltd.

## AMBIENT AIR METHOD TO14A/TO15

Data file : /chem/msd8.i/8-12apr.b/8041207.d  
 Lab Smp Id: 0704143-02A  
 Inj Date : 12-APR-2007 17:23  
 Operator : JG Inst ID: msd8.i  
 Smp Info : 200mL #12007  
 Misc Info : 5.0"Hg-5.0psi  
 Comment :  
 Method : /chem/msd8.i/8-12apr.b/t14q322b.m  
 Meth Date : 12-Apr-2007 23:51 jgray Quant Type: ISTD  
 Cal Date : 26-MAR-2007 13:11 Cal File: 8032608.d  
 Als bottle: 1  
 Dil Factor: 1.61000  
 Integrator: HP RTE Compound Sublist: AT04+Na.sub  
 Target Version: 3.50 Sample Matrix: AIR  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* CpndVariable

Cpnd Variable Local Compound Variable

CONCENTRATIONS								
RT	EXP RT	(REL RT)	MASS	RESPONSE		TARGET RANGE	RATIO	
				( PPBV)	( PPBV)			
==	=====	=====	====	=====	=====	=====	=====	=====
-----								
* 67	Bromochloromethane					CAS #: 74-97-5		
8.395	8.395	(1.000)	130	205442	25.0000	80.00- 120.00	100.00	
8.395	8.395	(1.000)	128	157039		45.68- 105.68	76.44	
8.395	8.395	(1.000)	49	591599		268.48- 328.48	287.96	
-----								
* 86	1,4-Difluorobenzene					CAS #: 540-36-3		
10.248	10.275	(1.000)	114	896316	25.0000	80.00- 120.00	100.00	
10.248	10.275	(1.000)	88	170965		0.00- 48.01	19.07	
-----								
* 123	Chlorobenzene-d5					CAS #: 3114-55-4		
15.225	15.225	(1.000)	117	675588	25.0000	80.00- 120.00	100.00	
15.225	15.225	(1.000)	82	430865		35.49- 95.49	63.78	
-----								
\$ 80	1,2-Dichloroethane-d4					CAS #: 17060-07-0		
9.473	9.474	(1.128)	65	428649	24.1379	80.00- 120.00	100.00	
9.473	9.474	(1.128)	67	199398		27.92- 87.92	46.52	
-----								
\$ 102	Toluene-d8					CAS #: 2037-26-5		
12.985	12.985	(1.267)	98	817614	23.1957	80.00- 120.00	100.00	
12.985	12.985	(1.267)	70	98954		0.00- 42.61	12.10	

CONCENTRATIONS

ON-COL FINAL

RT EXP RT (REL RT) MASS RESPONSE ( PPEV) ( PPBV) TARGET RANGE RATIO  
== =====

\$ 102 Toluene-d8 (continued)

12.985 12.985 (1.267) 100 540328 40.27- 100.27 66.09

\$ 138 Bromofluorobenzene

CAS #: 460-00-4

16.773 16.773 (1.102) 174 321776 23.2411 23.241 80.00- 120.00 100.00

16.745 16.773 (1.100) 95 565830 138.09- 198.09 175.85

16.773 16.773 (1.102) 176 304457 69.53- 129.53 94.62



Report Date: 19-Apr-2007 16:00

Air Toxics Ltd.

INTERNAL STANDARD COMPOUNDS  
AREA AND RT SUMMARY

Instrument ID: msd8.i  
 Lab File ID: 8041207.d  
 Lab Smp Id: 0704143-02A  
 Analysis Type: VOA  
 Quant Type: ISTD  
 Operator: JG  
 Method File: /chem/msd8.i/8-12apr.b/t14q322b.m  
 Misc Info: 5.0"Hg-5.0psi

Calibration Date: 12-APR-2007  
 Calibration Time: 09:53  
 Level: LOW  
 Sample Type: AIR

COMPOUND	STANDARD	AREA LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	269481	161689	377273	205442	-23.76
86 1,4-Difluorobenze	1204367	722620	1686114	896316	-25.58
123 Chlorobenzene-d5	906226	543736	1268716	675588	-25.45

COMPOUND	STANDARD	RT LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	8.40	8.07	8.73	8.40	0.00
86 1,4-Difluorobenze	10.28	9.95	10.61	10.25	-0.27
123 Chlorobenzene-d5	15.22	14.89	15.55	15.22	0.00

AREA UPPER LIMIT = + 40% of internal standard area.

AREA LOWER LIMIT = - 40% of internal standard area.

RT UPPER LIMIT = + 0.33 minutes of internal standard RT.

RT LOWER LIMIT = - 0.33 minutes of internal standard RT.

Air Toxics Ltd.

RECOVERY REPORT

Client Name: Client SDG: 8-12apr  
Sample Matrix: GAS Fraction: VOA  
Lab Smp Id: 0704143-02A  
Level: LOW Operator: JG  
Data Type: MS DATA SampleType: SAMPLE  
SpikeList File: AT04+NA-2.spk Quant Type: ISTD  
Sublist File: AT04+Na.sub  
Method File: /chem/msd8.i/8-12apr.b/t14q322b.m  
Misc Info: 5.0"Hg-5.0psi

SURROGATE COMPOUND	CONC ADDED PPBV	CONC RECOVERED PPBV	% RECOVERED	LIMITS
\$ 80 1,2-Dichloroethane	25.000	24.138	96.55	70-130
\$ 102 Toluene-d8	25.000	23.196	92.78	70-130
\$ 138 Bromofluorobenzene	25.000	23.241	92.96	70-130

Data File: /chem/msd8.1/8-12apr.1b/8041207.d

Date: 12-APR-2007 17:23

Client ID:


Sample Info: 200ML #12007

Column phase: RTX-624

Instrument: msd8.1

Operator: JG

Column diameter: 0.53



## **QC Results and Raw Data**



AN ENVIRONMENTAL ANALYTICAL LABORATORY

Client Sample ID: Lab Blank

Lab ID#: 0704143-03A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	8041205	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 4/12/07 02:12 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (uG/m3)	Amount (uG/m3)
Freon 12	0.50	Not Detected	2.5	Not Detected
Freon 114	0.50	Not Detected	3.5	Not Detected
Vinyl Chloride	0.50	Not Detected	1.3	Not Detected
Bromomethane	0.50	Not Detected	1.9	Not Detected
Chloroethane	0.50	Not Detected	1.3	Not Detected
Freon 11	0.50	Not Detected	2.8	Not Detected
1,1-Dichloroethene	0.50	Not Detected	2.0	Not Detected
Freon 113	0.50	Not Detected	3.8	Not Detected
Methylene Chloride	0.50	Not Detected	1.7	Not Detected
1,1-Dichloroethane	0.50	Not Detected	2.0	Not Detected
cis-1,2-Dichloroethene	0.50	Not Detected	2.0	Not Detected
Chloroform	0.50	Not Detected	2.4	Not Detected
1,1,1-Trichloroethane	0.50	Not Detected	2.7	Not Detected
Carbon Tetrachloride	0.50	Not Detected	3.1	Not Detected
Benzene	0.50	Not Detected	1.6	Not Detected
1,2-Dichloroethane	0.50	Not Detected	2.0	Not Detected
Trichloroethene	0.50	Not Detected	2.7	Not Detected
1,2-Dichloropropane	0.50	Not Detected	2.3	Not Detected
cis-1,3-Dichloropropene	0.50	Not Detected	2.3	Not Detected
Toluene	0.50	Not Detected	1.9	Not Detected
trans-1,3-Dichloropropene	0.50	Not Detected	2.3	Not Detected
1,1,2-Trichloroethane	0.50	Not Detected	2.7	Not Detected
Tetrachloroethene	0.50	Not Detected	3.4	Not Detected
1,2-Dibromoethane (EDB)	0.50	Not Detected	3.8	Not Detected
Chlorobenzene	0.50	Not Detected	2.3	Not Detected
Ethyl Benzene	0.50	Not Detected	2.2	Not Detected
m,p-Xylene	0.50	Not Detected	2.2	Not Detected
o-Xylene	0.50	Not Detected	2.2	Not Detected
Styrene	0.50	Not Detected	2.1	Not Detected
1,1,2,2-Tetrachloroethane	0.50	Not Detected	3.4	Not Detected
1,3,5-Trimethylbenzene	0.50	Not Detected	2.4	Not Detected
1,2,4-Trimethylbenzene	0.50	Not Detected	2.4	Not Detected
1,3-Dichlorobenzene	0.50	Not Detected	3.0	Not Detected
1,4-Dichlorobenzene	0.50	Not Detected	3.0	Not Detected
alpha-Chlorotoluene	0.50	Not Detected	2.6	Not Detected
1,2-Dichlorobenzene	0.50	Not Detected	3.0	Not Detected
1,3-Butadiene	0.50	Not Detected	1.1	Not Detected
Hexane	0.50	Not Detected	1.8	Not Detected
Cyclohexane	0.50	Not Detected	1.7	Not Detected



AN ENVIRONMENTAL ANALYTICAL LABORATORY

Client Sample ID: Lab Blank

Lab ID#: 0704143-03A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	8041205	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 4/12/07 02:12 PM

Compound	Rpt. Limit (ppbv)	Amount (ppbv)	Rpt. Limit (uG/m3)	Amount (uG/m3)
Heptane	0.50	Not Detected	2.0	Not Detected
Bromodichloromethane	0.50	Not Detected	3.4	Not Detected
Dibromochloromethane	0.50	Not Detected	4.2	Not Detected
Cumene	0.50	Not Detected	2.4	Not Detected
Propylbenzene	0.50	Not Detected	2.4	Not Detected
Chloromethane	2.0	Not Detected	4.1	Not Detected
1,2,4-Trichlorobenzene	2.0	Not Detected	15	Not Detected
Hexachlorobutadiene	2.0	Not Detected	21	Not Detected
Acetone	2.0	Not Detected	4.8	Not Detected
Carbon Disulfide	0.50	Not Detected	1.6	Not Detected
2-Propanol	2.0	Not Detected	4.9	Not Detected
trans-1,2-Dichloroethene	0.50	Not Detected	2.0	Not Detected
2-Butanone (Methyl Ethyl Ketone)	0.50	Not Detected	1.5	Not Detected
Tetrahydrofuran	0.50	Not Detected	1.5	Not Detected
1,4-Dioxane	2.0	Not Detected	7.2	Not Detected
4-Methyl-2-pentanone	0.50	Not Detected	2.0	Not Detected
2-Hexanone	2.0	Not Detected	8.2	Not Detected
Bromoform	0.50	Not Detected	5.2	Not Detected
4-Ethyltoluene	0.50	Not Detected	2.4	Not Detected
Ethanol	2.0	Not Detected	3.8	Not Detected
Methyl tert-butyl ether	0.50	Not Detected	1.8	Not Detected
3-Chloropropene	2.0	Not Detected	6.3	Not Detected
2,2,4-Trimethylpentane	0.50	Not Detected	2.3	Not Detected
Naphthalene	2.0	Not Detected	10	Not Detected

Container Type: NA - Not Applicable

Surrogates	%Recovery	Method Limits
Toluene-d8	100	70-130
1,2-Dichloroethane-d4	100	70-130
4-Bromofluorobenzene	94	70-130

Report Date: 12-Apr-2007 14:33

## Air Toxics Ltd.

## AMBIENT AIR METHOD TO14A/TO15

Data file : /chem/msd8.i/8-12apr.b/8041205.d  
 Lab Smp Id: Lab Blank Client Smp ID: Lab Blank  
 Inj Date : 12-APR-2007 14:12  
 Operator : JG Inst ID: msd8.i  
 Smp Info : 200mL #13673  
 Misc Info : Humid  
 Comment :  
 Method : /chem/msd8.i/8-12apr.b/t14q322b.m  
 Meth Date : 12-Apr-2007 13:04 jgray Quant Type: ISTD  
 Cal Date : 26-MAR-2007 13:11 Cal File: 8032608.d  
 Als bottle: 1  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: AT04+ENS.sub  
 Target Version: 3.50 Sample Matrix: AIR  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* CpndVariable

Cpnd Variable Local Compound Variable

CONCENTRATIONS								
RT	EXP RT	(REL RT)	MASS	RESPONSE		TARGET RANGE	RATIO	
				( PPBV)	( PPBV)			
==	=====	=====	====	=====	=====	=====	=====	=====
-----								
* 67	Bromochloromethane					CAS #: 74-97-5		
8.395	8.395	(1.000)	130	195237	25.0000	80.00- 120.00	100.00	
8.395	8.395	(1.000)	128	156432		45.68- 105.68	80.12	
8.395	8.395	(1.000)	49	593177		268.48- 328.48	303.82	
-----								
* 86	1,4-Difluorobenzene					CAS #: 540-36-3		
10.248	10.275	(1.000)	114	874623	25.0000	80.00- 120.00	100.00	
10.248	10.275	(1.000)	88	163816		0.00- 48.01	18.73	
-----								
* 123	Chlorobenzene-d5					CAS #: 3114-55-4		
15.224	15.225	(1.000)	117	669037	25.0000	80.00- 120.00	100.00	
15.224	15.225	(1.000)	82	461580		35.49- 95.49	68.99	
-----								
\$ 80	1,2-Dichloroethane-d4					CAS #: 17060-07-0		
9.473	9.474	(1.128)	65	423904	25.1184	80.00- 120.00	100.00	
9.473	9.474	(1.128)	67	203726		27.92- 87.92	48.06	
-----								
\$ 102	Toluene-d8					CAS #: 2037-26-5		
12.985	12.985	(1.267)	98	862817	25.0852	80.00- 120.00	100.00	
12.985	12.985	(1.267)	70	105957		0.00- 42.61	12.28	

CONCENTRATIONS

ON-COL FINAL

RT EXP RT (REL RT) MASS RESPONSE ( PPEV) ( PPBV) TARGET RANGE RATIO  
== =====

\$ 102 Toluene-d8 (continued)

12.985 12.985 (1.267) 100 559325 40.27- 100.27 64.83

\$ 138 Bromofluorobenzene

CAS #: 460-00-4

16.773 16.773 (1.102) 174 321940 23.4807 23.481 80.00- 120.00 100.00

16.745 16.773 (1.100) 95 568164 138.09- 198.09 176.48

16.773 16.773 (1.102) 176 322063 69.53- 129.53 100.04



Report Date: 12-Apr-2007 14:33

Air Toxics Ltd.

INTERNAL STANDARD COMPOUNDS  
AREA AND RT SUMMARY

Instrument ID: msd8.i  
 Lab File ID: 8041205.d  
 Lab Smp Id: Lab Blank  
 Analysis Type: VOA  
 Quant Type: ISTD  
 Operator: JG  
 Method File: /chem/msd8.i/8-12apr.b/t14q322b.m  
 Misc Info: Humid

Calibration Date: 12-APR-2007  
 Calibration Time: 09:53  
 Client Smp ID: Lab Blank  
 Level: LOW  
 Sample Type: AIR

COMPOUND	STANDARD	AREA LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	269481	161689	377273	195237	-27.55
86 1,4-Difluorobenze	1204367	722620	1686114	874623	-27.38
123 Chlorobenzene-d5	906226	543736	1268716	669037	-26.17

COMPOUND	STANDARD	RT LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	8.40	8.07	8.73	8.39	0.00
86 1,4-Difluorobenze	10.28	9.95	10.61	10.25	-0.27
123 Chlorobenzene-d5	15.22	14.89	15.55	15.22	0.00

AREA UPPER LIMIT = + 40% of internal standard area.

AREA LOWER LIMIT = - 40% of internal standard area.

RT UPPER LIMIT = + 0.33 minutes of internal standard RT.

RT LOWER LIMIT = - 0.33 minutes of internal standard RT.

Data File: /chem/msd8.1/8-12apr.1b/8041205.d

Date: 12-APR-2007 14:12

Client ID: Lab Blank


Sample Info: 200mL #13673

Column phase: RTX-624

Instrument: msd8.1

Operator: JG

Column diameter: 0.53



# LEVEL-IV VALIDATABLE

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

SURROGATE RECOVERY FORM

Lab Name: AIR TOXICS LIMITED.

SDG No.: 0704143

	CLIENT SAMPLE NO.	SURROGATE % RECOVERY						TOTAL OUT
		1,2-Dichloroethane-d 4	#	Toluene-d8	#	4-Bromofluorobenze ne	#	
01	000001016 Downwind	97		94		93		0
02	12007 Upwind	96		93		93		0
03	Lab Blank	100		100		94		0
04	CCV	95		99		102		0
05	LCS	105		96		106		0
06								0
07								0
08								0
09								0
10								0
11								0
12								0
13								0
14								0
15								0
16								0
17								0
18								0
19								0
20								0
21								0
22								0
23								0
24								0

Surrogate Recovery Limits

1,2-Dichloroethane-d4 70 - 130

Toluene-d8 70 - 130

4-Bromofluorobenzene 70 - 130

\* Designates values outside of QC limits

# LEVEL-IV VALIDATABLE

Modified EPA Method TO-15 GC/MS Full Scan

INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: AIR TOXICS, LTD  
 Lab File ID: 8041202.d  
 Instrument ID: msd8.i

SDG No: 0704143  
 Date Analyzed: 04/12/2007  
 Time Analyzed: 09:53 AM

	Chlorobenzene-d5		RT		1,4-Difluorobenzene		RT		Bromochloromethane		RT	
	Area	#		#	Area	#		#	Area	#		#
24-HOUR STD	906226		15.22		1204367		10.28		269481		8.4	
UPPER LIMIT	1268716		15.55		1686114		10.61		377273		08.73	
LOWER LIMIT	543736		14.89		722620		09.95		161689		08.07	
CLIENT SAMPLE NO												
01 000001016 Downwind	673199		15.22		923293		10.25		209705		8.4	
02 12007 Upwind	675588		15.22		896316		10.25		205442		8.4	
03 Lab Blank	669037		15.22		874623		10.25		195237		8.4	
04 CCV	906226		15.22		1204367		10.28		269481		8.4	
05 LCS	694716		15.22		907795		10.25		189537		8.4	
06												
07												
08												
09												
10												
11												
12												
13												
14												
15												
16												
17												
18												
19												
20												
21												
22												

'Area Upper Limit=+40% of internal standard area'  
 'Area Lower Limit=-40% of internal standard area'

RT Upper Limit=+0.33 minutes of internal standard RT  
 RT Lower Limit=-0.33 minutes of internal standard RT

\* Designates values outside of QC limits



Air Toxics Ltd.

INITIAL CALIBRATION DATA

Start Cal Date : 22-MAR-2007 10:41  
 End Cal Date : 26-MAR-2007 13:11  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 3.50  
 Integrator : HP RTE  
 Method file : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Cal Date : 27-Mar-2007 06:43 sscott  
 Curve Type : Average

Compound	0.20000 Level 1	0.50000 Level 2	2.000 Level 3	25.000 Level 4	50.000 Level 5	100.000 Level 6	Level 7	RRF	% RSD
8 Chloromethane	200.000 2.36981	+++++	4.29882	2.67832	2.54830	2.48799		2.87665	27.906
9 Butane	0.62557	+++++	1.15611	0.68183	0.64366	0.63598		0.74863	30.559 <-
11 1,3-Butadiene	2.36965	3.35923	3.70712	2.53676	2.41352	2.42576		2.80201	20.684
10 Vinyl Chloride	2.35886	2.96430	3.50969	2.54251	2.41065	2.42632		2.70205	16.758
12 Methanol	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++
13 Bromomethane	1.29205	2.04242	1.71377	1.27988	1.22262	1.31172		1.47708	22.266
14 Vinyl Bromide	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++
15 Isopentane	3.99290	+++++	6.64086	4.22299	4.13290	4.12583		4.62310	24.463
16 Chloroethane	1.22781	1.73630	1.73111	1.30147	1.19950	1.24630		1.40708	18.136
17 Dichlorofluoromethane/Fr21	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++



Air Toxics Ltd.

INITIAL CALIBRATION DATA

Start Cal Date : 22-MAR-2007 10:41  
 End Cal Date : 26-MAR-2007 13:11  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 3.50  
 Integrator : HP RTE  
 Method file : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Cal Date : 27-Mar-2007 06:43 sscott  
 Curve Type : Average

Compound	0.20000	0.50000	2.000	25.000	50.000	100.000	RRF	% RSD
	Level 1	Level 2	Level 3	Level 4	Level 5	Level 6		
	200.000							
	Level 7							
27 Freon 113	+++++	2.82232	3.22222	2.14765	2.03365	2.04910		
	1.95455						2.37158	22.064
29 1,1-Dichloroethene	+++++	5.25838	5.05771	3.62935	3.42759	3.47883		
	3.36907						4.03682	21.678
30 Acetone	+++++	+++++	2.05734	1.38732	1.24688	1.27242		
	1.25483						1.44376	24.081
31 Acetaldehyde	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
	+++++							
32 Freon143a	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
	+++++							
33 Carbon Disulfide	+++++	7.95920	7.71487	5.56651	5.39873	5.39231		
	5.27666						6.21805	20.261
34 2-Propanol	+++++	+++++	6.85744	5.44165	5.29419	5.39097		
	5.29799						5.65645	11.921
35 Acetonitrile	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
	+++++							
36 Cyclopentene	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
	+++++							
37 3-Chloropropene	+++++	+++++	1.33829	0.93944	0.92232	0.92704		
	0.89360						1.00414	18.678



Air Toxics Ltd.

INITIAL CALIBRATION DATA

Start Cal Date : 22-MAR-2007 10:41  
 End Cal Date : 26-MAR-2007 13:11  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 3.50  
 Integrator : HP RTE  
 Method file : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Cal Date : 27-Mar-2007 06:43 sscott  
 Curve Type : Average

Compound	0.20000 Level 1	0.50000 Level 2	2.000 Level 3	25.000 Level 4	50.000 Level 5	100.000 Level 6	Level 7	RRF	% RSD
38 tert-Butyl-Alcohol	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++
39 Methylene Chloride	+++++	4.53517	4.27698	3.10830	2.90127	2.94899		3.43297	22.248
40 Acrylonitrile	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++
41 2-Methyl-1-Butene	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++
42 MTBE	+++++	8.85745	8.57775	6.27521	5.93540	5.99233		6.91999	20.258
44 1-Pentene	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++
43 trans-1,2-Dichloroethene	+++++	2.94934	2.54247	1.79033	1.70500	1.72263		2.06297	26.460
45 Hexane	+++++	6.50743	6.29843	4.63124	4.39618	4.38297		5.07809	20.393
46 Ethyl Ether	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++
47 Ethanol-high	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++

Air Toxics Ltd.

INITIAL CALIBRATION DATA

Start Cal Date : 22-MAR-2007 10:41  
 End Cal Date : 26-MAR-2007 13:11  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 3.50  
 Integrator : HP RTE  
 Method file : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Cal Date : 27-Mar-2007 06:43 sscott  
 Curve Type : Average

Compound	0.20000	0.50000	2.000	25.000	50.000	100.000	—	% RSD
	Level 1	Level 2	Level 3	Level 4	Level 5	Level 6	RRF	
	200.000							
	Level 7							
48 Isopropyl ether	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
49 Propanal	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
50 Chloroprene	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
51 1-Propanol	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
52 1,1-Dichloroethane	+++++	5.18323	5.10569	3.98508	3.87923	3.82996		
	3.75100						4.28903	15.560
53 Bromoethane	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
54 Vinyl Acetate	+++++	+++++	0.51558	0.49264	0.52336	0.51139		
	0.50793						0.51018	2.229
55 Iodomethane	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
56 2,3-Dimethylbutane	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
57 Ethyl-tert-butyl Ether	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++



## Air Toxics Ltd.

## INITIAL CALIBRATION DATA

Start Cal Date : 22-MAR-2007 10:41  
 End Cal Date : 26-MAR-2007 13:11  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 3.50  
 Integrator : HP RTE  
 Method file : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Cal Date : 27-Mar-2007 06:43 sscott  
 Curve Type : Average

Compound	0.20000	0.50000	2.000	25.000	50.000	100.000		
	Level 1	Level 2	Level 3	Level 4	Level 5	Level 6	RRF	% RSD
	200.000							
	Level 7							
69 Chloroform	6.17803	5.06987	4.49947	3.38493	3.22658	3.26840		
	3.14998						4.11103	28.585
70 2-Butanol	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
71 1,1-Dichloropropene	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
72 Cyclohexane	+++++	3.97979	3.99673	2.74032	2.57034	2.59050		
	2.50871						3.06440	23.485
74 3-Methyl-1-Hexene	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
73 1,1,1-Trichloroethane	+++++	4.63376	4.65460	3.38033	3.20850	3.24716		
	3.14555						3.71165	19.572
75 Carbon Tetrachloride	+++++	3.35462	3.79930	2.81089	2.69706	2.70088		
	2.62506						2.99797	15.802
76 Isobutanol	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
77 tert-amyl-Methyl Ether	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
78 2,2,4-Trimethylpentane	+++++	16.43298	17.99115	13.37425	12.70715	12.77683		
	12.50824						14.29843	16.282

Air Toxics Ltd.

INITIAL CALIBRATION DATA

Start Cal Date : 22-MAR-2007 10:41  
 End Cal Date : 26-MAR-2007 13:11  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 3.50  
 Integrator : HP RTE  
 Method file : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Cal Date : 27-Mar-2007 06:43 sscott  
 Curve Type : Average

Compound	0.20000	0.50000	2.000	25.000	50.000	100.000	RRF	% RSD
	Level 1	Level 2	Level 3	Level 4	Level 5	Level 6		
	200.000							
	Level 7							
79 Benzene	2.03131	1.47626	1.61743	1.22633	1.16525	1.12959		
	1.10050						1.39238	24.519
81 1,2-Dichloroethane	+++++	0.84058	0.78407	0.61043	0.57149	0.55908		
	0.54673						0.65206	19.511
82 Heptane	+++++	0.28296	0.17842	0.14079	0.13652	0.13223		
	0.12701						0.16632	36.075
83 2,3-Dimethylpentane	+++++	+++++	1.93177	+++++	1.34790	+++++		
	1.27765						1.51911	23.639
84 1-Methoxy-2-Propanol	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
85 2,3,4-Trimethylpentane	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
87 1-Butanol	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
88 Methyl Methacrylate	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
89 2-Pentanone	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
90 Pentanal	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++

## Air Toxics Ltd.

## INITIAL CALIBRATION DATA

Start Cal Date : 22-MAR-2007 10:41  
 End Cal Date : 26-MAR-2007 13:11  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 3.50  
 Integrator : HP RTE  
 Method file : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Cal Date : 27-Mar-2007 06:43 sscott  
 Curve Type : Average

Compound	0.20000	0.50000	2.000	25.000	50.000	100.000	—	% RSD
	Level 1	Level 2	Level 3	Level 4	Level 5	Level 6	RRF	
	200.000							
	Level 7							
91 Ethyl Acrylate	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
92 Trichloroethene	+++++	0.48093	0.59807	0.43694	0.40476	0.38896		
	0.37753						0.44787	18.426
93 Methyl Cyclohexane	+++++	4.88062	4.87830	3.41153	3.30277	3.30451		
	3.22043						3.83302	21.206
94 Dibromomethane	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
95 1,2-Dichloropropane	+++++	0.67532	0.61108	0.46484	0.45017	0.43591		
	0.42410						0.51024	20.748
96 1,4-Dioxane	+++++	+++++	0.34178	0.26004	0.24908	0.24171		
	0.23737						0.26600	16.251
97 Octane	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
98 Bromodichloromethane	+++++	0.89246	0.94044	0.71928	0.67428	0.65833		
	0.64210						0.75448	17.094
99 1-Nitropropane	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
100 cis-1,3-Dichloropropene	+++++	0.76889	0.76757	0.60016	0.57911	0.58023		
	0.56488						0.64347	15.119

Air Toxics Ltd.

INITIAL CALIBRATION DATA

Start Cal Date : 22-MAR-2007 10:41  
 End Cal Date : 26-MAR-2007 13:11  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 3.50  
 Integrator : HP RTE  
 Method file : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Cal Date : 27-Mar-2007 06:43 sscott  
 Curve Type : Average

Compound	0.20000 Level 1	0.50000 Level 2	2.000 Level 3	25.000 Level 4	50.000 Level 5	100.000 Level 6	Level 7	RRF	% RSD
101 4-Methyl-2-pentanone	0.43377	0.67928	0.62422	0.46863	0.46135	0.44565		0.51882	20.265
103 Toluene	1.12371	1.34027	1.57212	1.22204	1.16762	1.13880		1.26076	13.605
104 1-Methoxy-2-propyl acetate	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++
105 Epichlorohydrin	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++
106 trans-1,3-Dichloropropene	0.76552	1.11760	0.96729	0.81239	0.78942	0.77860		0.87181	16.208
107 1,3-Dichloropropane	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++
108 1,1,2-Trichloroethane	0.42903	0.65309	0.58031	0.47394	0.45806	0.45084		0.50754	17.498
110 alpha-Pinene	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++
109 Tetrachloroethene	0.47984	0.68363	0.56664	0.50273	0.50196	0.49965		0.53907	14.229
111 Butyl Acetate	+++++	+++++	+++++	+++++	+++++	+++++		+++++	+++++

Air Toxics Ltd.

INITIAL CALIBRATION DATA

Start Cal Date : 22-MAR-2007 10:41  
 End Cal Date : 26-MAR-2007 13:11  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 3.50  
 Integrator : HP RTE  
 Method file : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Cal Date : 27-Mar-2007 06:43 sscott  
 Curve Type : Average

Compound	0.20000	0.50000	2.000	25.000	50.000	100.000	RRF	% RSD
	Level 1	Level 2	Level 3	Level 4	Level 5	Level 6		
	200.000							
	Level 7							
112 2-Hexanone	+++++	+++++	0.88087	0.79222	0.78548	0.79567		
	0.75095						0.80104	5.997
113 trans-1,4-dichloro-2-butene	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
114 Dibromochloromethane	+++++	0.83418	0.80967	0.62979	0.62927	0.62191		
	0.60765						0.68875	15.065
115 1,2-Dibromoethane	+++++	0.76691	0.88801	0.67253	0.65651	0.65470		
	0.63435						0.71217	13.750
116 beta-Pinene	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
117 Decane	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
118 Diisobutyl Ketone	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
119 Alphasethylstyrene	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
120 Dicyclopentadiene	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
121 1,1,1,2-Tetrachloroethane	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++



## Air Toxics Ltd.

## INITIAL CALIBRATION DATA

Start Cal Date : 22-MAR-2007 10:41  
 End Cal Date : 26-MAR-2007 13:11  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 3.50  
 Integrator : HP RTE  
 Method file : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Cal Date : 27-Mar-2007 06:43 sscott  
 Curve Type : Average

Compound	0.20000 Level 1	0.50000 Level 2	2.000 Level 3	25.000 Level 4	50.000 Level 5	100.000 Level 6	200.000 Level 7	RRF	% RSD
122 D-Limonene	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
124 Chlorobenzene	+++++	1.40231	1.40389	1.09599	1.09995	1.04079	1.00407	1.17450	15.381
125 Bis(2-chloroethyl) ether	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
126 Nonane	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
127 Ethyl Benzene	+++++	0.82062	0.74191	0.61784	0.60525	0.59030	0.58046	0.65940	14.931
128 m,p-Xylene	+++++	1.04206	0.97220	0.79323	0.78399	0.76243	0.73366	0.84792	14.971
129 Undecane	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
130 o-Xylene	+++++	0.96930	1.00716	0.76148	0.74636	0.72189	0.69957	0.81763	16.433
132 2-Heptanone	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
131 Styrene	1.69313	1.14679	1.49850	1.09415	1.11463	1.13886	1.12395	1.25857	18.889

## Air Toxics Ltd.

## INITIAL CALIBRATION DATA

Start Cal Date : 22-MAR-2007 10:41  
 End Cal Date : 26-MAR-2007 13:11  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 3.50  
 Integrator : HP RTE  
 Method file : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Cal Date : 27-Mar-2007 06:43 sscott  
 Curve Type : Average

Compound	0.20000	0.50000	2.000	25.000	50.000	100.000	—	% RSD
	Level 1	Level 2	Level 3	Level 4	Level 5	Level 6	RRF	
	200.000							
	Level 7							
133 Bromoform	+++++	0.61756	0.64606	0.53857	0.55383	0.53227		
	0.52666						0.56916	8.818
134 Cyclohexanone	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
135 Cumene	3.45662	2.66945	2.84554	2.18442	2.13966	2.10117		
	2.06896						2.49512	20.937
136 1-chloro-2-Bromopropane	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
137 Bromobenzene	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
139 1,2,3-Trichloropropane	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
140 Dodecane	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
141 2-Chlorotoluene	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
142 1,1,2,2-Tetrachloroethane	+++++	1.37134	1.42089	1.07599	1.06692	1.06370		
	1.02790						1.17112	15.006
143 Propylbenzene	+++++	3.30360	3.36767	2.68855	2.64283	2.69125		
	2.23406						2.82133	15.388

Air Toxics Ltd.

INITIAL CALIBRATION DATA

Start Cal Date : 22-MAR-2007 10:41  
 End Cal Date : 26-MAR-2007 13:11  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 3.50  
 Integrator : HP RTE  
 Method file : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Cal Date : 27-Mar-2007 06:43 sscott  
 Curve Type : Average

Compound	0.20000	0.50000	2.000	25.000	50.000	100.000	RRF	% RSD
	Level 1	Level 2	Level 3	Level 4	Level 5	Level 6		
	200.000							
	Level 7							
144 4-Chlorotoluene	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
145 4-Ethyltoluene	+++++	2.88824	2.85962	2.34847	2.37703	2.39701		
	1.84102						2.45190	15.787
146 1,3,5-Trimethylbenzene	+++++	2.72711	3.06967	2.08844	2.05129	2.04889		
	1.93827						2.32061	19.960
147 2,6-Dimethyl-1-propanol	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
148 tert-Butylbenzene	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
149 Pentachloroethane	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
150 sec-Butylbenzene	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
151 1,2,4-Trimethylbenzene	+++++	2.58151	2.58001	1.89703	1.87135	1.88091		
	1.79705						2.10131	17.749
152 p-Cymene	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
153 1,2,3-Trimethylbenzene	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++

Air Toxics Ltd.

INITIAL CALIBRATION DATA

Start Cal Date : 22-MAR-2007 10:41  
 End Cal Date : 26-MAR-2007 13:11  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 3.50  
 Integrator : HP RTE  
 Method file : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Cal Date : 27-Mar-2007 06:43 sscott  
 Curve Type : Average

Compound	0.20000	0.50000	2.000	25.000	50.000	100.000	RRF	% RSD
	Level 1	Level 2	Level 3	Level 4	Level 5	Level 6		
	200.000							
	Level 7							
154 1,3-Dichlorobenzene	+++++	1.07365	0.96922	0.85086	0.88100	0.84628		
	0.85516						0.91270	10.000
155 1,4-Dichlorobenzene	+++++	1.44035	1.26100	1.03804	1.08377	1.16840		
	0.99164						1.16387	14.264
156 alpha-Chlorotoluene	+++++	1.83854	1.73296	1.52280	1.65167	1.67027		
	1.69415						1.68507	6.144
157 Butylbenzene	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
158 1,2-Dichlorobenzene	+++++	1.26589	1.15974	0.89919	0.92239	0.88418		
	0.87873						1.00169	16.735
159 Hexachloroethane	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
160 Aniline	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
161 1,2-Dibromo-3-Chloropropane	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
162 Isooctyl Alcohol	+++++	+++++	+++++	+++++	+++++	+++++		
	+++++						+++++	+++++
163 1,2,4-Trichlorobenzene	+++++	+++++	0.72875	0.47508	0.58404	0.65031		
	0.68067						0.62377	15.759

Air Toxics Ltd.

INITIAL CALIBRATION DATA

Start Cal Date : 22-MAR-2007 10:41  
 End Cal Date : 26-MAR-2007 13:11  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 3.50  
 Integrator : HP RTE  
 Method file : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Cal Date : 27-Mar-2007 06:43 sscott  
 Curve Type : Average

Compound	0.20000 Level 1	0.50000 Level 2	2.000 Level 3	25.000 Level 4	50.000 Level 5	100.000 Level 6	RRF	% RSD
164 Hexachlorobutadiene	0.66959	+++++	1.22486	0.72372	0.69039	0.67531	0.79677	30.150
165 Naphthalene	1.77256	+++++	2.69108	1.63439	1.83942	2.02809	1.99311	20.828
166 1,2,3-Trichlorobenzene	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
167 1,3,5-Trichlorobenzene	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
168 Isooctyl Acrylate	+++++	+++++	+++++	+++++	+++++	+++++	+++++	+++++
194 2-Methylpentane	1.48655	+++++	2.44385	+++++	1.60008	+++++	1.84349	28.371
195 Thiopene	0.63204	+++++	0.98460	+++++	0.65581	+++++	0.75748	26.014
196 Indan	1.73600	+++++	2.66528	+++++	1.70640	+++++	2.03589	26.783
197 Indene	1.18332	+++++	1.74642	+++++	1.09694	+++++	1.34222	26.277
\$ 80 1,2-Dichloroethane-d4	2.36561	2.08429	2.14741	2.23426	2.05489	2.18260	2.16099	5.197

Air Toxics Ltd.

INITIAL CALIBRATION DATA

Start Cal Date : 22-MAR-2007 10:41  
 End Cal Date : 26-MAR-2007 13:11  
 Quant Method : ISTD  
 Origin : Disabled  
 Target Version : 3.50  
 Integrator : HP RTE  
 Method file : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Cal Date : 27-Mar-2007 06:43 sscott  
 Curve Type : Average

Compound	0.20000 Level 1	0.50000 Level 2	2.000 Level 3	25.000 Level 4	50.000 Level 5	100.000 Level 6	RRF	% RSD
	200.000 Level 7							
\$ 102 Toluene-d8	0.99667	0.96945	0.94994	0.98725	0.97848	0.99102		
	1.00924						0.98315	1.972
\$ 138 Bromofluorobenzene	0.49812	0.46909	0.50906	0.51341	0.54503	0.52622		
	0.52541						0.51234	4.726

# Calibration History

Method : /chem/msd8.i/8-26mar.b/t14q322b.m  
Start Cal Date: 22-MAR-2007 10:41  
End Cal Date : 26-MAR-2007 13:11

## Initial Calibration

Injection Date	Sublist	Calibration File
Cal Level: 1 , Cal Amount: 0.20000		
22-MAR-2007 10:41	AFCEElow	/chem/msd8.i/8-22mar.b/8032203.d
Cal Level: 2 , Cal Amount: 0.50000		
22-MAR-2007 17:30	ICALlevel2	/chem/msd8.i/8-22mar.b/8032215.d
Cal Level: 3 , Cal Amount: 2.00000		
26-MAR-2007 12:15	sp5b	/chem/msd8.i/8-26mar.b/8032606.d
22-MAR-2007 11:38	AT04mdl+Na+ENSR	/chem/msd8.i/8-22mar.b/8032205.d
Cal Level: 4 , Cal Amount: 25.00000		
22-MAR-2007 12:06	AT04+ENS	/chem/msd8.i/8-22mar.b/8032206.d
Cal Level: 5 , Cal Amount: 50.00000		
26-MAR-2007 12:43	sp5b	/chem/msd8.i/8-26mar.b/8032607.d
22-MAR-2007 12:34	AT04+ENS	/chem/msd8.i/8-22mar.b/8032207.d
Cal Level: 6 , Cal Amount: 100.00000		
22-MAR-2007 13:02	AT04+ENS	/chem/msd8.i/8-22mar.b/8032208.d
Cal Level: 7 , Cal Amount: 200.00000		
26-MAR-2007 13:11	sp5b	/chem/msd8.i/8-26mar.b/8032608.d
22-MAR-2007 13:30	AT04+ENS	/chem/msd8.i/8-22mar.b/8032209.d

Continuing Calibration

Ccal Level Mode: GLOBAL LEVEL 8

```
+-----+-----+-----+-----+-----+-----+-----+-----+-----+
| Ccal Level: 8 , Ccal Amount: 50.000                                                                                       |
+-----+-----+-----+-----+-----+-----+-----+-----+-----+
| 26-MAR-2007 12:43 |sp5bccv                               |/chem/msd8.i/8-26mar.b/8032607a.d |
+-----+-----+-----+-----+-----+-----+-----+-----+-----+
```



@ Air Toxics Ltd.

ION ABUNDANCE CRITERIA

% REL. ABUNDANCE

MSD-8

Logbook #: 1478

m/z	REL. ABUNDANCE
50	15.0 - 40.0% of mass 95
75	30.0 - 60.0% of mass 95
95	Base peak, 100.00% relative abundance
96	5.0 - 9.0% of mass 95
173	Less than 2.0% of mass 174
174	Greater than 50.0% of mass 95
175	5.0 - 9.0% of mass 174
176	Greater than 95.0% but less than 101.0% of mass 174
177	5.0 - 9.0% of mass 176

1 - value in parenthesis is % mass 174  
 2 - value in parenthesis is % mass 176  
 Verify 176/174 m/z Ratio:  $675517/701792 \times 100 = 96.26$

NOAH Cart #: NA

File #: NA

BFB Injection Date: 3/22/07  
 BFB Injection Time: 0929  
 BFB File ID: 8032201  
 Tekmar Purge Flow: 14.8 mL/min  
 Vacuum: 6.5e-6  
 IS/Std #: 1443-3 Exp. Date: 5/24/07  
 BCM: 283735  
 1,4-DFB: 1370859  
 CB-d5: 1067063  
 Verified CCV IS vs ICAL mid-point (-40%D) [Signature]

Calculation Check:

ppbv of compound =  $\frac{\text{Area}_{\text{Sample}}}{\text{Area}_{\text{IS}}} \times \text{Conc}_{\text{IS}} \times \text{RRF}$

$= \frac{(1341361)}{(1370859)} \times (25) \times (0.98315) = 24.881$

Reported Result 24.881

File ID: 8032207  
 Compound: Toluene-d8  
 Initials: [Signature]

Use	File #	Sample / Client Name	Can #	Pressure	Amt Loaded	DF	Loader Init.	Date Analyzed	Time Analyzed	Review Init.	Comments
✓	8032201	BFB Tune check	843-174	50g	2 ml	1.0	TSS	3-22-07	9:29	TD	
✓	02	Lab Blank	13673	Humid	200ml		[Signature]		1014	[Signature]	
✓	03	ICM (200ppb) Lvl 1	1487-115	0.2ppbv	0.2ml				1041	[Signature]	ES THQ3224
✓	04			Lvl 2	0.5				1110	[Signature]	
✓	05			Lvl 3	2				1138	[Signature]	
✓	06			Lvl 4	25				1206	[Signature]	
✓	07			Lvl 5	50				1234	[Signature]	
✓	08			Lvl 6	100				1302	[Signature]	
✓	09			Lvl 7	200				1330	[Signature]	

Date: 3/22/07

10	X	8032210	Lab blank	13673	Humid	200ml	100	gpc	3/23/07	1359	gpc	-
11	X									1428	gpc	-
12	X									1456	gpc	-
13	X									1606	gpc	-
14	X									1702	gpc	-
15	✓		TCM (200gpc) Lvl 2	1487-115	0.5gpc	0.5ml				1730	gpc	-
16	✓		LS #1408-408	200gpc	50gpc	50ml				1819	gpc	-
17	X		Lab Blank	13673	Humid	second	100	gpc		1949	gpc	-
18	X		Lab Blank							2000	gpc	-
19	X		Lab Blank							2041	gpc	-
20	X		Lab Blank							2224	gpc	-
21												
22												
23												
24												
25												
26												
27												
28												
29												
30												
31												
32												

Comments: WST Flow meter SN: US05E27601 Flow Controller SN: AA9506172

Exp: 8/19/07 Actual: 22.4 ml/min

Nominal: 25.5 ml/min

Signature [Signature]

Date 3/23/07

@ Air Toxics Ltd.

**MSD-8**

**ION ABUNDANCE CRITERIA**

m/z	REL. ABUNDANCE	% REL. ABUNDANCE
50	15.0 - 40.0% of mass 95	29.0%
75	30.0 - 60.0% of mass 95	54.1%
95	Base peak; 100.00% relative abundance	100.00
96	5.0 - 9.0% of mass 95	6.4%
173	Less than 2.0% of mass 174	(0.10)¹
174	Greater than 50.0% of mass 95	52.0%
175	5.0 - 9.0% of mass 174	(2.95)¹
176	Greater than 95.0% but less than 101.0% of mass 174	(95.40)¹
177	5.0 - 9.0% of mass 176	(6.53)²

Logbook #: 1478  
 BFB Injection Date: 3/20/07  
 BFB Injection Time: 09:05  
 BFB File ID: 0032601  
 Tekmar Purge Flow: 2.3/2.0/0.7/0.4  
 Vacuum: \_\_\_\_\_

IS/Std.#:	<u>1143-3</u>	Exp. Date:	<u>9/10/07</u>
BCM	<u>31809</u>		
1.4-DFB	<u>190342</u>		
CB-d5	<u>1370416</u>		

Verified CCV IS vs ICAL mid-point (-40%D) E7D

Verify 176/174 m/z Ratio:  $\frac{(1731969)}{(1020742)} \times \frac{(1020742)}{(1020742)} = 95.40$

NOAH Cart #: 9/5 File #: F032308 / 1403267

**Calculation Check:**

ppbv of compound =  $\frac{\text{Area}_{\text{sample}}}{\text{Area}_{\text{std}}} \times \text{Conc.}_{\text{std}} \times \text{RRF}$

$= \frac{(1731969)}{(1020742)} \times (0.99315) = 24.189$

Reported Result: 24.189

File ID:	<u>0032602</u>
Compound:	<u>161-d4</u>
Initials:	<u>E7D</u>

Use	File #	Sample / Client Name	Can #	Pressure	Amnt Loaded	DF	Loader Init.	Date Analyzed	Time Analyzed	Review Init.	Comments
X	0032601	BFB TUNE CHECK	142-242	5.0ppv	2.0ML	1.00	E7D	3/20/07	1903	E7D	
	02	(UN-1 (200ppb))	147-307	5.0ppv	5.0ml		E7D		0932	E7D	NO Naps
	03	(US-1 (200ppb))	148-408	5.0ppv	5.0ml		E7D		1000	E7D	
	04	Gas Std (100ppb)	149-509	5.0ppv	5.0ml		E7D		1111	E7D	
	05	System blank	150-610	5.0ppv	2.0ml		E7D		1139	E7D	
X	06	ITL Level 3 (100ppb)	151-711	2.0ppv	2.0ml		E7D		1205	AS	SPCB
X	07	ITL Level 3 (100ppb)	152-812	5.0ppv	5.0ml		E7D		1243	AS	SPCB/SPC ON
X	08	ITL Level 3 (100ppb)	153-913	2.0ppv	2.0ml		E7D		1311	AS	
X	09	System blank	154-014	5.0ppv	2.0ml		E7D		1430	AS	

**Initial Calibration Narrative for t14q322a.m**

A seven-point initial calibration was analyzed on MSD-8 on March 22, 2007. As noted on the accompanying analytical run log(s), the following point (Level 2) was re-analyzed due to:

- a. anomalous unacceptable linearity for THF, trans-1,2-DCE, Bromomethane

The following compounds used 0.2 as the lowest calibration concentration:  
Benzene, Chloroform, Styrene and Cumene.

Air Toxics Ltd.  
 Modified EPA Methods TO-14A/TO-15  
 Internal Standard and Associated Target Compounds and Surrogates

<b>Bromochloromethane</b>
<b>Target Compounds:</b>
Freon 12
Freon 114
Chloromethane
Vinyl Chloride
1,3-Butadiene
Bromomethane
Chloroethane
Freon 11
Ethanol
Freon 113
1,1-Dichloroethene
Acetone
2-Propanol
Carbon Disulfide
3-Chloropropene
Methylene Chloride
Methyl tert-butyl ether
trans-1,2-Dichloroethene
Hexane
1,1-Dichloroethane
2-Butanone (Methyl Ethyl Ketone)
cis-1,2-Dichloroethene
Tetrahydrofuran
Chloroform
1,1,1-Trichloroethane
Cyclohexane
Carbon Tetrachloride
2,2,4-Trimethylpentane
<b>Surrogates:</b>
1,2-Dichloroethane-d4

<b>1,4-Difluorobenzene</b>
<b>Target Compounds:</b>
Benzene
1,2-Dichloroethane
Heptane
Trichloroethene
1,2-Dichloropropane
1,4-Dioxane
Bromodichloromethane
cis-1,3-Dichloropropene
4-Methyl-2-pentanone
Toluene
<b>Surrogates:</b>
Toluene-d8

<b>Chlorobenzene-d5</b>
<b>Target Compounds:</b>
trans-1,3-Dichloropropene
1,1,2-Trichloroethane
Tetrachloroethene
2-Hexanone
Dibromochloromethane
1,2-Dibromoethane (EDB)
Chlorobenzene
Ethyl Benzene
m,p-Xylene
o-Xylene
Styrene
Bromoform
Cumene
1,1,2,2-Tetrachloroethane
Propylbenzene
4-Ethyltoluene
1,3,5-Trimethylbenzene
1,2,4-Trimethylbenzene
1,3-Dichlorobenzene
1,4-Dichlorobenzene
alpha-Chlorotoluene
1,2-Dichlorobenzene
1,2,4-Trichlorobenzene
Hexachlorobutadiene
<b>Surrogates:</b>
Bromofluorobenzene

Report Date: 23-Mar-2007 08:11

## Air Toxics Ltd.

## AMBIENT AIR METHOD TO14A/TO15

Data file : /chem/msd8.i/8-22mar.b/8032216.d  
 Lab Smp Id: LCS-1 Client Smp ID: LCS-1  
 Inj Date : 22-MAR-2007 18:19  
 Operator : ej Inst ID: msd8.i  
 Smp Info : 50mL #1408-408  
 Misc Info : 200ppbv->50ppbv  
 Comment :  
 Method : /chem/msd8.i/8-22mar.b/t14q322a.m  
 Meth Date : 23-Mar-2007 08:11 sscott Quant Type: ISTD  
 Cal Date : 22-MAR-2007 13:30 Cal File: 8032209.d  
 Als bottle: 1 QC Sample: LCS  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: AT04+ENS.sub  
 Target Version: 3.50 Sample Matrix: AIR  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* CpndVariable

Cpnd Variable Local Compound Variable

CONCENTRATIONS								
		ON-COL	FINAL					
RT	EXP RT (REL RT)	MASS	RESPONSE ( PPBV)	( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====		
-----								
* 67	Bromochloromethane				CAS #: 74-97-5			
8.368	8.395 (1.000)	130	278522	25.0000	70.00-	130.00	100.00	
8.368	8.395 (1.000)	128	223686		46.35-	106.35	80.31	
8.368	8.395 (1.000)	49	840124		255.78-	315.78	301.64	
-----								
* 86	1,4-Difluorobenzene				CAS #: 540-36-3			
10.248	10.248 (1.000)	114	1372474	25.0000	70.00-	130.00	100.00	
10.248	10.248 (1.000)	88	244699		0.00-	48.88	17.83	
-----								
* 123	Chlorobenzene-d5				CAS #: 3114-55-4			
15.225	15.224 (1.000)	117	1036564	25.0000	70.00-	130.00	100.00	
15.197	15.224 (1.000)	82	689191		36.68-	96.68	66.49	
-----								
\$ 80	1,2-Dichloroethane-d4				CAS #: 17060-07-0			
9.446	9.473 (1.129)	65	612394	25.4366	25.436	70.00-	130.00	100.00
9.446	9.473 (1.129)	67	348350		27.92-	87.92	56.88	
-----								
\$ 102	Toluene-d8				CAS #: 2037-26-5			
12.985	12.985 (1.267)	98	1319513	24.4472	24.447	70.00-	130.00	100.00
12.985	12.985 (1.267)	70	167973		0.00-	42.61	12.73	

CONCENTRATIONS

ON-COL FINAL

RT	EXP RT	(REL RT)	MASS	RESPONSE	( PPEV)	( PPBV)	TARGET RANGE	RATIO
==	=====	=====	=====	=====	=====	=====	=====	=====

\$ 102 Toluene-d8 (continued)

12.985	12.985	(1.267)	100	974319			40.27- 100.27	73.84
--------	--------	---------	-----	--------	--	--	---------------	-------

\$ 138 Bromofluorobenzene

CAS #: 460-00-4

16.745	16.745	(1.100)	174	512983	24.1486	24.149	70.00- 130.00	100.00
16.745	16.745	(1.100)	95	994391			137.66- 197.66	193.84
16.745	16.745	(1.100)	176	500594			63.61- 123.61	97.58

3 Propylene

CAS #: 115-07-1

2.395	2.395	(0.286)	41	1876430	63.6731	63.673	70.00- 130.00	100.00
2.395	2.395	(0.286)	42	1242992			35.20- 95.20	66.24
2.395	2.395	(0.286)	39	1349895			42.80- 102.80	71.94

4 Dichlorodifluoromethane/Fr12

CAS #: 75-71-8

2.451	2.450	(0.293)	85	3599527	57.7849	57.785	70.00- 130.00	100.00
2.451	2.450	(0.293)	87	1162658			1.89- 61.89	32.30

6 Freon 114

CAS #: 76-14-2

2.589	2.616	(0.309)	135	1892311	53.5618	53.562	70.00- 130.00	100.00
2.589	2.616	(0.309)	137	615847			1.88- 61.88	32.54

8 Chloromethane

CAS #: 74-87-3

2.727	2.727	(0.326)	50	1778031	55.4796	55.480	70.00- 130.00	100.00
2.727	2.727	(0.326)	52	545218			0.00- 59.46	30.66

9 Butane

CAS #: 106-97-8

2.810	2.837	(0.336)	58	444668	53.3149	53.315	70.00- 130.00	100.00
2.810	2.837	(0.336)	43	3526100			772.78- 832.78	792.97

10 Vinyl Chloride

CAS #: 75-01-4

2.893	2.893	(0.346)	62	1691484	56.1894	56.189	70.00- 130.00	100.00
2.893	2.893	(0.346)	64	497353			0.33- 60.33	29.40

11 1,3-Butadiene

CAS #: 106-99-0

2.893	2.920	(0.346)	54	1615957	51.7656	51.766	70.00- 130.00	100.00
2.893	2.920	(0.346)	39	1828365			86.60- 146.60	113.14

13 Bromomethane

CAS #: 74-83-9

3.418	3.446	(0.408)	94	924272	56.1665	56.166	70.00- 130.00	100.00
3.418	3.446	(0.408)	96	853544			64.77- 124.77	92.35

15 Isopentane

CAS #: 78-78-4

3.529	3.556	(0.422)	43	2709546	52.6071	52.607	70.00- 130.00	100.00
3.529	3.556	(0.422)	57	1581385			27.51- 87.51	58.36
3.529	3.556	(0.422)	72	171194			0.00- 35.95	6.32

CONCENTRATIONS

ON-COL FINAL

RT EXP RT (REL RT) MASS RESPONSE ( PPBV) ( PPBV) TARGET RANGE RATIO  
 == == ===== == ===== ===== =====

16 Chloroethane CAS #: 75-00-3  
 3.557 3.612 (0.425) 64 887054 56.5863 56.586 70.00- 130.00 100.00  
 3.557 3.612 (0.425) 49 272964 2.42- 62.42 30.77  
 3.557 3.612 (0.425) 66 261670 0.00- 58.92 29.50

18 Trichlorofluoromethane/Fr11 CAS #: 75-69-4  
 3.888 3.916 (0.465) 101 2978271 53.3780 53.378 70.00- 130.00 100.00  
 3.888 3.916 (0.465) 103 1900616 34.71- 94.71 63.82

21 Ethanol CAS #: 64-17-5  
 4.275 4.331 (0.511) 45 661304 52.2111 52.211 70.00- 130.00 100.00  
 4.275 4.331 (0.511) 43 132468 0.00- 50.43 20.03  
 4.248 4.331 (0.508) 46 280779 12.21- 72.21 42.46

27 Freon 113 CAS #: 76-13-1  
 4.718 4.745 (0.564) 151 1414844 53.5490 53.549 70.00- 130.00 100.00  
 4.718 4.745 (0.564) 153 905176 32.67- 92.67 63.98  
 4.718 4.745 (0.564) 101 2268866 124.83- 184.83 160.36

29 1,1-Dichloroethene CAS #: 75-35-4  
 4.773 4.801 (0.570) 61 2448396 54.4406 54.440 70.00- 130.00 100.00  
 4.773 4.801 (0.570) 96 1111548 15.10- 75.10 45.40  
 4.773 4.801 (0.570) 98 688548 0.00- 58.17 28.12

30 Acetone CAS #: 67-64-1  
 4.939 4.939 (0.590) 58 832492 51.7567 51.757 70.00- 130.00 100.00  
 4.939 4.939 (0.590) 43 3109773 337.35- 397.35 373.55

33 Carbon Disulfide CAS #: 75-15-0  
 5.133 5.160 (0.613) 76 3456368 49.8938 49.894 70.00- 130.00 100.00

34 2-Propanol CAS #: 67-63-0  
 5.133 5.160 (0.613) 45 3220225 51.1002 51.100 70.00- 130.00 100.00  
 5.133 5.160 (0.613) 43 622351 0.00- 50.64 19.33  
 5.133 5.160 (0.613) 59 109991 0.00- 33.21 3.42

37 3-Chloropropene CAS #: 107-05-1  
 5.437 5.437 (0.650) 76 582413 52.0618 52.062 70.00- 130.00 100.00  
 5.437 5.437 (0.650) 41 2574669 414.53- 474.53 442.07

39 Methylene Chloride CAS #: 75-09-2  
 5.686 5.713 (0.679) 49 1928181 50.4149 50.415 70.00- 130.00 100.00  
 5.686 5.713 (0.679) 84 959383 18.44- 78.44 49.76  
 5.686 5.713 (0.679) 51 576250 0.00- 59.93 29.89

42 MTBE CAS #: 1634-04-4  
 6.045 6.045 (0.722) 73 3752786 48.6776 48.678 70.00- 130.00 100.00



CONCENTRATIONS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	ON-COL ( PPEV)	FINAL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
42 MTBE (continued)									
6.045	6.045	(0.722)	57	1080934			0.00- 59.33	28.80	
6.045	6.045	(0.722)	41	1183273			1.95- 61.95	31.53	
-----									
43 trans-1,2-Dichloroethene					CAS #: 156-60-5				
6.100	6.100	(0.729)	96	1090503	47.4477	47.448	70.00- 130.00	100.00	
6.073	6.100	(0.726)	61	2105343			169.97- 229.97	193.06	
6.100	6.100	(0.729)	98	673150			34.71- 94.71	61.73	
-----									
45 Hexane					CAS #: 110-54-3				
6.432	6.432	(0.769)	57	2673909	47.2636	47.264	70.00- 130.00	100.00	
6.432	6.432	(0.769)	43	1951618			44.61- 104.61	72.99	
6.432	6.432	(0.769)	86	377558			0.00- 43.77	14.12	
-----									
52 1,1-Dichloroethane					CAS #: 75-34-3				
6.874	6.902	(0.822)	63	2404587	50.3224	50.322	70.00- 130.00	100.00	
6.874	6.902	(0.822)	65	709306			0.00- 59.58	29.50	
-----									
54 Vinyl Acetate					CAS #: 108-05-4				
6.957	6.957	(0.831)	86	303994	53.4836	53.484	70.00- 130.00	100.00	
6.957	6.957	(0.831)	43	4738804			1469.62-1529.62	1558.85	
6.957	6.957	(0.831)	42	407364			100.08- 160.08	134.00	
-----									
63 cis-1,2-Dichloroethene					CAS #: 156-59-2				
7.925	7.953	(0.947)	61	1766344	48.5859	48.586	70.00- 130.00	100.00	
7.925	7.953	(0.947)	96	996178			27.36- 87.36	56.40	
7.925	7.953	(0.947)	98	603006			5.83- 65.83	34.14	
-----									
64 2-Butanone					CAS #: 78-93-3				
7.980	7.980	(0.954)	72	564657	43.2924	43.292	70.00- 130.00	100.00	
7.980	7.980	(0.954)	43	3472194			574.29- 634.29	614.92	
7.980	7.980	(0.954)	57	225366			10.61- 70.61	39.91	
-----									
66 Tetrahydrofuran					CAS #: 109-99-9				
8.368	8.367	(1.000)	42	2017236	41.5387	41.539	70.00- 130.00	100.00	
8.368	8.367	(1.000)	71	508113			0.00- 54.93	25.19	
8.368	8.367	(1.000)	72	558622			0.00- 56.52	27.69	
-----									
69 Chloroform					CAS #: 67-66-3				
8.506	8.506	(1.017)	83	1925190	42.0342	42.034	70.00- 130.00	100.00	
8.506	8.506	(1.017)	85	1250807			33.71- 93.71	64.97	
-----									
72 Cyclohexane					CAS #: 110-82-7				
8.727	8.727	(1.043)	84	1538360	45.0602	45.060	70.00- 130.00	100.00	
8.727	8.727	(1.043)	56	2601694			138.68- 198.68	169.12	
8.727	8.727	(1.043)	41	1546180			73.81- 133.81	100.51	
-----									

CONCENTRATIONS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	ON-COL ( PPEV)	FINAL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
-----										
73	1,1,1-Trichloroethane					CAS #:	71-55-6			
8.755	8.754	(1.046)	97	1902106	45.9990	45.999	70.00-	130.00	100.00	
8.755	8.754	(1.046)	99	1246434			33.95-	93.95	65.53	
-----										
75	Carbon Tetrachloride					CAS #:	56-23-5			
8.976	9.003	(1.073)	119	1575718	47.1772	47.177	70.00-	130.00	100.00	
8.976	9.003	(1.073)	117	1744559			81.67-	141.67	110.72	
-----										
78	2,2,4-Trimethylpentane					CAS #:	540-84-1			
9.418	9.446	(1.126)	57	7315932	45.9263	45.926	70.00-	130.00	100.00	
9.418	9.446	(1.126)	56	2562625			4.45-	64.45	35.03	
9.418	9.446	(1.126)	41	2159630			0.00-	59.44	29.52	
-----										
79	Benzene					CAS #:	71-43-2			
9.418	9.418	(0.919)	78	3340948	43.7066	43.706	70.00-	130.00	100.00	
9.418	9.418	(0.919)	77	774792			0.00-	54.39	23.19	
-----										
81	1,2-Dichloroethane					CAS #:	107-06-2			
9.612	9.612	(0.938)	62	1669206	46.6289	46.629	70.00-	130.00	100.00	
9.612	9.612	(0.938)	64	507016			1.44-	61.44	30.37	
-----										
82	Heptane					CAS #:	142-82-5			
9.805	9.805	(0.957)	100	377747	41.3701	41.370	70.00-	130.00	100.00	
9.805	9.805	(0.957)	43	3122573			789.66-	849.66	826.63	
9.805	9.805	(0.957)	71	1222459			294.27-	354.27	323.62	
-----										
92	Trichloroethene					CAS #:	79-01-6			
10.662	10.662	(1.040)	95	1155008	46.9757	46.976	70.00-	130.00	100.00	
10.662	10.662	(1.040)	130	960049			57.13-	117.13	83.12	
10.662	10.662	(1.040)	97	737325			32.50-	92.50	63.84	
-----										
93	Methyl Cyclohexane					CAS #:	108-87-2			
10.884	10.883	(1.301)	83	1922839	45.0279	45.028	70.00-	130.00	100.00	
10.884	10.883	(1.301)	98	908713			15.00-	75.00	47.26	
10.884	10.883	(1.301)	55	2207231			87.18-	147.18	114.79	
-----										
95	1,2-Dichloropropane					CAS #:	78-87-5			
11.188	11.188	(1.092)	63	1250358	44.6373	44.637	70.00-	130.00	100.00	
11.188	11.188	(1.092)	62	934234			45.03-	105.03	74.72	
11.188	11.188	(1.092)	41	962531			45.52-	105.52	76.98	
-----										
96	1,4-Dioxane					CAS #:	123-91-1			
11.409	11.409	(1.113)	88	645482	44.2024	44.202	70.00-	130.00	100.00	
11.409	11.409	(1.113)	58	619887			70.27-	130.27	96.03	
11.409	11.409	(1.113)	57	198043			0.54-	60.54	30.68	
-----										

CONCENTRATIONS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	ON-COL ( PPEV)	FINAL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
-----									
98 Bromodichloromethane						CAS #: 75-27-4			
11.713	11.741	(1.143)	83	1880847	45.4089	45.409	70.00- 130.00	100.00	
11.713	11.741	(1.143)	85	1201006			32.38- 92.38	63.85	
-----									
100 cis-1,3-Dichloropropene						CAS #: 10061-01-5			
12.598	12.625	(1.229)	75	1654311	46.8298	46.830	70.00- 130.00	100.00	
12.598	12.625	(1.229)	77	523109			1.52- 61.52	31.62	
12.598	12.625	(1.229)	39	1241394			48.26- 108.26	75.04	
-----									
101 4-Methyl-2-pentanone						CAS #: 108-10-1			
12.902	12.902	(1.259)	58	1256304	44.1080	44.108	70.00- 130.00	100.00	
12.902	12.902	(1.259)	43	3592445			251.14- 311.14	285.95	
12.902	12.902	(1.259)	85	387104			2.33- 62.33	30.81	
-----									
103 Toluene						CAS #: 108-88-3			
13.096	13.095	(1.278)	91	3428164	49.5296	49.530	70.00- 130.00	100.00	
13.096	13.095	(1.278)	92	2053740			29.45- 89.45	59.91	
-----									
106 trans-1,3-Dichloropropene						CAS #: 10061-02-6			
13.621	13.621	(0.895)	75	1733104	47.9456	47.946	70.00- 130.00	100.00	
13.621	13.621	(0.895)	77	544880			1.42- 61.42	31.44	
13.621	13.621	(0.895)	39	1268743			43.57- 103.57	73.21	
-----									
108 1,1,2-Trichloroethane						CAS #: 79-00-5			
13.870	13.870	(0.911)	97	953253	45.2979	45.298	70.00- 130.00	100.00	
13.870	13.870	(0.911)	99	587815			31.40- 91.40	61.66	
13.870	13.870	(0.911)	83	899793			60.06- 120.06	94.39	
-----									
109 Tetrachloroethene						CAS #: 127-18-4			
13.925	13.925	(0.915)	166	1075635	48.1240	48.124	70.00- 130.00	100.00	
13.925	13.925	(0.915)	129	859308			46.36- 106.36	79.89	
13.925	13.925	(0.915)	131	824537			45.00- 105.00	76.66	
-----									
112 2-Hexanone						CAS #: 591-78-6			
14.257	14.257	(0.936)	58	1547930	46.6060	46.606	70.00- 130.00	100.00	
14.257	14.257	(0.936)	43	3177061			176.52- 236.52	205.25	
14.257	14.257	(0.936)	100	239749			0.00- 45.68	15.49	
-----									
114 Dibromochloromethane						CAS #: 124-48-1			
14.423	14.423	(0.947)	129	1351154	47.3140	47.314	70.00- 130.00	100.00	
14.423	14.423	(0.947)	127	1074573			46.96- 106.96	79.53	
-----									
115 1,2-Dibromoethane						CAS #: 106-93-4			
14.589	14.588	(0.958)	107	1439089	48.7359	48.736	70.00- 130.00	100.00	
14.589	14.588	(0.958)	109	1320391			62.88- 122.88	91.75	
-----									

CONCENTRATIONS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	( PPEV)	FINAL	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
124 Chlorobenzene						CAS #: 108-90-7				
15.252	15.252	(1.002)	112	2364677	48.5581	48.558	70.00- 130.00	100.00		
15.252	15.252	(1.002)	114	723136			0.73- 60.73	30.58		
15.252	15.252	(1.002)	77	1645570			40.58- 100.58	69.59		
-----										
127 Ethyl Benzene						CAS #: 100-41-4				
15.363	15.363	(1.009)	106	1311932	47.9853	47.985	70.00- 130.00	100.00		
15.363	15.363	(1.009)	91	4511338			309.90- 369.90	343.87		
-----										
128 m,p-Xylene						CAS #: 108-38-3				
15.529	15.529	(1.020)	106	1672409	47.5696	47.570	70.00- 130.00	100.00		
15.529	15.529	(1.020)	91	3565698			177.38- 237.38	213.21		
-----										
130 o-Xylene						CAS #: 95-47-6				
16.054	16.054	(1.054)	106	1625053	47.9353	47.935	70.00- 130.00	100.00		
16.054	16.054	(1.054)	91	3693663			195.99- 255.99	227.29		
-----										
131 Styrene						CAS #: 100-42-5				
16.082	16.082	(1.056)	104	2458659	47.1155	47.115	70.00- 130.00	100.00		
16.082	16.082	(1.056)	78	1484068			28.78- 88.78	60.36		
-----										
133 Bromoform						CAS #: 75-25-2				
16.358	16.358	(1.074)	173	1107122	46.9144	46.914	70.00- 130.00	100.00		
16.358	16.358	(1.074)	171	566742			21.17- 81.17	51.19		
-----										
135 Cumene						CAS #: 98-82-8				
16.524	16.524	(1.085)	105	4785876	46.2609	46.261	70.00- 130.00	100.00		
16.524	16.524	(1.085)	120	1083750			0.00- 53.14	22.64		
16.496	16.524	(1.084)	51	737073			0.00- 45.39	15.40		
-----										
142 1,1,2,2-Tetrachloroethane						CAS #: 79-34-5				
16.966	16.966	(1.114)	83	2314942	47.6739	47.674	70.00- 130.00	100.00		
16.966	16.966	(1.114)	85	1479700			34.35- 94.35	63.92		
-----										
143 Propylbenzene						CAS #: 103-65-1				
16.994	16.994	(1.116)	91	6112597	52.2536	52.254	70.00- 130.00	100.00		
16.994	16.994	(1.116)	120	1134482			0.00- 48.79	18.56		
16.994	16.994	(1.116)	105	209533			0.00- 33.35	3.43		
-----										
145 4-Ethyltoluene						CAS #: 622-96-8				
17.132	17.132	(1.125)	105	5307532	52.2076	52.208	70.00- 130.00	100.00		
17.132	17.132	(1.125)	120	1325326			0.00- 54.57	24.97		
-----										
146 1,3,5-Trimethylbenzene						CAS #: 108-67-8				
17.215	17.215	(1.131)	105	4361594	45.3301	45.330	70.00- 130.00	100.00		
17.215	17.215	(1.131)	120	1879411			12.64- 72.64	43.09		
-----										

CONCENTRATIONS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CONCENTRATIONS		TARGET RANGE	RATIO	
					ON-COL ( PPEV)	FINAL ( PPBV)			
==	=====	=====	=====	=====	=====	=====	=====	=====	=====
-----									
151	1,2,4-Trimethylbenzene					CAS #: 95-63-6			
17.602	17.602	(1.156)	105	4025096	46.1987	46.199	70.00-	130.00	100.00
17.602	17.602	(1.156)	120	1666459			11.41-	71.41	41.40
-----									
154	1,3-Dichlorobenzene					CAS #: 541-73-1			
17.907	17.906	(1.176)	146	1763512	46.6011	46.601	70.00-	130.00	100.00
17.907	17.906	(1.176)	148	1109821			32.20-	92.20	62.93
17.879	17.906	(1.174)	111	831855			13.28-	73.28	47.17
-----									
155	1,4-Dichlorobenzene					CAS #: 106-46-7			
17.990	17.989	(1.182)	146	2412413	49.9911	49.991	70.00-	130.00	100.00
17.990	17.989	(1.182)	148	1546218			34.16-	94.16	64.09
17.990	17.989	(1.182)	111	1088761			16.41-	76.41	45.13
-----									
156	alpha-Chlorotoluene					CAS #: 100-44-7			
18.128	18.128	(1.191)	91	3406668	48.7592	48.759	70.00-	130.00	100.00
18.128	18.128	(1.191)	126	594643			0.00-	46.97	17.46
-----									
158	1,2-Dichlorobenzene					CAS #: 95-50-1			
18.321	18.321	(1.203)	146	1870369	45.0339	45.034	70.00-	130.00	100.00
18.321	18.321	(1.203)	148	1179854			33.83-	93.83	63.08
18.321	18.321	(1.203)	111	1039079			21.29-	81.29	55.55
-----									
163	1,2,4-Trichlorobenzene					CAS #: 120-82-1			
19.593	19.593	(1.287)	180	1227104	47.4461	47.446	70.00-	130.00	100.00
19.593	19.593	(1.287)	182	1149823			62.06-	122.06	93.70
-----									
164	Hexachlorobutadiene					CAS #: 87-68-3			
19.676	19.676	(1.292)	225	1408481	42.6344	42.634	70.00-	130.00	100.00
19.676	19.676	(1.292)	223	915348			35.92-	95.92	64.99
-----									
165	Naphthalene					CAS #: 91-20-3			
19.787	19.787	(1.300)	128	5855301	70.8537	70.854	70.00-	130.00	100.00(R)
19.787	19.787	(1.300)	127	815433			0.00-	43.72	13.93
-----									

QC Flag Legend

R - Spike/Surrogate failed recovery limits.

Report Date: 23-Mar-2007 08:11

Air Toxics Ltd.

INTERNAL STANDARD COMPOUNDS  
AREA AND RT SUMMARY

Instrument ID: msd8.i

Calibration Date: 22-MAR-2007

Lab File ID: 8032216.d

Calibration Time: 12:34

Lab Smp Id: LCS-1

Client Smp ID: LCS-1

Analysis Type: VOA

Level: LOW

Quant Type: ISTD

Sample Type: AIR

Operator: ej

Method File: /chem/msd8.i/8-22mar.b/t14q322a.m

Misc Info: 200ppbv-&gt;50ppbv

COMPOUND	STANDARD	AREA LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	283735	170241	397229	278522	-1.84
86 1,4-Difluorobenze	1370859	822515	1919203	1372474	0.12
123 Chlorobenzene-d5	1067063	640238	1493888	1036564	-2.86

COMPOUND	STANDARD	RT LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	8.39	8.06	8.72	8.37	-0.33
86 1,4-Difluorobenze	10.25	9.92	10.58	10.25	0.00
123 Chlorobenzene-d5	15.22	14.89	15.55	15.22	0.00

AREA UPPER LIMIT = + 40% of internal standard area.

AREA LOWER LIMIT = - 40% of internal standard area.

RT UPPER LIMIT = + 0.33 minutes of internal standard RT.

RT LOWER LIMIT = - 0.33 minutes of internal standard RT.

Air Toxics Ltd.

RECOVERY REPORT

Client Name: Client SDG: 8-22mar  
 Sample Matrix: GAS Fraction: VOA  
 Lab Smp Id: LCS-1 Client Smp ID: LCS-1  
 Level: LOW Operator: ej  
 Data Type: MS DATA SampleType: LCS  
 SpikeList File: Spectra+ENS.spk Quant Type: ISTD  
 Sublist File: AT04+ENS.sub  
 Method File: /chem/msd8.i/8-22mar.b/t14q322a.m  
 Misc Info: 200ppbv->50ppbv


SPIKE COMPOUND	CONC ADDED PPBV	CONC RECOVERED PPBV	% RECOVERED	LIMITS
131 Styrene	50.000	47.115	94.23	70-130
106 trans-1,3-Dichloro	50.000	47.946	95.89	70-130
4 Dichlorodifluorome	50.000	57.785	115.57	70-130
6 Freon 114	50.000	53.562	107.12	70-130
8 Chloromethane	50.000	55.480	110.96	70-130
10 Vinyl Chloride	50.000	56.189	112.38	70-130
11 1,3-Butadiene	50.000	51.766	103.53	60-140
13 Bromomethane	50.000	56.166	112.33	70-130
16 Chloroethane	50.000	56.586	113.17	70-130
18 Trichlorofluoromet	50.000	53.378	106.76	70-130
21 Ethanol	50.000	52.211	104.42	60-140
27 Freon 113	50.000	53.549	107.10	70-130
29 1,1-Dichloroethene	50.000	54.440	108.88	70-130
30 Acetone	50.000	51.757	103.51	60-140
33 Carbon Disulfide	50.000	49.894	99.79	60-140
34 2-Propanol	50.000	51.100	102.20	60-140
39 Methylene Chloride	50.000	50.415	100.83	70-130
42 MTBE	50.000	48.678	97.36	60-140
43 trans-1,2-Dichloro	50.000	47.448	94.90	60-140
45 Hexane	50.000	47.264	94.53	60-140
52 1,1-Dichloroethane	50.000	50.322	100.64	70-130
63 cis-1,2-Dichloroet	50.000	48.586	97.17	70-130
64 2-Butanone	50.000	43.292	86.58	60-140
66 Tetrahydrofuran	50.000	41.539	83.08	60-140
69 Chloroform	50.000	42.034	84.07	70-130
72 Cyclohexane	50.000	45.060	90.12	60-140
73 1,1,1-Trichloroeth	50.000	45.999	92.00	70-130
75 Carbon Tetrachlori	50.000	47.177	94.35	70-130
79 Benzene	50.000	43.706	87.41	70-130
81 1,2-Dichloroethane	50.000	46.629	93.26	70-130
82 Heptane	50.000	41.370	82.74	60-140
92 Trichloroethene	50.000	46.976	93.95	70-130
95 1,2-Dichloropropan	50.000	44.637	89.27	70-130

Report Date: 23-Mar-2007 08:11

SPIKE COMPOUND	CONC ADDED PPBV	CONC RECOVERED PPBV	% RECOVERED	LIMITS
96 1,4-Dioxane	50.000	44.202	88.40	60-140
98 Bromodichlorometha	50.000	45.409	90.82	60-140
100 cis-1,3-Dichloropr	50.000	46.830	93.66	70-130
101 4-Methyl-2-pentano	50.000	44.108	88.22	60-140
103 Toluene	50.000	49.530	99.06	70-130
108 1,1,2-Trichloroeth	50.000	45.298	90.60	70-130
109 Tetrachloroethene	50.000	48.124	96.25	70-130
112 2-Hexanone	50.000	46.606	93.21	60-140
114 Dibromochlorometha	50.000	47.314	94.63	60-140
115 1,2-Dibromoethane	50.000	48.736	97.47	70-130
124 Chlorobenzene	50.000	48.558	97.12	70-130
127 Ethyl Benzene	50.000	47.985	95.97	70-130
128 m,p-Xylene	50.000	47.570	95.14	70-130
130 o-Xylene	50.000	47.935	95.87	70-130
133 Bromoform	50.000	46.914	93.83	60-140
142 1,1,2,2-Tetrachlor	50.000	47.674	95.35	70-130
145 4-Ethyltoluene	50.000	52.208	104.42	60-140
146 1,3,5-Trimethylben	50.000	45.330	90.66	70-130
151 1,2,4-Trimethylben	50.000	46.199	92.40	70-130
154 1,3-Dichlorobenzen	50.000	46.601	93.20	70-130
155 1,4-Dichlorobenzen	50.000	49.991	99.98	70-130
156 alpha-Chlorotoluen	50.000	48.759	97.52	70-130
158 1,2-Dichlorobenzen	50.000	45.034	90.07	70-130
163 1,2,4-Trichloroben	50.000	47.446	94.89	70-130
164 Hexachlorobutadien	50.000	42.634	85.27	70-130
135 Cumene	50.000	46.261	92.52	60-140
143 Propylbenzene	50.000	52.254	104.51	60-140
37 3-Chloropropene	50.000	52.062	104.12	60-140
78 2,2,4-Trimethylpen	50.000	45.926	91.85	60-140
9 Butane	50.000	53.315	106.63	70-130
15 Isopentane	50.000	52.607	105.21	70-130
93 Methyl Cyclohexane	50.000	45.028	90.06	70-130
165 Naphthalene	50.000	70.854	141.71*	60-140

SURROGATE COMPOUND	CONC ADDED PPBV	CONC RECOVERED PPBV	% RECOVERED	LIMITS
\$ 80 1,2-Dichloroethane	25.000	25.436	101.75	70-130
\$ 102 Toluene-d8	25.000	24.447	97.79	70-130
\$ 138 Bromofluorobenzene	25.000	24.149	96.59	70-130





Report Date: 23-Mar-2007 08:08

## Air Toxics Ltd.

## AMBIENT AIR METHOD TO14A/TO15

Data file : /chem/msd8.i/8-22mar.b/8032203.d  
 Lab Smp Id: ICAL Client Smp ID: Level 1  
 Inj Date : 22-MAR-2007 10:41  
 Operator : sjr Inst ID: msd8.i  
 Smp Info : 0.2ml 1487-115  
 Misc Info : 200ppbv -> 0.2ppbv  
 Comment :  
 Method : /chem/msd8.i/8-22mar.b/t14q322a.m  
 Meth Date : 23-Mar-2007 08:08 sscott Quant Type: ISTD  
 Cal Date : 22-MAR-2007 10:41 Cal File: 8032203.d  
 Als bottle: 1 Calibration Sample, Level: 1  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: AFCEElow.sub  
 Target Version: 3.50 Sample Matrix: AIR  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* CpndVariable

Cpnd Variable Local Compound Variable

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	( PPBV)	( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
* 67 Bromochloromethane CAS #: 74-97-5									
8.395	8.395	(1.000)	130	285447	25.0000		70.00- 130.00	100.00	
8.395	8.395	(1.000)	128	222762			46.35- 106.35	78.04	
8.395	8.395	(1.000)	49	841026			255.78- 315.78	294.63	
-----									
* 86 1,4-Difluorobenzene CAS #: 540-36-3									
10.248	10.248	(1.000)	114	1360758	25.0000		70.00- 130.00	100.00	
10.248	10.248	(1.000)	88	259943			0.00- 48.88	19.10	
-----									
* 123 Chlorobenzene-d5 CAS #: 3114-55-4									
15.224	15.224	(1.000)	117	1082018	25.0000		70.00- 130.00	100.00	
15.224	15.224	(1.000)	82	726655			0.00- 30.00	67.16	
-----									
\$ 80 1,2-Dichloroethane-d4 CAS #: 17060-07-0									
9.473	9.473	(1.128)	65	587414	25.0000	25.018	70.00- 130.00	100.00	
9.473	9.473	(1.128)	67	292096			0.00- 30.00	49.73	
-----									
\$ 102 Toluene-d8 CAS #: 2037-26-5									
12.985	12.985	(1.267)	98	1356224	25.0000	25.230	70.00- 130.00	100.00	
12.985	12.985	(1.267)	70	172342			0.00- 30.00	12.71	

AMOUNTS											
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO			
==	=====	=====	=====	=====	=====	=====	=====	=====			
-----											
\$ 102 Toluene-d8 (continued)											
12.985	12.985	(1.267)	100	890926			0.00- 30.00	65.69			
-----											
\$ 138 Bromofluorobenzene											
						CAS #: 460-00-4					
16.773	16.773	(1.102)	174	538970	25.0000	23.876	70.00- 130.00	100.00			
16.745	16.745	(1.100)	95	902384			137.66- 197.66	167.43			
16.773	16.773	(1.102)	176	528806			63.61- 123.61	98.11			
-----											
69 Chloroform											
						CAS #: 67-66-3					
8.506	8.506	(1.013)	83	14108	0.20000	0.2628	70.00- 130.00	100.00(a)			
8.533	8.533	(1.016)	85	9243			33.71- 93.71	65.52			
-----											
79 Benzene											
						CAS #: 71-43-2					
9.418	9.418	(0.919)	78	22113	0.20000	0.2542	70.00- 130.00	100.00(a)			
9.418	9.418	(0.919)	77	8130			0.00- 30.00	36.77			
-----											
131 Styrene											
						CAS #: 100-42-5					
16.109	16.109	(1.058)	104	14656	0.20000	0.2412	70.00- 130.00	100.00(a)			
16.082	16.082	(1.056)	78	11399			28.78- 88.78	77.78			
-----											
135 Cumene											
						CAS #: 98-82-8					
16.524	16.524	(1.085)	105	29921	0.20000	0.2471	70.00- 130.00	100.00(a)			
16.524	16.524	(1.085)	120	6949			0.00- 30.00	23.22			
16.496	16.496	(1.084)	51	6483			0.00- 30.00	21.67			
-----											

QC Flag Legend

a - Target compound detected but, quantitated amount  
 Below Limit Of Quantitation(BLOQ).

Report Date: 23-Mar-2007 08:08

Air Toxics Ltd.

INTERNAL STANDARD COMPOUNDS  
AREA AND RT SUMMARY

Instrument ID: msd8.i

Calibration Date: 22-MAR-2007

Lab File ID: 8032203.d

Calibration Time: 12:34

Lab Smp Id: ICAL

Client Smp ID: Level 1

Analysis Type: VOA

Level: LOW

Quant Type: ISTD

Sample Type: AIR

Operator: sjr

Method File: /chem/msd8.i/8-22mar.b/t14q322a.m

Misc Info: 200ppbv -&gt; 0.2ppbv

COMPOUND	STANDARD	AREA LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	283735	170241	397229	285447	0.60
86 1,4-Difluorobenze	1370859	822515	1919203	1360758	-0.74
123 Chlorobenzene-d5	1067063	640238	1493888	1082018	1.40

COMPOUND	STANDARD	RT LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	8.39	8.06	8.72	8.40	0.00
86 1,4-Difluorobenze	10.25	9.92	10.58	10.25	0.00
123 Chlorobenzene-d5	15.22	14.89	15.55	15.22	0.00

AREA UPPER LIMIT = + 40% of internal standard area.

AREA LOWER LIMIT = - 40% of internal standard area.

RT UPPER LIMIT = + 0.33 minutes of internal standard RT.

RT LOWER LIMIT = - 0.33 minutes of internal standard RT.

Data File: /chem/msd8.1/8-22mar.1b/8032203.d

Date : 22-MAR-2007 10:41

Client ID: Level 1


Sample Info: 0.2ml 1487-115

Column phase: RTX-624

Instrument: msd8.1

Operator: sjr

Column diameter: 0.53



Report Date: 23-Mar-2007 08:09

## Air Toxics Ltd.

## AMBIENT AIR METHOD TO14A/TO15

Data file : /chem/msd8.i/8-22mar.b/8032215.d  
 Lab Smp Id: ICAL Client Smp ID: Level 2  
 Inj Date : 22-MAR-2007 17:30  
 Operator : sjr Inst ID: msd8.i  
 Smp Info : 0.5ml 1487-115  
 Misc Info : 200ppbv -> 0.5ppbv  
 Comment :  
 Method : /chem/msd8.i/8-22mar.b/t14q322a.m  
 Meth Date : 23-Mar-2007 08:09 sscott Quant Type: ISTD  
 Cal Date : 22-MAR-2007 17:30 Cal File: 8032215.d  
 Als bottle: 1 Calibration Sample, Level: 2  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: ICALlevel2.sub  
 Target Version: 3.50 Sample Matrix: AIR  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* CpndVariable

Cpnd Variable Local Compound Variable

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	( PPBV)	ON-COL	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
* 67 Bromochloromethane CAS #: 74-97-5									
8.395	8.395	(1.000)	130	270806	25.0000		70.00- 130.00	100.00	
8.395	8.395	(1.000)	128	205077			46.35- 106.35	75.73	
8.367	8.367	(1.000)	49	794498			255.78- 315.78	293.38	
-----									
* 86 1,4-Difluorobenzene CAS #: 540-36-3									
10.248	10.248	(1.000)	114	1311826	25.0000		70.00- 130.00	100.00	
10.248	10.248	(1.000)	88	240488			0.00- 48.88	18.33	
-----									
* 123 Chlorobenzene-d5 CAS #: 3114-55-4									
15.225	15.225	(1.000)	117	1039819	25.0000		70.00- 130.00	100.00	
15.197	15.197	(1.000)	82	701224			0.00- 30.00	67.44	
-----									
\$ 80 1,2-Dichloroethane-d4 CAS #: 17060-07-0									
9.473	9.473	(1.128)	65	564438	25.0000	25.225	70.00- 130.00	100.00	
9.473	9.473	(1.128)	67	303193			0.00- 30.00	53.72	
-----									
\$ 102 Toluene-d8 CAS #: 2037-26-5									
12.985	12.985	(1.267)	98	1271754	25.0000	24.692	70.00- 130.00	100.00	
12.985	12.985	(1.267)	70	162520			0.00- 30.00	12.78	

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
\$ 102 Toluene-d8 (continued)									
12.985	12.985	(1.267)	100	881454			0.00- 30.00	69.31	
-----									
\$ 138 Bromofluorobenzene CAS #: 460-00-4									
16.773	16.773	(1.102)	174	487769	25.0000	23.265	70.00- 130.00	100.00	
16.745	16.745	(1.100)	95	897698			137.66- 197.66	184.04	
16.773	16.773	(1.102)	176	495471			63.61- 123.61	101.58	
-----									
4 Dichlorodifluoromethane/Fr12 CAS #: 75-71-8									
2.450	2.450	(0.292)	85	42030	0.50000	0.6249	70.00- 130.00	100.00	
2.478	2.478	(0.295)	87	13507			0.00- 30.00	32.14	
-----									
6 Freon 114 CAS #: 76-14-2									
2.616	2.616	(0.312)	135	22788	0.50000	0.6009	70.00- 130.00	100.00	
2.644	2.644	(0.315)	137	8952			1.88- 61.88	39.28	
-----									
10 Vinyl Chloride CAS #: 75-01-4									
2.921	2.921	(0.348)	62	16055	0.50000	0.5515	70.00- 130.00	100.00	
2.893	2.893	(0.345)	64	2845			0.00- 30.00	17.72	
-----									
11 1,3-Butadiene CAS #: 106-99-0									
2.893	2.893	(0.345)	54	18194	0.50000	0.5819	70.00- 130.00	100.00	
2.893	2.893	(0.345)	39	71814			0.00- 30.00	394.71	
-----									
13 Bromomethane CAS #: 74-83-9									
3.446	3.446	(0.410)	94	11062	0.50000	0.6255	70.00- 130.00	100.00	
3.418	3.418	(0.407)	96	21718			64.77- 124.77	196.33	
-----									
16 Chloroethane CAS #: 75-00-3									
3.584	3.584	(0.427)	64	9404	0.50000	0.5914	70.00- 130.00	100.00	
3.612	3.612	(0.430)	49	4248			0.00- 30.00	45.17	
3.695	3.695	(0.440)	66	2350			0.00- 30.00	24.99	
-----									
18 Trichlorofluoromethane/Fr11 CAS #: 75-69-4									
3.916	3.916	(0.466)	101	29095	0.50000	0.5434	70.00- 130.00	100.00	
3.916	3.916	(0.466)	103	24541			34.71- 94.71	84.35	
-----									
27 Freon 113 CAS #: 76-13-1									
4.745	4.745	(0.565)	151	15286	0.50000	0.5812	70.00- 130.00	100.00	
4.745	4.745	(0.565)	153	10476			32.67- 92.67	68.53	
4.745	4.745	(0.565)	101	25478			124.83- 184.83	166.68	
-----									
29 1,1-Dichloroethene CAS #: 75-35-4									
4.773	4.773	(0.569)	61	28480	0.50000	0.6054	70.00- 130.00	100.00	
4.801	4.801	(0.572)	96	16319			15.10- 75.10	57.30	
4.773	4.773	(0.569)	98	11574			0.00- 58.17	40.64	
-----									

AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
-----										
33	Carbon Disulfide					CAS #:	75-15-0			
5.160	5.160	(0.615)	76	43108	0.50000	0.5958	70.00-	130.00	100.00	
-----										
39	Methylene Chloride					CAS #:	75-09-2			
5.713	5.713	(0.681)	49	24563	0.50000	0.6098	70.00-	130.00	100.00	
5.713	5.713	(0.681)	84	13379			18.44-	78.44	54.47	
5.713	5.713	(0.681)	51	6981			0.00-	30.00	28.42	
-----										
42	MTBE					CAS #:	1634-04-4			
6.045	6.045	(0.720)	73	47973	0.50000	0.5988	70.00-	130.00	100.00	
6.045	6.045	(0.720)	57	13535			0.00-	59.33	28.21	
6.045	6.045	(0.720)	41	17986			0.00-	30.00	37.49	
-----										
43	trans-1,2-Dichloroethene					CAS #:	156-60-5			
6.100	6.100	(0.727)	96	15974	0.50000	0.6337	70.00-	130.00	100.00	
6.100	6.100	(0.727)	61	28116			169.97-	229.97	176.01	
6.100	6.100	(0.727)	98	12293			0.00-	30.00	76.96	
-----										
45	Hexane					CAS #:	110-54-3			
6.460	6.460	(0.769)	57	35245	0.50000	0.5968	70.00-	130.00	100.00	
6.432	6.432	(0.766)	43	24536			0.00-	30.00	69.62	
6.460	6.460	(0.769)	86	5992			0.00-	30.00	17.00	
-----										
52	1,1-Dichloroethane					CAS #:	75-34-3			
6.902	6.902	(0.822)	63	28073	0.50000	0.5719	70.00-	130.00	100.00	
6.902	6.902	(0.822)	65	12707			0.00-	59.58	45.26	
-----										
64	2-Butanone					CAS #:	78-93-3			
7.980	7.980	(0.951)	72	9064	0.50000	0.6310	70.00-	130.00	100.00	
7.980	7.980	(0.951)	43	37052			574.29-	634.29	408.78	
7.980	7.980	(0.951)	57	3317			0.00-	30.00	36.60	
-----										
63	cis-1,2-Dichloroethene					CAS #:	156-59-2			
7.953	7.953	(0.947)	61	19760	0.50000	0.5612	70.00-	130.00	100.00	
7.953	7.953	(0.947)	96	10181			27.36-	87.36	51.52	
7.953	7.953	(0.947)	98	11221			5.83-	65.83	56.79	
-----										
66	Tetrahydrofuran					CAS #:	109-99-9			
8.367	8.367	(0.997)	42	36347	0.50000	0.6537	70.00-	130.00	100.00	
8.395	8.395	(1.000)	71	10429			0.00-	54.93	28.69	
8.367	8.367	(0.997)	72	9696			0.00-	30.00	26.68	
-----										
69	Chloroform					CAS #:	67-66-3			
8.506	8.506	(1.013)	83	27459	0.50000	0.5254	70.00-	130.00	100.00	
8.533	8.533	(1.016)	85	14750			33.71-	93.71	53.72	
-----										



AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
-----										
73	1,1,1-Trichloroethane					CAS #:	71-55-6			
8.755	8.755	(1.043)	97	25097	0.50000	0.5909	70.00-	130.00	100.00	
8.755	8.755	(1.043)	99	15727			33.95-	93.95	62.66	
-----										
72	Cyclohexane					CAS #:	110-82-7			
8.727	8.727	(1.040)	84	21555	0.50000	0.6076	70.00-	130.00	100.00	
8.727	8.727	(1.040)	56	31282			138.68-	198.68	145.13	
8.727	8.727	(1.040)	41	25266			73.81-	133.81	117.22	
-----										
75	Carbon Tetrachloride					CAS #:	56-23-5			
9.003	9.003	(1.072)	119	18169	0.50000	0.5543	70.00-	130.00	100.00	
9.003	9.003	(1.072)	117	20156			81.67-	141.67	110.94	
-----										
79	Benzene					CAS #:	71-43-2			
9.418	9.418	(0.919)	78	38732	0.50000	0.4739	70.00-	130.00	100.00(a)	
9.418	9.418	(0.919)	77	13594			0.00-	30.00	35.10	
-----										
81	1,2-Dichloroethane					CAS #:	107-06-2			
9.612	9.612	(0.938)	62	22054	0.50000	0.5953	70.00-	130.00	100.00	
9.612	9.612	(0.938)	64	7025			0.00-	30.00	31.85	
-----										
82	Heptane					CAS #:	142-82-5			
9.805	9.805	(0.957)	100	7424	0.50000	0.6746	70.00-	130.00	100.00	
9.805	9.805	(0.957)	43	52953			0.00-	30.00	713.27	
9.833	9.833	(0.960)	71	18647			0.00-	30.00	251.17	
-----										
92	Trichloroethene					CAS #:	79-01-6			
10.662	10.662	(1.040)	95	12618	0.50000	0.5430	70.00-	130.00	100.00	
10.690	10.690	(1.043)	130	13346			57.13-	117.13	105.77	
10.662	10.662	(1.040)	97	8061			32.50-	92.50	63.88	
-----										
95	1,2-Dichloropropane					CAS #:	78-87-5			
11.188	11.188	(1.092)	63	17718	0.50000	0.6000	70.00-	130.00	100.00	
11.188	11.188	(1.092)	62	12940			45.03-	105.03	73.03	
11.188	11.188	(1.092)	41	12462			45.52-	105.52	70.34	
-----										
98	Bromodichloromethane					CAS #:	75-27-4			
11.741	11.741	(1.146)	83	23415	0.50000	0.5696	70.00-	130.00	100.00	
11.741	11.741	(1.146)	85	15224			32.38-	92.38	65.02	
-----										
100	cis-1,3-Dichloropropene					CAS #:	10061-01-5			
12.625	12.625	(1.232)	75	20173	0.50000	0.5704	70.00-	130.00	100.00	
12.625	12.625	(1.232)	77	7291			1.52-	61.52	36.14	
12.625	12.625	(1.232)	39	11695			48.26-	108.26	57.97	
-----										
101	4-Methyl-2-pentanone					CAS #:	108-10-1			
12.902	12.902	(1.259)	58	17822	0.50000	0.5955	70.00-	130.00	100.00	

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
101 4-Methyl-2-pentanone (continued)									
12.902	12.902	(1.259)	43	51793			0.00- 30.00	290.61	
12.902	12.902	(1.259)	85	9743			0.00- 30.00	54.67	
-----									
103 Toluene CAS #: 108-88-3									
13.096	13.096	(1.278)	91	35164	0.50000	0.5344	70.00- 130.00	100.00	
13.096	13.096	(1.278)	92	25326			29.45- 89.45	72.02	
-----									
106 trans-1,3-Dichloropropene CAS #: 10061-02-6									
13.621	13.621	(0.895)	75	23242	0.50000	0.5860	70.00- 130.00	100.00	
13.621	13.621	(0.895)	77	8854			1.42- 61.42	38.09	
13.621	13.621	(0.895)	39	14308			43.57- 103.57	61.56	
-----									
108 1,1,2-Trichloroethane CAS #: 79-00-5									
13.897	13.897	(0.913)	97	13582	0.50000	0.5878	70.00- 130.00	100.00	
13.870	13.870	(0.911)	99	8275			31.40- 91.40	60.93	
13.870	13.870	(0.911)	83	13152			60.06- 120.06	96.83	
-----									
109 Tetrachloroethene CAS #: 127-18-4									
13.925	13.925	(0.915)	166	14217	0.50000	0.5766	70.00- 130.00	100.00	
13.925	13.925	(0.915)	129	11924			46.36- 106.36	83.87	
13.925	13.925	(0.915)	131	8940			45.00- 105.00	62.88	
-----									
112 2-Hexanone CAS #: 591-78-6									
14.257	14.257	(0.936)	58	20545	0.50000	0.6288	70.00- 130.00	100.00	
14.257	14.257	(0.936)	43	38619			176.52- 236.52	187.97	
14.257	14.257	(0.936)	100	6148			0.00- 30.00	29.92	
-----									
114 Dibromochloromethane CAS #: 124-48-1									
14.423	14.423	(0.947)	129	17348	0.50000	0.5700	70.00- 130.00	100.00	
14.423	14.423	(0.947)	127	15277			0.00- 30.00	88.06	
-----									
115 1,2-Dibromoethane CAS #: 106-93-4									
14.589	14.589	(0.958)	107	15949	0.50000	0.5388	70.00- 130.00	100.00	
14.589	14.589	(0.958)	109	17920			62.88- 122.88	112.36	
-----									
124 Chlorobenzene CAS #: 108-90-7									
15.252	15.252	(1.002)	112	29163	0.50000	0.5604	70.00- 130.00	100.00	
15.252	15.252	(1.002)	114	13422			0.73- 60.73	46.02	
15.252	15.252	(1.002)	77	29566			40.58- 100.58	101.38	
-----									
127 Ethyl Benzene CAS #: 100-41-4									
15.363	15.363	(1.009)	106	17066	0.50000	0.5755	70.00- 130.00	100.00	
15.363	15.363	(1.009)	91	51748			0.00- 30.00	303.22	
-----									
128 m,p-Xylene CAS #: 108-38-3									
15.529	15.529	(1.020)	106	21671	0.50000	0.5707	70.00- 130.00	100.00	

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT (PPEV)	ON-COL (PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
128 m,p-Xylene (continued)									
15.529	15.529	(1.020)	91	45416			0.00- 30.00	209.57	
-----									
130 o-Xylene CAS #: 95-47-6									
16.054	16.054	(1.054)	106	20158	0.50000	0.5650	70.00- 130.00	100.00	
16.054	16.054	(1.054)	91	42807			195.99- 255.99	212.36	
-----									
131 Styrene CAS #: 100-42-5									
16.109	16.109	(1.058)	104	23849	0.50000	0.4350	70.00- 130.00	100.00(a)	
16.082	16.082	(1.056)	78	14289			28.78- 88.78	59.91	
-----									
133 Bromoform CAS #: 75-25-2									
16.358	16.358	(1.074)	173	12843	0.50000	0.5272	70.00- 130.00	100.00	
16.358	16.358	(1.074)	171	8122			21.17- 81.17	63.24	
-----									
142 1,1,2,2-Tetrachloroethane CAS #: 79-34-5									
16.966	16.966	(1.114)	83	28519	0.50000	0.5624	70.00- 130.00	100.00	
16.966	16.966	(1.114)	85	16875			34.35- 94.35	59.17	
-----									
145 4-Ethyltoluene CAS #: 622-96-8									
17.132	17.132	(1.125)	105	60065	0.50000	0.5485	70.00- 130.00	100.00	
17.132	17.132	(1.125)	120	14682			0.00- 54.57	24.44	
-----									
146 1,3,5-Trimethylbenzene CAS #: 108-67-8									
17.215	17.215	(1.131)	105	56714	0.50000	0.5707	70.00- 130.00	100.00	
17.215	17.215	(1.131)	120	22159			0.00- 30.00	39.07	
-----									
151 1,2,4-Trimethylbenzene CAS #: 95-63-6									
17.602	17.602	(1.156)	105	53686	0.50000	0.5797	70.00- 130.00	100.00	
17.602	17.602	(1.156)	120	22150			11.41- 71.41	41.26	
-----									
154 1,3-Dichlorobenzene CAS #: 541-73-1									
17.907	17.907	(1.176)	146	22328	0.50000	0.5493	70.00- 130.00	100.00	
17.907	17.907	(1.176)	148	16082			0.00- 30.00	72.03	
17.879	17.879	(1.174)	111	10660			0.00- 30.00	47.74	
-----									
155 1,4-Dichlorobenzene CAS #: 106-46-7									
17.989	17.989	(1.182)	146	29954	0.50000	0.5706	70.00- 130.00	100.00	
17.989	17.989	(1.182)	148	13724			0.00- 30.00	45.82	
17.989	17.989	(1.182)	111	11705			0.00- 30.00	39.08	
-----									
156 alpha-Chlorotoluene CAS #: 100-44-7									
18.128	18.128	(1.191)	91	38235	0.50000	0.5268	70.00- 130.00	100.00	
18.128	18.128	(1.191)	126	7448			0.00- 30.00	19.48	
-----									
158 1,2-Dichlorobenzene CAS #: 95-50-1									
18.321	18.321	(1.203)	146	26326	0.50000	0.5785	70.00- 130.00	100.00	

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
158 1,2-Dichlorobenzene (continued)									
18.321	18.321	(1.203)	148	16880			33.83- 93.83	64.12	
18.321	18.321	(1.203)	111	14335			21.29- 81.29	54.45	
-----									
135 Cumene CAS #: 98-82-8									
16.524	16.524	(1.085)	105	55515	0.50000	0.4844	70.00- 130.00	100.00(a)	
16.524	16.524	(1.085)	120	14849			0.00- 30.00	26.75	
16.496	16.496	(1.084)	51	10327			0.00- 30.00	18.60	
-----									
143 Propylbenzene CAS #: 103-65-1									
16.994	16.994	(1.116)	91	68703	0.50000	0.5556	70.00- 130.00	100.00	
16.994	16.994	(1.116)	120	14050			0.00- 30.00	20.45	
16.994	16.994	(1.116)	105	6002			0.00- 30.00	8.74	
-----									
78 2,2,4-Trimethylpentane CAS #: 540-84-1									
9.418	9.418	(1.122)	57	89003	0.50000	0.5639	70.00- 130.00	100.00	
9.418	9.418	(1.122)	56	33063			0.00- 30.00	37.15	
9.418	9.418	(1.122)	41	31452			0.00- 30.00	35.34	
-----									
93 Methyl Cyclohexane CAS #: 108-87-2									
10.911	10.911	(1.300)	83	26434	0.50000	0.5964	70.00- 130.00	100.00	
10.884	10.884	(1.296)	98	14042			0.00- 30.00	53.12	
10.884	10.884	(1.296)	55	27351			0.00- 30.00	103.47	
-----									

QC Flag Legend

a - Target compound detected but, quantitated amount  
Below Limit Of Quantitation(BLOQ).

Report Date: 23-Mar-2007 08:09

Air Toxics Ltd.

INTERNAL STANDARD COMPOUNDS  
AREA AND RT SUMMARY

Instrument ID: msd8.i

Calibration Date: 22-MAR-2007

Lab File ID: 8032215.d

Calibration Time: 12:34

Lab Smp Id: ICAL

Client Smp ID: Level 2

Analysis Type: VOA

Level: LOW

Quant Type: ISTD

Sample Type: AIR

Operator: sjr

Method File: /chem/msd8.i/8-22mar.b/t14q322a.m

Misc Info: 200ppbv -&gt; 0.5ppbv

COMPOUND	STANDARD	AREA LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	283735	170241	397229	270806	-4.56
86 1,4-Difluorobenze	1370859	822515	1919203	1311826	-4.31
123 Chlorobenzene-d5	1067063	640238	1493888	1039819	-2.55

COMPOUND	STANDARD	RT LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	8.39	8.06	8.72	8.40	0.00
86 1,4-Difluorobenze	10.25	9.92	10.58	10.25	0.00
123 Chlorobenzene-d5	15.22	14.89	15.55	15.22	0.00

AREA UPPER LIMIT = + 40% of internal standard area.

AREA LOWER LIMIT = - 40% of internal standard area.

RT UPPER LIMIT = + 0.33 minutes of internal standard RT.

RT LOWER LIMIT = - 0.33 minutes of internal standard RT.

Data File: /chem/msd8.1/8-22mar.lb/8032215.d

Date: 22-MAR-2007 17:30

Client ID: Level 2


Sample Info: 0.5ml 1487-115

Column phase: RTX-624

Instrument: msd8.1

Operator: sjr

Column diameter: 0.53



Report Date: 26-Mar-2007 13:58

## Air Toxics Ltd.

## AMBIENT AIR METHOD TO14A/TO15

Data file : /chem/msd8.i/8-26mar.b/8032606.d  
 Lab Smp Id: ICAL Client Smp ID: LEVEL 3  
 Inj Date : 26-MAR-2007 12:15  
 Operator : ea Inst ID: msd8.i  
 Smp Info : 2.0ml #1487-42  
 Misc Info : 200ppbv-2.0ppbv  
 Comment :  
 Method : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Meth Date : 26-Mar-2007 13:58 ealcan Quant Type: ISTD  
 Cal Date : 26-MAR-2007 12:15 Cal File: 8032606.d  
 Als bottle: 1 Calibration Sample, Level: 3  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: sp5b.sub  
 Target Version: 3.50 Sample Matrix: AIR  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* CpndVariable

Cpnd Variable Local Compound Variable

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	( PPBV)	ON-COL	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
* 67 Bromochloromethane CAS #: 74-97-5									
8.395	8.395	(1.000)	130	293727	25.0000		70.00- 130.00	100.00	
8.395	8.395	(1.000)	128	225479			46.88- 106.88	76.76	
8.367	8.367	(1.000)	49	828524			254.92- 314.92	282.07	
-----									
* 86 1,4-Difluorobenzene CAS #: 540-36-3									
10.248	10.248	(1.000)	114	1372405	25.0000		70.00- 130.00	100.00	
10.248	10.248	(1.000)	88	259807			0.00- 48.81	18.93	
-----									
* 123 Chlorobenzene-d5 CAS #: 3114-55-4									
15.224	15.224	(1.000)	117	1065309	25.0000		70.00- 130.00	100.00	
15.224	15.224	(1.000)	82	720886			0.00- 30.00	67.67	
-----									
194 2-Methylpentane CAS #: 107-83-5									
5.547	5.547	(0.661)	71	57426	2.00000	2.417	70.00- 130.00	100.00	
5.547	5.547	(0.661)	43	218120			0.00- 30.00	379.83	
5.547	5.547	(0.661)	42	109426			0.00- 30.00	190.55	
-----									
195 Thiopene CAS #: 110-02-1									
9.833	9.833	(0.960)	84	108102	2.00000	2.401	70.00- 130.00	100.00(T)	

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	====	=====	=====	=====	=====	=====	
195 Thiopene (continued)									
9.833	9.833	(0.960)	58	79477			0.00- 30.00	73.52	
0.000	1.000	(0.000)	0	0			0.00- 30.00	0.00	
-----									
196 Indan									
						CAS #: 496-11-7			
18.183	18.183	(1.194)	117	227148	2.00000	2.439	70.00- 130.00	100.00	
18.183	18.183	(1.194)	118	124147			0.00- 30.00	54.65	
18.183	18.183	(1.194)	91	53358			0.00- 30.00	23.49	
-----									
197 Indene									
						CAS #: 95-13-6			
18.404	18.404	(1.209)	115	148838	2.00000	2.457	70.00- 130.00	100.00(T)	
0.000	1.000	(0.000)	16	0			0.00- 30.00	0.00	
-----									
83 2,3-Dimethylpentane									
						CAS #: 565-59-3			
8.782	8.782	(1.046)	71	45393	2.00000	2.356	70.00- 130.00	100.00	
8.810	8.810	(1.049)	56	153558			314.83- 374.83	338.29	
8.782	8.782	(1.046)	43	145616			0.00- 30.00	320.79	
-----									

QC Flag Legend

T - Target compound detected outside RT window.



Report Date: 26-Mar-2007 13:58

Air Toxics Ltd.

INTERNAL STANDARD COMPOUNDS  
AREA AND RT SUMMARY

Instrument ID: msd8.i

Calibration Date: 26-MAR-2007

Lab File ID: 8032606.d

Calibration Time: 12:43

Lab Smp Id: ICAL

Client Smp ID: LEVEL 3

Analysis Type: VOA

Level: LOW

Quant Type: ISTD

Sample Type: AIR

Operator: ea

Method File: /chem/msd8.i/8-26mar.b/t14q322b.m

Misc Info: 200ppbv-2.0ppbv

COMPOUND	STANDARD	AREA LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	291821	175093	408549	293727	0.65
86 1,4-Difluorobenze	1354208	812525	1895891	1372405	1.34
123 Chlorobenzene-d5	1071895	643137	1500653	1065309	-0.61

COMPOUND	STANDARD	RT LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	8.40	8.07	8.73	8.39	0.00
86 1,4-Difluorobenze	10.25	9.92	10.58	10.25	0.00
123 Chlorobenzene-d5	15.22	14.89	15.55	15.22	0.00

AREA UPPER LIMIT = + 40% of internal standard area.

AREA LOWER LIMIT = - 40% of internal standard area.

RT UPPER LIMIT = + 0.33 minutes of internal standard RT.

RT LOWER LIMIT = - 0.33 minutes of internal standard RT.

Data File: /chem/msd8.1/8-26mar.lb/8032606.d

Date: 26-MAR-2007 12:15

Client ID: LEVEL 3


Sample Info: 2.0ml #1487-42

Column phase: RTX-624

Instrument: msd8.1

Operator: ea

Column diameter: 0.53



Report Date: 23-Mar-2007 08:09

## Air Toxics Ltd.

## AMBIENT AIR METHOD TO14A/TO15

Data file : /chem/msd8.i/8-22mar.b/8032205.d  
 Lab Smp Id: ICAL Client Smp ID: Level 3  
 Inj Date : 22-MAR-2007 11:38  
 Operator : sjr Inst ID: msd8.i  
 Smp Info : 2.0ml 1487-115  
 Misc Info : 200ppbv -> 2ppbv  
 Comment :  
 Method : /chem/msd8.i/8-22mar.b/t14q322a.m  
 Meth Date : 23-Mar-2007 08:09 sscott Quant Type: ISTD  
 Cal Date : 22-MAR-2007 11:38 Cal File: 8032205.d  
 Als bottle: 1 Calibration Sample, Level: 3  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: AT04mdl+Na+ENSR.sub  
 Target Version: 3.50 Sample Matrix: AIR  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* CpndVariable

Cpnd Variable Local Compound Variable

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	( PPBV)	CAL-AMT	ON-COL	TARGET RANGE	RATIO
==	=====	=====	=====	=====	=====	=====	=====	=====	=====
* 67 Bromochloromethane CAS #: 74-97-5									
8.395	8.395	(1.000)	130	275416	25.0000			70.00- 130.00	100.00
8.395	8.395	(1.000)	128	214844				46.35- 106.35	78.01
8.395	8.395	(1.000)	49	843588				255.78- 315.78	306.30
-----									
* 86 1,4-Difluorobenzene CAS #: 540-36-3									
10.248	10.248	(1.000)	114	1361295	25.0000			70.00- 130.00	100.00
10.248	10.248	(1.000)	88	247284				0.00- 48.88	18.17
-----									
* 123 Chlorobenzene-d5 CAS #: 3114-55-4									
15.224	15.224	(1.000)	117	1067969	25.0000			70.00- 130.00	100.00
15.224	15.224	(1.000)	82	699917				0.00- 30.00	65.54
-----									
\$ 80 1,2-Dichloroethane-d4 CAS #: 17060-07-0									
9.473	9.473	(1.128)	65	591430	25.0000	25.734		70.00- 130.00	100.00
9.473	9.473	(1.128)	67	299310				0.00- 30.00	50.61
-----									
\$ 102 Toluene-d8 CAS #: 2037-26-5									
12.985	12.985	(1.267)	98	1293142	25.0000	24.391		70.00- 130.00	100.00
12.985	12.985	(1.267)	70	177809				0.00- 30.00	13.75

AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
-----										
\$ 102 Toluene-d8 (continued)										
12.985	12.985	(1.267)	100	845124			0.00- 30.00	65.35		
-----										
\$ 138 Bromofluorobenzene										
						CAS #: 460-00-4				
16.773	16.773	(1.102)	174	543659	25.0000	25.185	70.00- 130.00	100.00		
16.745	16.745	(1.100)	95	861291			137.66- 197.66	158.42		
16.773	16.773	(1.102)	176	504513			63.61- 123.61	92.80		
-----										
3 Propylene										
						CAS #: 115-07-1				
2.395	2.395	(0.285)	41	80391	2.00000	2.431	70.00- 130.00	100.00		
2.395	2.395	(0.285)	42	56887			0.00- 30.00	70.76		
2.395	2.395	(0.285)	39	58801			0.00- 30.00	73.14		
-----										
4 Dichlorodifluoromethane/Fr12										
						CAS #: 75-71-8				
2.450	2.450	(0.292)	85	150641	2.00000	2.130	70.00- 130.00	100.00		
2.450	2.450	(0.292)	87	52547			0.00- 30.00	34.88		
-----										
6 Freon 114										
						CAS #: 76-14-2				
2.616	2.616	(0.312)	135	80893	2.00000	2.064	70.00- 130.00	100.00(H)		
2.616	2.616	(0.312)	137	25279			1.88- 61.88	31.25		
-----										
8 Chloromethane										
						CAS #: 74-87-3				
2.755	2.755	(0.328)	50	94717	2.00000	2.511	70.00- 130.00	100.00		
2.755	2.755	(0.328)	52	32017			0.00- 30.00	33.80		
-----										
10 Vinyl Chloride										
						CAS #: 75-01-4				
2.920	2.920	(0.348)	62	77330	2.00000	2.370	70.00- 130.00	100.00		
2.893	2.893	(0.345)	64	28295			0.00- 30.00	36.59		
-----										
11 1,3-Butadiene										
						CAS #: 106-99-0				
2.893	2.893	(0.345)	54	81680	2.00000	2.346	70.00- 130.00	100.00		
2.893	2.893	(0.345)	39	193417			0.00- 30.00	236.80		
-----										
13 Bromomethane										
						CAS #: 74-83-9				
3.446	3.446	(0.410)	94	37760	2.00000	2.065	70.00- 130.00	100.00		
3.446	3.446	(0.410)	96	36806			64.77- 124.77	97.47		
-----										
16 Chloroethane										
						CAS #: 75-00-3				
3.612	3.612	(0.430)	64	38142	2.00000	2.226	70.00- 130.00	100.00		
3.612	3.612	(0.430)	49	14286			0.00- 30.00	37.45		
3.612	3.612	(0.430)	66	9687			0.00- 30.00	25.40		
-----										
18 Trichlorofluoromethane/Fr11										
						CAS #: 75-69-4				
3.916	3.916	(0.466)	101	139609	2.00000	2.343	70.00- 130.00	100.00		
3.916	3.916	(0.466)	103	94957			34.71- 94.71	68.02		
-----										

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
21 Ethanol						CAS #: 64-17-5			
4.275	4.275	(0.509)	45	32514	2.00000	2.322	70.00- 130.00	100.00	
4.248	4.248	(0.506)	43	13134			0.00- 30.00	40.39	
4.248	4.248	(0.506)	46	17713			0.00- 30.00	54.48	
-----									
27 Freon 113						CAS #: 76-13-1			
4.745	4.745	(0.565)	151	70996	2.00000	2.393	70.00- 130.00	100.00	
4.745	4.745	(0.565)	153	39185			32.67- 92.67	55.19	
4.745	4.745	(0.565)	101	96260			124.83- 184.83	135.59	
-----									
29 1,1-Dichloroethene						CAS #: 75-35-4			
4.801	4.801	(0.572)	61	111438	2.00000	2.208	70.00- 130.00	100.00	
4.801	4.801	(0.572)	96	54692			15.10- 75.10	49.08	
4.801	4.801	(0.572)	98	30302			0.00- 58.17	27.19	
-----									
30 Acetone						CAS #: 67-64-1			
4.939	4.939	(0.588)	58	45330	2.00000	2.490	70.00- 130.00	100.00	
4.939	4.939	(0.588)	43	136965			0.00- 30.00	302.15	
-----									
34 2-Propanol						CAS #: 67-63-0			
5.160	5.160	(0.615)	45	151092	2.00000	2.257	70.00- 130.00	100.00	
5.160	5.160	(0.615)	43	47357			0.00- 30.00	31.34	
5.160	5.160	(0.615)	59	7137			0.00- 30.00	4.72	
-----									
33 Carbon Disulfide						CAS #: 75-15-0			
5.160	5.160	(0.615)	76	169984	2.00000	2.197	70.00- 130.00	100.00	
-----									
39 Methylene Chloride						CAS #: 75-09-2			
5.713	5.713	(0.681)	49	94236	2.00000	2.191	70.00- 130.00	100.00	
5.713	5.713	(0.681)	84	43931			18.44- 78.44	46.62	
5.713	5.713	(0.681)	51	30460			0.00- 30.00	32.32	
-----									
42 MTBE						CAS #: 1634-04-4			
6.045	6.045	(0.720)	73	188996	2.00000	2.202	70.00- 130.00	100.00	
6.045	6.045	(0.720)	57	57183			0.00- 59.33	30.26	
6.045	6.045	(0.720)	41	64052			0.00- 30.00	33.89	
-----									
43 trans-1,2-Dichloroethene						CAS #: 156-60-5			
6.100	6.100	(0.727)	96	56019	2.00000	2.120	70.00- 130.00	100.00	
6.100	6.100	(0.727)	61	104489			169.97- 229.97	186.52	
6.100	6.100	(0.727)	98	37274			0.00- 30.00	66.54	
-----									
45 Hexane						CAS #: 110-54-3			
6.460	6.460	(0.769)	57	138775	2.00000	2.197	70.00- 130.00	100.00	
6.432	6.432	(0.766)	43	105650			0.00- 30.00	76.13	
6.460	6.460	(0.769)	86	17560			0.00- 30.00	12.65	
-----									

AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
54 Vinyl Acetate										
						CAS #:	108-05-4			
6.957	6.957	(0.829)	86	11360	2.00000	1.985	70.00-	130.00	100.00(a)	
6.985	6.985	(0.832)	43	222873			0.00-	30.00	1961.91	
6.985	6.985	(0.832)	42	30814			0.00-	30.00	271.25	
-----										
52 1,1-Dichloroethane										
						CAS #:	75-34-3			
6.902	6.902	(0.822)	63	112495	2.00000	2.162	70.00-	130.00	100.00	
6.902	6.902	(0.822)	65	38326			0.00-	59.58	34.07	
-----										
64 2-Butanone										
						CAS #:	78-93-3			
8.008	8.008	(0.954)	72	31076	2.00000	2.083	70.00-	130.00	100.00	
7.980	7.980	(0.951)	43	189499			574.29-	634.29	609.79	
8.008	8.008	(0.954)	57	13265			0.00-	30.00	42.69	
-----										
63 cis-1,2-Dichloroethene										
						CAS #:	156-59-2			
7.953	7.953	(0.947)	61	93793	2.00000	2.374	70.00-	130.00	100.00	
7.953	7.953	(0.947)	96	53841			27.36-	87.36	57.40	
7.953	7.953	(0.947)	98	35782			5.83-	65.83	38.15	
-----										
66 Tetrahydrofuran										
						CAS #:	109-99-9			
8.367	8.367	(0.997)	42	112088	2.00000	1.988	70.00-	130.00	100.00	
8.367	8.367	(0.997)	71	31794			0.00-	54.93	28.37	
8.367	8.367	(0.997)	72	36828			0.00-	30.00	32.86	
-----										
69 Chloroform										
						CAS #:	67-66-3			
8.506	8.506	(1.013)	83	99138	2.00000	1.897	70.00-	130.00	100.00	
8.506	8.506	(1.013)	85	60557			33.71-	93.71	61.08	
-----										
73 1,1,1-Trichloroethane										
						CAS #:	71-55-6			
8.754	8.754	(1.043)	97	102556	2.00000	2.235	70.00-	130.00	100.00	
8.754	8.754	(1.043)	99	65946			33.95-	93.95	64.30	
-----										
72 Cyclohexane										
						CAS #:	110-82-7			
8.727	8.727	(1.040)	84	88061	2.00000	2.274	70.00-	130.00	100.00	
8.727	8.727	(1.040)	56	144374			138.68-	198.68	163.95	
8.727	8.727	(1.040)	41	83275			73.81-	133.81	94.57	
-----										
75 Carbon Tetrachloride										
						CAS #:	56-23-5			
9.003	9.003	(1.072)	119	83711	2.00000	2.314	70.00-	130.00	100.00	
9.003	9.003	(1.072)	117	90621			81.67-	141.67	108.25	
-----										
79 Benzene										
						CAS #:	71-43-2			
9.418	9.418	(0.919)	78	176144	2.00000	2.057	70.00-	130.00	100.00	
9.418	9.418	(0.919)	77	42211			0.00-	30.00	23.96	
-----										
81 1,2-Dichloroethane										
						CAS #:	107-06-2			
9.612	9.612	(0.938)	62	85388	2.00000	2.142	70.00-	130.00	100.00	

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	====	=====	=====	=====	=====	=====	
81 1,2-Dichloroethane (continued)									
9.612	9.612	(0.938)	64	28911			0.00- 30.00	33.86	
-----									
82 Heptane CAS #: 142-82-5									
9.805	9.805	(0.957)	100	19431	2.00000	1.790	70.00- 130.00	100.00	
9.805	9.805	(0.957)	43	173362			0.00- 30.00	892.19	
9.805	9.805	(0.957)	71	61608			0.00- 30.00	317.06	
-----									
92 Trichloroethene CAS #: 79-01-6									
10.662	10.662	(1.040)	95	65132	2.00000	2.418	70.00- 130.00	100.00	
10.690	10.690	(1.043)	130	44283			57.13- 117.13	67.99	
10.662	10.662	(1.040)	97	42276			32.50- 92.50	64.91	
-----									
95 1,2-Dichloropropane CAS #: 78-87-5									
11.188	11.188	(1.092)	63	66549	2.00000	2.111	70.00- 130.00	100.00	
11.188	11.188	(1.092)	62	52517			45.03- 105.03	78.91	
11.188	11.188	(1.092)	41	54992			45.52- 105.52	82.63	
-----									
96 1,4-Dioxane CAS #: 123-91-1									
11.409	11.409	(1.113)	88	37221	2.00000	2.314	70.00- 130.00	100.00	
11.409	11.409	(1.113)	58	35515			70.27- 130.27	95.42	
11.409	11.409	(1.113)	57	15135			0.00- 30.00	40.66	
-----									
98 Bromodichloromethane CAS #: 75-27-4									
11.741	11.741	(1.146)	83	102417	2.00000	2.250	70.00- 130.00	100.00	
11.741	11.741	(1.146)	85	67122			32.38- 92.38	65.54	
-----									
100 cis-1,3-Dichloropropene CAS #: 10061-01-5									
12.625	12.625	(1.232)	75	83591	2.00000	2.177	70.00- 130.00	100.00	
12.625	12.625	(1.232)	77	23874			1.52- 61.52	28.56	
12.625	12.625	(1.232)	39	64418			48.26- 108.26	77.06	
-----									
101 4-Methyl-2-pentanone CAS #: 108-10-1									
12.902	12.902	(1.259)	58	67980	2.00000	2.122	70.00- 130.00	100.00	
12.902	12.902	(1.259)	43	203570			0.00- 30.00	299.46	
12.902	12.902	(1.259)	85	22270			0.00- 30.00	32.76	
-----									
103 Toluene CAS #: 108-88-3									
13.095	13.095	(1.278)	91	171210	2.00000	2.312	70.00- 130.00	100.00	
13.095	13.095	(1.278)	92	93386			29.45- 89.45	54.54	
-----									
106 trans-1,3-Dichloropropene CAS #: 10061-02-6									
13.621	13.621	(0.895)	75	82643	2.00000	2.019	70.00- 130.00	100.00	
13.621	13.621	(0.895)	77	31993			1.42- 61.42	38.71	
13.621	13.621	(0.895)	39	78281			43.57- 103.57	94.72	
-----									

AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
-----										
108	1,1,2-Trichloroethane					CAS #:	79-00-5			
13.897	13.897	(0.913)	97	49580	2.00000	2.058	70.00-	130.00	100.00	
13.897	13.897	(0.913)	99	31406			31.40-	91.40	63.34	
13.897	13.897	(0.913)	83	46646			60.06-	120.06	94.08	
-----										
109	Tetrachloroethene					CAS #:	127-18-4			
13.925	13.925	(0.915)	166	48412	2.00000	1.940	70.00-	130.00	100.00	
13.925	13.925	(0.915)	129	47041			46.36-	106.36	97.17	
13.925	13.925	(0.915)	131	40366			45.00-	105.00	83.38	
-----										
112	2-Hexanone					CAS #:	591-78-6			
14.257	14.257	(0.936)	58	75259	2.00000	2.114	70.00-	130.00	100.00	
14.257	14.257	(0.936)	43	175210			176.52-	236.52	232.81	
14.257	14.257	(0.936)	100	13078			0.00-	30.00	17.38	
-----										
114	Dibromochloromethane					CAS #:	124-48-1			
14.423	14.423	(0.947)	129	69176	2.00000	2.137	70.00-	130.00	100.00	
14.423	14.423	(0.947)	127	50548			0.00-	30.00	73.07	
-----										
115	1,2-Dibromoethane					CAS #:	106-93-4			
14.589	14.589	(0.958)	107	75869	2.00000	2.305	70.00-	130.00	100.00	
14.589	14.589	(0.958)	109	73073			62.88-	122.88	96.31	
-----										
124	Chlorobenzene					CAS #:	108-90-7			
15.252	15.252	(1.002)	112	119945	2.00000	2.156	70.00-	130.00	100.00	
15.252	15.252	(1.002)	114	44278			0.73-	60.73	36.92	
15.252	15.252	(1.002)	77	94503			40.58-	100.58	78.79	
-----										
127	Ethyl Benzene					CAS #:	100-41-4			
15.363	15.363	(1.009)	106	63387	2.00000	2.053	70.00-	130.00	100.00	
15.363	15.363	(1.009)	91	227641			0.00-	30.00	359.13	
-----										
128	m,p-Xylene					CAS #:	108-38-3			
15.529	15.529	(1.020)	106	83062	2.00000	2.084	70.00-	130.00	100.00	
15.529	15.529	(1.020)	91	188273			0.00-	30.00	226.67	
-----										
130	o-Xylene					CAS #:	95-47-6			
16.054	16.054	(1.054)	106	86049	2.00000	2.219	70.00-	130.00	100.00	
16.054	16.054	(1.054)	91	194934			195.99-	255.99	226.54	
-----										
131	Styrene					CAS #:	100-42-5			
16.082	16.082	(1.056)	104	128028	2.00000	2.198	70.00-	130.00	100.00	
16.082	16.082	(1.056)	78	77317			28.78-	88.78	60.39	
-----										
133	Bromoform					CAS #:	75-25-2			
16.358	16.358	(1.074)	173	55198	2.00000	2.133	70.00-	130.00	100.00	



AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
133 Bromoform (continued)									
16.358	16.358	(1.074)	171	30668			21.17- 81.17	55.56	
-----									
142 1,1,2,2-Tetrachloroethane CAS #: 79-34-5									
16.966	16.966	(1.114)	83	121397	2.00000	2.209	70.00- 130.00	100.00	
16.966	16.966	(1.114)	85	81065			34.35- 94.35	66.78	
-----									
145 4-Ethyltoluene CAS #: 622-96-8									
17.132	17.132	(1.125)	105	244319	2.00000	2.112	70.00- 130.00	100.00	
17.132	17.132	(1.125)	120	57573			0.00- 54.57	23.56	
-----									
146 1,3,5-Trimethylbenzene CAS #: 108-67-8									
17.215	17.215	(1.131)	105	262265	2.00000	2.347	70.00- 130.00	100.00	
17.215	17.215	(1.131)	120	111702			0.00- 30.00	42.59	
-----									
151 1,2,4-Trimethylbenzene CAS #: 95-63-6									
17.602	17.602	(1.156)	105	220430	2.00000	2.201	70.00- 130.00	100.00	
17.602	17.602	(1.156)	120	92573			11.41- 71.41	42.00	
-----									
154 1,3-Dichlorobenzene CAS #: 541-73-1									
17.906	17.906	(1.176)	146	82808	2.00000	1.989	70.00- 130.00	100.00	
17.906	17.906	(1.176)	148	57275			0.00- 30.00	69.17	
17.906	17.906	(1.176)	111	36240			0.00- 30.00	43.76	
-----									
155 1,4-Dichlorobenzene CAS #: 106-46-7									
17.989	17.989	(1.182)	146	107737	2.00000	1.999	70.00- 130.00	100.00	
17.989	17.989	(1.182)	148	72586			0.00- 30.00	67.37	
17.989	17.989	(1.182)	111	44084			0.00- 30.00	40.92	
-----									
156 alpha-Chlorotoluene CAS #: 100-44-7									
18.128	18.128	(1.191)	91	148060	2.00000	1.991	70.00- 130.00	100.00	
18.128	18.128	(1.191)	126	30145			0.00- 30.00	20.36	
-----									
158 1,2-Dichlorobenzene CAS #: 95-50-1									
18.321	18.321	(1.203)	146	99085	2.00000	2.078	70.00- 130.00	100.00	
18.321	18.321	(1.203)	148	58889			33.83- 93.83	59.43	
18.321	18.321	(1.203)	111	46285			21.29- 81.29	46.71	
-----									
163 1,2,4-Trichlorobenzene CAS #: 120-82-1									
19.593	19.593	(1.287)	180	62263	2.00000	2.220	70.00- 130.00	100.00	
19.593	19.593	(1.287)	182	58921			62.06- 122.06	94.63	
-----									
164 Hexachlorobutadiene CAS #: 87-68-3									
19.676	19.676	(1.292)	225	104649	2.00000	2.558	70.00- 130.00	100.00	
19.676	19.676	(1.292)	223	57449			35.92- 95.92	54.90	
-----									

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
-----									
135 Cumene						CAS #: 98-82-8			
16.524	16.524	(1.085)	105	243116	2.00000	2.049	70.00- 130.00	100.00	
16.524	16.524	(1.085)	120	52203			0.00- 30.00	21.47	
16.496	16.496	(1.084)	51	38911			0.00- 30.00	16.01	
-----									
143 Propylbenzene						CAS #: 103-65-1			
16.994	16.994	(1.116)	91	287725	2.00000	2.169	70.00- 130.00	100.00	
16.994	16.994	(1.116)	120	52173			0.00- 30.00	18.13	
16.994	16.994	(1.116)	105	13051			0.00- 30.00	4.54	
-----									
37 3-Chloropropene						CAS #: 107-05-1			
5.437	5.437	(0.648)	76	29487	2.00000	2.368	70.00- 130.00	100.00	
5.437	5.437	(0.648)	41	139623			0.00- 30.00	473.51	
-----									
78 2,2,4-Trimethylpentane						CAS #: 540-84-1			
9.446	9.446	(1.125)	57	396404	2.00000	2.290	70.00- 130.00	100.00	
9.418	9.418	(1.122)	56	133818			0.00- 30.00	33.76	
9.418	9.418	(1.122)	41	121704			0.00- 30.00	30.70	
-----									
165 Naphthalene						CAS #: 91-20-3			
19.787	19.787	(1.300)	128	229919	2.00000	2.376	70.00- 130.00	100.00	
19.787	19.787	(1.300)	127	33934			0.00- 30.00	14.76	
-----									
9 Butane						CAS #: 106-97-8			
2.837	2.837	(0.338)	58	25473	2.00000	2.569	70.00- 130.00	100.00	
2.837	2.837	(0.338)	43	171107			0.00- 30.00	671.72	
-----									
15 Isopentane						CAS #: 78-78-4			
3.556	3.556	(0.424)	43	146320	2.00000	2.466	70.00- 130.00	100.00	
3.556	3.556	(0.424)	57	75097			0.00- 30.00	51.32	
3.556	3.556	(0.424)	72	9820			0.00- 30.00	6.71	
-----									
93 Methyl Cyclohexane						CAS #: 108-87-2			
10.883	10.883	(1.296)	83	107485	2.00000	2.241	70.00- 130.00	100.00	
10.911	10.911	(1.300)	98	43831			0.00- 30.00	40.78	
10.883	10.883	(1.296)	55	121330			0.00- 30.00	112.88	
-----									

QC Flag Legend

- a - Target compound detected but, quantitated amount Below Limit Of Quantitation(BLOQ).
- H - Operator selected an alternate compound hit.

Report Date: 23-Mar-2007 08:09

Air Toxics Ltd.

INTERNAL STANDARD COMPOUNDS  
AREA AND RT SUMMARY

Instrument ID: msd8.i

Calibration Date: 22-MAR-2007

Lab File ID: 8032205.d

Calibration Time: 12:34

Lab Smp Id: ICAL

Client Smp ID: Level 3

Analysis Type: VOA

Level: LOW

Quant Type: ISTD

Sample Type: AIR

Operator: sjr

Method File: /chem/msd8.i/8-22mar.b/t14q322a.m

Misc Info: 200ppbv -&gt; 2ppbv

COMPOUND	STANDARD	AREA LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	283735	170241	397229	275416	-2.93
86 1,4-Difluorobenze	1370859	822515	1919203	1361295	-0.70
123 Chlorobenzene-d5	1067063	640238	1493888	1067969	0.08

COMPOUND	STANDARD	RT LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	8.39	8.06	8.72	8.40	0.00
86 1,4-Difluorobenze	10.25	9.92	10.58	10.25	0.00
123 Chlorobenzene-d5	15.22	14.89	15.55	15.22	0.00

AREA UPPER LIMIT = + 40% of internal standard area.

AREA LOWER LIMIT = - 40% of internal standard area.


RT UPPER LIMIT = + 0.33 minutes of internal standard RT.

RT LOWER LIMIT = - 0.33 minutes of internal standard RT.

Data File: /chem/msd8.1/8-22mar.lb/8032205.d  
Date: 22-MAR-2007 11:38  
Client ID: Level 3  
Sample Info: 2.0ml 1487-115

Column phase: RTX-624

Instrument: msd8.1  
Operator: sjr  
Column diameter: 0.53



Report Date: 23-Mar-2007 08:09

## Air Toxics Ltd.

## AMBIENT AIR METHOD TO14A/TO15

Data file : /chem/msd8.i/8-22mar.b/8032206.d  
 Lab Smp Id: ICAL Client Smp ID: Level 4  
 Inj Date : 22-MAR-2007 12:06  
 Operator : sjr Inst ID: msd8.i  
 Smp Info : 25ml 1487-115  
 Misc Info : 200ppbv->25ppbv  
 Comment :  
 Method : /chem/msd8.i/8-22mar.b/t14q322a.m  
 Meth Date : 23-Mar-2007 08:09 sscott Quant Type: ISTD  
 Cal Date : 22-MAR-2007 12:06 Cal File: 8032206.d  
 Als bottle: 1 Calibration Sample, Level: 4  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: AT04+ENS.sub  
 Target Version: 3.50 Sample Matrix: AIR  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* CpndVariable

Cpnd Variable Local Compound Variable

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT	ON-COL	TARGET RANGE	RATIO	
==	=====	=====	====	=====	=====	=====	=====	=====	
* 67 Bromochloromethane CAS #: 74-97-5									
8.395	8.395	(1.000)	130	284047	25.0000		70.00- 130.00	100.00	
8.395	8.395	(1.000)	128	223837			46.35- 106.35	78.80	
8.368	8.368	(1.000)	49	821968			255.78- 315.78	289.38	
-----									
* 86 1,4-Difluorobenzene CAS #: 540-36-3									
10.248	10.248	(1.000)	114	1370589	25.0000		70.00- 130.00	100.00	
10.248	10.248	(1.000)	88	261113			0.00- 48.88	19.05	
-----									
* 123 Chlorobenzene-d5 CAS #: 3114-55-4									
15.225	15.225	(1.000)	117	1101300	25.0000		70.00- 130.00	100.00	
15.197	15.197	(1.000)	82	733265			0.00- 30.00	66.58	
-----									
\$ 80 1,2-Dichloroethane-d4 CAS #: 17060-07-0									
9.474	9.474	(1.128)	65	634634	25.0000	26.400	70.00- 130.00	100.00	
9.474	9.474	(1.128)	67	318417			0.00- 30.00	50.17	
-----									
\$ 102 Toluene-d8 CAS #: 2037-26-5									
12.985	12.985	(1.267)	98	1353120	25.0000	25.279	70.00- 130.00	100.00	
12.985	12.985	(1.267)	70	171655			0.00- 30.00	12.69	

AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
\$ 102 Toluene-d8 (continued)										
12.985	12.985	(1.267)	100	916907			0.00- 30.00	67.76		
-----										
\$ 138 Bromofluorobenzene										
						CAS #: 460-00-4				
16.745	16.745	(1.100)	174	565422	25.0000	25.319	70.00- 130.00	100.00		
16.745	16.745	(1.100)	95	952630			137.66- 197.66	168.48		
16.745	16.745	(1.100)	176	564473			63.61- 123.61	99.83		
-----										
3 Propylene										
						CAS #: 115-07-1				
2.395	2.395	(0.285)	41	699671	25.0000	21.819	70.00- 130.00	100.00		
2.395	2.395	(0.285)	42	459264			0.00- 30.00	65.64		
2.395	2.395	(0.285)	39	505920			0.00- 30.00	72.31		
-----										
4 Dichlorodifluoromethane/Fr12										
						CAS #: 75-71-8				
2.451	2.451	(0.292)	85	1439765	25.0000	20.838	70.00- 130.00	100.00		
2.451	2.451	(0.292)	87	454966			0.00- 30.00	31.60		
-----										
6 Freon 114										
						CAS #: 76-14-2				
2.589	2.589	(0.308)	135	827106	25.0000	21.434	70.00- 130.00	100.00		
2.589	2.589	(0.308)	137	245941			1.88- 61.88	29.74		
-----										
8 Chloromethane										
						CAS #: 74-87-3				
2.727	2.727	(0.325)	50	760770	25.0000	21.088	70.00- 130.00	100.00		
2.727	2.727	(0.325)	52	240634			0.00- 30.00	31.63		
-----										
9 Butane										
						CAS #: 106-97-8				
2.810	2.810	(0.335)	58	193671	25.0000	20.606	70.00- 130.00	100.00		
2.810	2.810	(0.335)	43	1529744			0.00- 30.00	789.87		
-----										
10 Vinyl Chloride										
						CAS #: 75-01-4				
2.893	2.893	(0.345)	62	722193	25.0000	22.250	70.00- 130.00	100.00		
2.921	2.921	(0.348)	64	215629			0.00- 30.00	29.86		
-----										
11 1,3-Butadiene										
						CAS #: 106-99-0				
2.865	2.865	(0.341)	54	720558	25.0000	21.110	70.00- 130.00	100.00		
2.865	2.865	(0.341)	39	880604			0.00- 30.00	122.21		
-----										
13 Bromomethane										
						CAS #: 74-83-9				
3.418	3.418	(0.407)	94	363545	25.0000	20.450	70.00- 130.00	100.00		
3.418	3.418	(0.407)	96	329744			64.77- 124.77	90.70		
-----										
15 Isopentane										
						CAS #: 78-78-4				
3.557	3.557	(0.424)	43	1199529	25.0000	21.120	70.00- 130.00	100.00		
3.557	3.557	(0.424)	57	727109			0.00- 30.00	60.62		
3.557	3.557	(0.424)	72	67241			0.00- 30.00	5.61		
-----										

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	=====
-----									
16 Chloroethane						CAS #: 75-00-3			
3.557	3.557	(0.424)	64	369678	25.0000	21.806	70.00- 130.00	100.00	
3.557	3.557	(0.424)	49	113454			0.00- 30.00	30.69	
3.557	3.557	(0.424)	66	115437			0.00- 30.00	31.23	
-----									
18 Trichlorofluoromethane/Fr11						CAS #: 75-69-4			
3.888	3.888	(0.463)	101	1352144	25.0000	22.686	70.00- 130.00	100.00	
3.888	3.888	(0.463)	103	878754			34.71- 94.71	64.99	
-----									
21 Ethanol						CAS #: 64-17-5			
4.248	4.248	(0.506)	45	314916	25.0000	22.776	70.00- 130.00	100.00	
4.248	4.248	(0.506)	43	69502			0.00- 30.00	22.07	
4.248	4.248	(0.506)	46	143603			0.00- 30.00	45.60	
-----									
27 Freon 113						CAS #: 76-13-1			
4.745	4.745	(0.565)	151	610034	25.0000	21.002	70.00- 130.00	100.00	
4.745	4.745	(0.565)	153	389700			32.67- 92.67	63.88	
4.745	4.745	(0.565)	101	945928			124.83- 184.83	155.06	
-----									
29 1,1-Dichloroethene						CAS #: 75-35-4			
4.773	4.773	(0.569)	61	1030905	25.0000	20.891	70.00- 130.00	100.00	
4.773	4.773	(0.569)	96	462006			15.10- 75.10	44.82	
4.773	4.773	(0.569)	98	282914			0.00- 58.17	27.44	
-----									
30 Acetone						CAS #: 67-64-1			
4.939	4.939	(0.588)	58	394063	25.0000	22.178	70.00- 130.00	100.00	
4.939	4.939	(0.588)	43	1348833			0.00- 30.00	342.29	
-----									
33 Carbon Disulfide						CAS #: 75-15-0			
5.133	5.133	(0.611)	76	1581151	25.0000	20.896	70.00- 130.00	100.00	
-----									
34 2-Propanol						CAS #: 67-63-0			
5.133	5.133	(0.611)	45	1545684	25.0000	23.198	70.00- 130.00	100.00	
5.133	5.133	(0.611)	43	332266			0.00- 30.00	21.50	
5.133	5.133	(0.611)	59	50332			0.00- 30.00	3.26	
-----									
37 3-Chloropropene						CAS #: 107-05-1			
5.437	5.437	(0.648)	76	266845	25.0000	22.018	70.00- 130.00	100.00	
5.437	5.437	(0.648)	41	1223525			0.00- 30.00	458.52	
-----									
39 Methylene Chloride						CAS #: 75-09-2			
5.686	5.686	(0.677)	49	882904	25.0000	20.971	70.00- 130.00	100.00	
5.713	5.713	(0.681)	84	418715			18.44- 78.44	47.42	
5.686	5.686	(0.677)	51	255897			0.00- 30.00	28.98	
-----									
42 MTBE						CAS #: 1634-04-4			
6.045	6.045	(0.720)	73	1782455	25.0000	21.167	70.00- 130.00	100.00	

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
42 MTBE (continued)									
6.045	6.045	(0.720)	57	538572			0.00- 59.33	30.22	
6.045	6.045	(0.720)	41	553460			0.00- 30.00	31.05	
-----									
43 trans-1,2-Dichloroethene						CAS #: 156-60-5			
6.100	6.100	(0.727)	96	508539	25.0000	19.921	70.00- 130.00	100.00	
6.100	6.100	(0.727)	61	1024177			169.97- 229.97	201.40	
6.100	6.100	(0.727)	98	320921			0.00- 30.00	63.11	
-----									
45 Hexane						CAS #: 110-54-3			
6.432	6.432	(0.766)	57	1315489	25.0000	21.212	70.00- 130.00	100.00	
6.432	6.432	(0.766)	43	960958			0.00- 30.00	73.05	
6.460	6.460	(0.769)	86	174067			0.00- 30.00	13.23	
-----									
52 1,1-Dichloroethane						CAS #: 75-34-3			
6.902	6.902	(0.822)	63	1131951	25.0000	21.952	70.00- 130.00	100.00	
6.902	6.902	(0.822)	65	333923			0.00- 59.58	29.50	
-----									
54 Vinyl Acetate						CAS #: 108-05-4			
6.957	6.957	(0.829)	86	139932	25.0000	24.124	70.00- 130.00	100.00	
6.957	6.957	(0.829)	43	2265817			0.00- 30.00	1619.23	
6.957	6.957	(0.829)	42	212529			0.00- 30.00	151.88	
-----									
63 cis-1,2-Dichloroethene						CAS #: 156-59-2			
7.953	7.953	(0.947)	61	864179	25.0000	22.046	70.00- 130.00	100.00	
7.953	7.953	(0.947)	96	478959			27.36- 87.36	55.42	
7.953	7.953	(0.947)	98	310381			5.83- 65.83	35.92	
-----									
64 2-Butanone						CAS #: 78-93-3			
7.980	7.980	(0.951)	72	288208	25.0000	19.985	70.00- 130.00	100.00	
7.980	7.980	(0.951)	43	1731315			574.29- 634.29	600.72	
7.980	7.980	(0.951)	57	111897			0.00- 30.00	38.83	
-----									
66 Tetrahydrofuran						CAS #: 109-99-9			
8.368	8.368	(0.997)	42	1059865	25.0000	19.552	70.00- 130.00	100.00	
8.368	8.368	(0.997)	71	251229			0.00- 54.93	23.70	
8.368	8.368	(0.997)	72	268980			0.00- 30.00	25.38	
-----									
69 Chloroform						CAS #: 67-66-3			
8.506	8.506	(1.013)	83	961478	25.0000	18.924	70.00- 130.00	100.00	
8.506	8.506	(1.013)	85	617819			33.71- 93.71	64.26	
-----									
72 Cyclohexane						CAS #: 110-82-7			
8.727	8.727	(1.040)	84	778381	25.0000	20.624	70.00- 130.00	100.00	
8.727	8.727	(1.040)	56	1305092			138.68- 198.68	167.67	
8.727	8.727	(1.040)	41	792458			73.81- 133.81	101.81	
-----									



AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
-----									
73	1,1,1-Trichloroethane					CAS #:	71-55-6		
8.755	8.755	(1.043)	97	960173	25.0000	21.290	70.00-	130.00	100.00
8.755	8.755	(1.043)	99	617103			33.95-	93.95	64.27
-----									
75	Carbon Tetrachloride					CAS #:	56-23-5		
9.003	9.003	(1.072)	119	798425	25.0000	22.200	70.00-	130.00	100.00
9.003	9.003	(1.072)	117	880568			81.67-	141.67	110.29
-----									
78	2,2,4-Trimethylpentane					CAS #:	540-84-1		
9.418	9.418	(1.122)	57	3798916	25.0000	22.104	70.00-	130.00	100.00
9.418	9.418	(1.122)	56	1325359			0.00-	30.00	34.89
9.418	9.418	(1.122)	41	1128592			0.00-	30.00	29.71
-----									
79	Benzene					CAS #:	71-43-2		
9.418	9.418	(0.919)	78	1680797	25.0000	20.394	70.00-	130.00	100.00
9.418	9.418	(0.919)	77	402439			0.00-	30.00	23.94
-----									
81	1,2-Dichloroethane					CAS #:	107-06-2		
9.612	9.612	(0.938)	62	836643	25.0000	21.750	70.00-	130.00	100.00
9.612	9.612	(0.938)	64	261907			0.00-	30.00	31.30
-----									
82	Heptane					CAS #:	142-82-5		
9.805	9.805	(0.957)	100	192971	25.0000	19.060	70.00-	130.00	100.00
9.805	9.805	(0.957)	43	1616070			0.00-	30.00	837.47
9.805	9.805	(0.957)	71	627603			0.00-	30.00	325.23
-----									
92	Trichloroethene					CAS #:	79-01-6		
10.662	10.662	(1.040)	95	598872	25.0000	22.749	70.00-	130.00	100.00
10.662	10.662	(1.040)	130	522988			57.13-	117.13	87.33
10.662	10.662	(1.040)	97	363933			32.50-	92.50	60.77
-----									
93	Methyl Cyclohexane					CAS #:	108-87-2		
10.884	10.884	(1.296)	83	969035	25.0000	20.710	70.00-	130.00	100.00
10.911	10.911	(1.300)	98	462139			0.00-	30.00	47.69
10.884	10.884	(1.296)	55	1129900			0.00-	30.00	116.60
-----									
95	1,2-Dichloropropane					CAS #:	78-87-5		
11.188	11.188	(1.092)	63	637102	25.0000	21.115	70.00-	130.00	100.00
11.188	11.188	(1.092)	62	495731			45.03-	105.03	77.81
11.188	11.188	(1.092)	41	503325			45.52-	105.52	79.00
-----									
96	1,4-Dioxane					CAS #:	123-91-1		
11.409	11.409	(1.113)	88	356412	25.0000	22.920	70.00-	130.00	100.00
11.409	11.409	(1.113)	58	364281			70.27-	130.27	102.21
11.409	11.409	(1.113)	57	104086			0.00-	30.00	29.20
-----									

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
-----									
98 Bromodichloromethane						CAS #: 75-27-4			
11.741	11.741	(1.146)	83	985831	25.0000	22.293	70.00- 130.00	100.00	
11.741	11.741	(1.146)	85	589435			32.38- 92.38	59.79	
-----									
100 cis-1,3-Dichloropropene						CAS #: 10061-01-5			
12.626	12.626	(1.232)	75	822568	25.0000	22.099	70.00- 130.00	100.00	
12.626	12.626	(1.232)	77	258696			1.52- 61.52	31.45	
12.626	12.626	(1.232)	39	632684			48.26- 108.26	76.92	
-----									
101 4-Methyl-2-pentanone						CAS #: 108-10-1			
12.902	12.902	(1.259)	58	642296	25.0000	20.982	70.00- 130.00	100.00	
12.902	12.902	(1.259)	43	1883753			0.00- 30.00	293.28	
12.902	12.902	(1.259)	85	212522			0.00- 30.00	33.09	
-----									
103 Toluene						CAS #: 108-88-3			
13.096	13.096	(1.278)	91	1674918	25.0000	23.048	70.00- 130.00	100.00	
13.096	13.096	(1.278)	92	988440			29.45- 89.45	59.01	
-----									
106 trans-1,3-Dichloropropene						CAS #: 10061-02-6			
13.621	13.621	(0.895)	75	894690	25.0000	22.036	70.00- 130.00	100.00	
13.621	13.621	(0.895)	77	266256			1.42- 61.42	29.76	
13.621	13.621	(0.895)	39	639695			43.57- 103.57	71.50	
-----									
108 1,1,2-Trichloroethane						CAS #: 79-00-5			
13.897	13.897	(0.913)	97	521949	25.0000	21.887	70.00- 130.00	100.00	
13.897	13.897	(0.913)	99	318557			31.40- 91.40	61.03	
13.870	13.870	(0.911)	83	451486			60.06- 120.06	86.50	
-----									
109 Tetrachloroethene						CAS #: 127-18-4			
13.925	13.925	(0.915)	166	553653	25.0000	22.294	70.00- 130.00	100.00	
13.925	13.925	(0.915)	129	429899			46.36- 106.36	77.65	
13.925	13.925	(0.915)	131	425518			45.00- 105.00	76.86	
-----									
112 2-Hexanone						CAS #: 591-78-6			
14.257	14.257	(0.936)	58	872471	25.0000	24.167	70.00- 130.00	100.00	
14.257	14.257	(0.936)	43	1780965			176.52- 236.52	204.13	
14.257	14.257	(0.936)	100	123096			0.00- 30.00	14.11	
-----									
114 Dibromochloromethane						CAS #: 124-48-1			
14.423	14.423	(0.947)	129	693588	25.0000	21.695	70.00- 130.00	100.00	
14.423	14.423	(0.947)	127	548637			0.00- 30.00	79.10	
-----									
115 1,2-Dibromoethane						CAS #: 106-93-4			
14.589	14.589	(0.958)	107	740660	25.0000	22.538	70.00- 130.00	100.00	
14.589	14.589	(0.958)	109	692469			62.88- 122.88	93.49	
-----									

AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
124 Chlorobenzene						CAS #:	108-90-7			
15.252	15.252	(1.002)	112	1207014	25.0000	21.910	70.00-	130.00	100.00	
15.252	15.252	(1.002)	114	382995			0.73-	60.73	31.73	
15.252	15.252	(1.002)	77	865791			40.58-	100.58	71.73	
-----										
127 Ethyl Benzene						CAS #:	100-41-4			
15.363	15.363	(1.009)	106	680428	25.0000	22.180	70.00-	130.00	100.00	
15.363	15.363	(1.009)	91	2265782			0.00-	30.00	332.99	
-----										
128 m,p-Xylene						CAS #:	108-38-3			
15.529	15.529	(1.020)	106	873579	25.0000	22.086	70.00-	130.00	100.00	
15.529	15.529	(1.020)	91	1802209			0.00-	30.00	206.30	
-----										
130 o-Xylene						CAS #:	95-47-6			
16.054	16.054	(1.054)	106	838616	25.0000	21.854	70.00-	130.00	100.00	
16.054	16.054	(1.054)	91	1922174			195.99-	255.99	229.21	
-----										
131 Styrene						CAS #:	100-42-5			
16.082	16.082	(1.056)	104	1204987	25.0000	20.890	70.00-	130.00	100.00	
16.082	16.082	(1.056)	78	711465			28.78-	88.78	59.04	
-----										
133 Bromoform						CAS #:	75-25-2			
16.358	16.358	(1.074)	173	593128	25.0000	22.859	70.00-	130.00	100.00	
16.358	16.358	(1.074)	171	316737			21.17-	81.17	53.40	
-----										
135 Cumene						CAS #:	98-82-8			
16.524	16.524	(1.085)	105	2405704	25.0000	20.537	70.00-	130.00	100.00	
16.524	16.524	(1.085)	120	559413			0.00-	30.00	23.25	
16.496	16.496	(1.084)	51	372808			0.00-	30.00	15.50	
-----										
142 1,1,2,2-Tetrachloroethane						CAS #:	79-34-5			
16.966	16.966	(1.114)	83	1184990	25.0000	21.803	70.00-	130.00	100.00	
16.966	16.966	(1.114)	85	748184			34.35-	94.35	63.14	
-----										
143 Propylbenzene						CAS #:	103-65-1			
16.994	16.994	(1.116)	91	2960905	25.0000	22.400	70.00-	130.00	100.00	
16.994	16.994	(1.116)	120	542929			0.00-	30.00	18.34	
16.994	16.994	(1.116)	105	98090			0.00-	30.00	3.31	
-----										
145 4-Ethyltoluene						CAS #:	622-96-8			
17.132	17.132	(1.125)	105	2586366	25.0000	22.423	70.00-	130.00	100.00	
17.132	17.132	(1.125)	120	631838			0.00-	54.57	24.43	
-----										
146 1,3,5-Trimethylbenzene						CAS #:	108-67-8			
17.215	17.215	(1.131)	105	2299995	25.0000	21.018	70.00-	130.00	100.00	
17.215	17.215	(1.131)	120	981553			0.00-	30.00	42.68	
-----										

AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
-----										
151	17.602	17.602	(1.156)	105	2089195	25.0000	21.244	70.00-	130.00	100.00
	17.602	17.602	(1.156)	120	847393			11.41-	71.41	40.56
	CAS #: 95-63-6									
-----										
154	17.907	17.907	(1.176)	146	937051	25.0000	22.541	70.00-	130.00	100.00
	17.907	17.907	(1.176)	148	591360			0.00-	30.00	63.11
	17.907	17.907	(1.176)	111	405175			0.00-	30.00	43.24
	CAS #: 541-73-1									
-----										
155	17.990	17.990	(1.182)	146	1143190	25.0000	21.522	70.00-	130.00	100.00
	17.990	17.990	(1.182)	148	726109			0.00-	30.00	63.52
	17.990	17.990	(1.182)	111	570439			0.00-	30.00	49.90
	CAS #: 106-46-7									
-----										
156	18.128	18.128	(1.191)	91	1677056	25.0000	22.573	70.00-	130.00	100.00
	18.128	18.128	(1.191)	126	286781			0.00-	30.00	17.10
	CAS #: 100-44-7									
-----										
158	18.321	18.321	(1.203)	146	990283	25.0000	21.171	70.00-	130.00	100.00
	18.321	18.321	(1.203)	148	626661			33.83-	93.83	63.28
	18.321	18.321	(1.203)	111	534286			21.29-	81.29	53.95
	CAS #: 95-50-1									
-----										
163	19.593	19.593	(1.287)	180	523210	25.0000	19.929	70.00-	130.00	100.00
	19.593	19.593	(1.287)	182	494582			62.06-	122.06	94.53
	CAS #: 120-82-1									
-----										
164	19.676	19.676	(1.292)	225	797028	25.0000	20.568	70.00-	130.00	100.00
	19.676	19.676	(1.292)	223	512717			35.92-	95.92	64.33
	CAS #: 87-68-3									
-----										
165	19.787	19.787	(1.300)	128	1799958	25.0000	19.884	70.00-	130.00	100.00
	19.787	19.787	(1.300)	127	250333			0.00-	30.00	13.91
	CAS #: 91-20-3									
-----										

Report Date: 23-Mar-2007 08:09

Air Toxics Ltd.

INTERNAL STANDARD COMPOUNDS  
AREA AND RT SUMMARY

Instrument ID: msd8.i

Calibration Date: 22-MAR-2007

Lab File ID: 8032206.d

Calibration Time: 12:34

Lab Smp Id: ICAL

Client Smp ID: Level 4

Analysis Type: VOA

Level: LOW

Quant Type: ISTD

Sample Type: AIR

Operator: sjr

Method File: /chem/msd8.i/8-22mar.b/t14q322a.m

Misc Info: 200ppbv-&gt;25ppbv

COMPOUND	STANDARD	AREA LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	283735	170241	397229	284047	0.11
86 1,4-Difluorobenze	1370859	822515	1919203	1370589	-0.02
123 Chlorobenzene-d5	1067063	640238	1493888	1101300	3.21


COMPOUND	STANDARD	RT LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	8.39	8.06	8.72	8.40	0.00
86 1,4-Difluorobenze	10.25	9.92	10.58	10.25	0.00
123 Chlorobenzene-d5	15.22	14.89	15.55	15.22	0.00

AREA UPPER LIMIT = + 40% of internal standard area.

AREA LOWER LIMIT = - 40% of internal standard area.

RT UPPER LIMIT = + 0.33 minutes of internal standard RT.

RT LOWER LIMIT = - 0.33 minutes of internal standard RT.



Report Date: 26-Mar-2007 13:58

## Air Toxics Ltd.

## AMBIENT AIR METHOD TO14A/TO15

Data file : /chem/msd8.i/8-26mar.b/8032607.d  
 Lab Smp Id: ICAL Client Smp ID: LEVEL 5  
 Inj Date : 26-MAR-2007 12:43  
 Operator : ea Inst ID: msd8.i  
 Smp Info : 50ml #1487-42  
 Misc Info : 200ppbv-50ppbv  
 Comment :  
 Method : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Meth Date : 26-Mar-2007 13:58 ealcan Quant Type: ISTD  
 Cal Date : 26-MAR-2007 12:43 Cal File: 8032607.d  
 Als bottle: 1 Calibration Sample, Level: 5  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: sp5b.sub  
 Target Version: 3.50 Sample Matrix: AIR  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* CpndVariable

Cpnd Variable Local Compound Variable

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	( PPBV)	ON-COL	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
* 67 Bromochloromethane CAS #: 74-97-5									
8.395	8.395	(1.000)	130	291821	25.0000		70.00- 130.00	100.00	
8.395	8.395	(1.000)	128	224346			46.88- 106.88	76.88	
8.367	8.367	(1.000)	49	831446			254.92- 314.92	284.92	
-----									
* 86 1,4-Difluorobenzene CAS #: 540-36-3									
10.248	10.248	(1.000)	114	1354208	25.0000		70.00- 130.00	100.00	
10.248	10.248	(1.000)	88	254759			0.00- 48.81	18.81	
-----									
* 123 Chlorobenzene-d5 CAS #: 3114-55-4									
15.224	15.224	(1.000)	117	1071895	25.0000		70.00- 130.00	100.00	
15.224	15.224	(1.000)	82	702024			35.49- 95.49	65.49	
-----									
194 2-Methylpentane CAS #: 107-83-5									
5.547	5.547	(0.661)	71	933874	50.0000	39.567	70.00- 130.00	100.00	
5.547	5.547	(0.661)	43	3483599			343.03- 403.03	373.03	
5.547	5.547	(0.661)	42	1817422			164.61- 224.61	194.61	
-----									
195 Thiopene CAS #: 110-02-1									
9.833	9.833	(0.960)	84	1776200	50.0000	39.978	70.00- 130.00	100.00(T)	

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
195 Thiopene (continued)									
9.833	9.833	(0.960)	58	1416688			49.76- 109.76	79.76	
0.000	1.000	(0.000)	0	0			0.00- 30.00	0.00	
-----									
196 Indan									
						CAS #: 496-11-7			
18.183	18.183	(1.194)	117	3658164	50.0000	39.033	70.00- 130.00	100.00	
18.183	18.183	(1.194)	118	2078338			26.81- 86.81	56.81	
18.183	18.183	(1.194)	91	672898			0.00- 48.39	18.39	
-----									
197 Indene									
						CAS #: 95-13-6			
18.404	18.404	(1.209)	115	2351599	50.0000	38.579	70.00- 130.00	100.00(T)	
0.000	1.000	(0.000)	16	0			0.00- 30.00	0.00	
-----									
83 2,3-Dimethylpentane									
						CAS #: 565-59-3			
8.810	8.810	(1.049)	71	786693	50.0000	41.099	70.00- 130.00	100.00	
8.782	8.782	(1.046)	56	2712744			314.83- 374.83	344.83	
8.782	8.782	(1.046)	43	2371887			271.50- 331.50	301.50	
-----									

QC Flag Legend

T - Target compound detected outside RT window.



Report Date: 26-Mar-2007 13:58

Air Toxics Ltd.

INTERNAL STANDARD COMPOUNDS  
AREA AND RT SUMMARY

Instrument ID: msd8.i

Calibration Date: 26-MAR-2007

Lab File ID: 8032607.d

Calibration Time: 12:43

Lab Smp Id: ICAL

Client Smp ID: LEVEL 5

Analysis Type: VOA

Level: LOW

Quant Type: ISTD

Sample Type: AIR

Operator: ea

Method File: /chem/msd8.i/8-26mar.b/t14q322b.m

Misc Info: 200ppbv-50ppbv

COMPOUND	STANDARD	AREA LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	291821	175093	408549	291821	0.00
86 1,4-Difluorobenze	1354208	812525	1895891	1354208	0.00
123 Chlorobenzene-d5	1071895	643137	1500653	1071895	0.00

COMPOUND	STANDARD	RT LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	8.40	8.07	8.73	8.40	0.00
86 1,4-Difluorobenze	10.25	9.92	10.58	10.25	0.00
123 Chlorobenzene-d5	15.22	14.89	15.55	15.22	0.00

AREA UPPER LIMIT = + 40% of internal standard area.

AREA LOWER LIMIT = - 40% of internal standard area.

RT UPPER LIMIT = + 0.33 minutes of internal standard RT.

RT LOWER LIMIT = - 0.33 minutes of internal standard RT.

Data File: /chem/msd8.1/8-26mar.1b/8032607.d

Date: 26-MAR-2007 12:43

Client ID: LEVEL 5

Sample Info: 50ml #1487-42


Column phase: RTX-624

Instrument: msd8.1

Operator: ea

Column diameter: 0.53

/chem/msd8.1/8-26mar.1b/8032607.d



Report Date: 23-Mar-2007 08:09

## Air Toxics Ltd.

## AMBIENT AIR METHOD TO14A/TO15

Data file : /chem/msd8.i/8-22mar.b/8032207.d  
 Lab Smp Id: ICAL Client Smp ID: Level 5  
 Inj Date : 22-MAR-2007 12:34  
 Operator : sjr Inst ID: msd8.i  
 Smp Info : 50ml 1487-115  
 Misc Info : 200ppbv->50ppbv  
 Comment :  
 Method : /chem/msd8.i/8-22mar.b/t14q322a.m  
 Meth Date : 23-Mar-2007 08:09 sscott Quant Type: ISTD  
 Cal Date : 22-MAR-2007 12:34 Cal File: 8032207.d  
 Als bottle: 1 Calibration Sample, Level: 5  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: AT04+ENS.sub  
 Target Version: 3.50 Sample Matrix: AIR  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* CpndVariable

Cpnd Variable Local Compound Variable

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	( PPBV)	( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
-----									
* 67	Bromochloromethane						CAS #:	74-97-5	
8.395	8.395	(1.000)	130	283735	25.0000		70.00-	130.00	100.00
8.395	8.395	(1.000)	128	216638			46.35-	106.35	76.35
8.367	8.367	(1.000)	49	810859			255.78-	315.78	285.78
-----									
* 86	1,4-Difluorobenzene						CAS #:	540-36-3	
10.247	10.247	(1.000)	114	1370859	25.0000		70.00-	130.00	100.00
10.247	10.247	(1.000)	88	258843			0.00-	48.88	18.88
-----									
* 123	Chlorobenzene-d5						CAS #:	3114-55-4	
15.224	15.224	(1.000)	117	1067063	25.0000		70.00-	130.00	100.00
15.197	15.197	(1.000)	82	711546			36.68-	96.68	66.68
-----									
\$ 80	1,2-Dichloroethane-d4						CAS #:	17060-07-0	
9.473	9.473	(1.128)	65	583045	25.0000	24.281	70.00-	130.00	100.00
9.473	9.473	(1.128)	67	337682			27.92-	87.92	57.92
-----									
\$ 102	Toluene-d8						CAS #:	2037-26-5	
12.985	12.985	(1.267)	98	1341361	25.0000	25.054	70.00-	130.00	100.00
12.985	12.985	(1.267)	70	169183			0.00-	42.61	12.61

AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
\$ 102 Toluene-d8 (continued)										
12.985	12.985	(1.267)	100	942630			40.27- 100.27	70.27		
-----										
\$ 138 Bromofluorobenzene										
						CAS #:	460-00-4			
16.745	16.745	(1.100)	174	581586	25.0000	26.878	70.00- 130.00	100.00		
16.745	16.745	(1.100)	95	975068			137.66- 197.66	167.66		
16.773	16.773	(1.102)	176	544422			63.61- 123.61	93.61		
-----										
3 Propylene						CAS #:	115-07-1			
2.395	2.395	(0.285)	41	1336408	50.0000	41.722	70.00- 130.00	100.00		
2.395	2.395	(0.285)	42	871400			35.20- 95.20	65.20		
2.395	2.395	(0.285)	39	972963			42.80- 102.80	72.80		
-----										
4 Dichlorodifluoromethane/Fr12						CAS #:	75-71-8			
2.450	2.450	(0.292)	85	2643697	50.0000	38.305	70.00- 130.00	100.00		
2.450	2.450	(0.292)	87	843092			1.89- 61.89	31.89		
-----										
6 Freon 114						CAS #:	76-14-2			
2.588	2.588	(0.308)	135	1585935	50.0000	41.143	70.00- 130.00	100.00		
2.588	2.588	(0.308)	137	505651			1.88- 61.88	31.88		
-----										
8 Chloromethane						CAS #:	74-87-3			
2.727	2.727	(0.325)	50	1446083	50.0000	40.129	70.00- 130.00	100.00		
2.727	2.727	(0.325)	52	426042			0.00- 59.46	29.46		
-----										
9 Butane						CAS #:	106-97-8			
2.810	2.810	(0.335)	58	365258	50.0000	38.906	70.00- 130.00	100.00		
2.810	2.810	(0.335)	43	2932236			772.78- 832.78	802.78		
-----										
10 Vinyl Chloride						CAS #:	75-01-4			
2.920	2.920	(0.348)	62	1367972	50.0000	42.192	70.00- 130.00	100.00		
2.893	2.893	(0.345)	64	414838			0.33- 60.33	30.33		
-----										
11 1,3-Butadiene						CAS #:	106-99-0			
2.893	2.893	(0.345)	54	1369603	50.0000	40.170	70.00- 130.00	100.00		
2.893	2.893	(0.345)	39	1597009			86.60- 146.60	116.60		
-----										
13 Bromomethane						CAS #:	74-83-9			
3.446	3.446	(0.410)	94	693802	50.0000	39.070	70.00- 130.00	100.00		
3.446	3.446	(0.410)	96	657527			64.77- 124.77	94.77		
-----										
15 Isopentane						CAS #:	78-78-4			
3.556	3.556	(0.424)	43	2345298	50.0000	41.338	70.00- 130.00	100.00		
3.556	3.556	(0.424)	57	1348763			27.51- 87.51	57.51		
3.556	3.556	(0.424)	72	139482			0.00- 35.95	5.95		
-----										

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
-----									
16 Chloroethane						CAS #: 75-00-3			
3.584	3.584	(0.427)	64	680681	50.0000	40.195	70.00- 130.00	100.00	
3.556	3.556	(0.424)	49	220676			2.42- 62.42	32.42	
3.556	3.556	(0.424)	66	196848			0.00- 58.92	28.92	
-----									
18 Trichlorofluoromethane/Fr11						CAS #: 75-69-4			
3.888	3.888	(0.463)	101	2561964	50.0000	43.032	70.00- 130.00	100.00	
3.888	3.888	(0.463)	103	1657808			34.71- 94.71	64.71	
-----									
21 Ethanol						CAS #: 64-17-5			
4.247	4.247	(0.506)	45	605204	50.0000	43.818	70.00- 130.00	100.00	
4.247	4.247	(0.506)	43	123622			0.00- 50.43	20.43	
4.247	4.247	(0.506)	46	255462			12.21- 72.21	42.21	
-----									
27 Freon 113						CAS #: 76-13-1			
4.745	4.745	(0.565)	151	1154034	50.0000	39.775	70.00- 130.00	100.00	
4.745	4.745	(0.565)	153	723187			32.67- 92.67	62.67	
4.745	4.745	(0.565)	101	1786742			124.83- 184.83	154.83	
-----									
29 1,1-Dichloroethene						CAS #: 75-35-4			
4.773	4.773	(0.569)	61	1945054	50.0000	39.459	70.00- 130.00	100.00	
4.773	4.773	(0.569)	96	877311			15.10- 75.10	45.10	
4.773	4.773	(0.569)	98	547880			0.00- 58.17	28.17	
-----									
30 Acetone						CAS #: 67-64-1			
4.939	4.939	(0.588)	58	707567	50.0000	39.866	70.00- 130.00	100.00	
4.939	4.939	(0.588)	43	2599224			337.35- 397.35	367.35	
-----									
33 Carbon Disulfide						CAS #: 75-15-0			
5.132	5.132	(0.611)	76	3063615	50.0000	40.532	70.00- 130.00	100.00	
-----									
34 2-Propanol						CAS #: 67-63-0			
5.160	5.160	(0.615)	45	3004294	50.0000	45.138	70.00- 130.00	100.00	
5.132	5.132	(0.611)	43	620101			0.00- 50.64	20.64	
5.160	5.160	(0.615)	59	96296			0.00- 33.21	3.21	
-----									
37 3-Chloropropene						CAS #: 107-05-1			
5.436	5.436	(0.648)	76	523387	50.0000	43.233	70.00- 130.00	100.00	
5.436	5.436	(0.648)	41	2326588			414.53- 474.53	444.53	
-----									
39 Methylene Chloride						CAS #: 75-09-2			
5.685	5.685	(0.677)	49	1646381	50.0000	39.149	70.00- 130.00	100.00	
5.713	5.713	(0.681)	84	797447			18.44- 78.44	48.44	
5.685	5.685	(0.677)	51	492691			0.00- 59.93	29.93	
-----									
42 MTBE						CAS #: 1634-04-4			
6.045	6.045	(0.720)	73	3368161	50.0000	40.042	70.00- 130.00	100.00	

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
42 MTBE (continued)									
6.045	6.045	(0.720)	57	987717			0.00- 59.33	29.33	
6.045	6.045	(0.720)	41	1076052			1.95- 61.95	31.95	
-----									
43 trans-1,2-Dichloroethene						CAS #: 156-60-5			
6.100	6.100	(0.727)	96	967537	50.0000	37.943	70.00- 130.00	100.00	
6.100	6.100	(0.727)	61	1934805			169.97- 229.97	199.97	
6.100	6.100	(0.727)	98	626055			34.71- 94.71	64.71	
-----									
45 Hexane						CAS #: 110-54-3			
6.432	6.432	(0.766)	57	2494698	50.0000	40.270	70.00- 130.00	100.00	
6.432	6.432	(0.766)	43	1861340			44.61- 104.61	74.61	
6.459	6.459	(0.769)	86	343624			0.00- 43.77	13.77	
-----									
52 1,1-Dichloroethane						CAS #: 75-34-3			
6.902	6.902	(0.822)	63	2201347	50.0000	42.739	70.00- 130.00	100.00	
6.902	6.902	(0.822)	65	651146			0.00- 59.58	29.58	
-----									
54 Vinyl Acetate						CAS #: 108-05-4			
6.957	6.957	(0.829)	86	296991	50.0000	51.257	70.00- 130.00	100.00	
6.957	6.957	(0.829)	43	4453738			1469.62-1529.62	1499.62	
6.957	6.957	(0.829)	42	386333			100.08- 160.08	130.08	
-----									
63 cis-1,2-Dichloroethene						CAS #: 156-59-2			
7.952	7.952	(0.947)	61	1618525	50.0000	41.336	70.00- 130.00	100.00	
7.952	7.952	(0.947)	96	928352			27.36- 87.36	57.36	
7.952	7.952	(0.947)	98	579920			5.83- 65.83	35.83	
-----									
64 2-Butanone						CAS #: 78-93-3			
7.980	7.980	(0.951)	72	555279	50.0000	38.546	70.00- 130.00	100.00	
7.980	7.980	(0.951)	43	3355513			574.29- 634.29	604.29	
7.980	7.980	(0.951)	57	225503			10.61- 70.61	40.61	
-----									
66 Tetrahydrofuran						CAS #: 109-99-9			
8.367	8.367	(0.997)	42	2017304	50.0000	37.255	70.00- 130.00	100.00	
8.367	8.367	(0.997)	71	502892			0.00- 54.93	24.93	
8.367	8.367	(0.997)	72	534966			0.00- 56.52	26.52	
-----									
69 Chloroform						CAS #: 67-66-3			
8.505	8.505	(1.013)	83	1830986	50.0000	36.077	70.00- 130.00	100.00	
8.505	8.505	(1.013)	85	1166454			33.71- 93.71	63.71	
-----									
72 Cyclohexane						CAS #: 110-82-7			
8.727	8.727	(1.040)	84	1458593	50.0000	38.689	70.00- 130.00	100.00	
8.727	8.727	(1.040)	56	2460409			138.68- 198.68	168.68	
8.727	8.727	(1.040)	41	1514191			73.81- 133.81	103.81	
-----									

AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
-----										
73	1,1,1-Trichloroethane					CAS #: 71-55-6				
8.754	8.754	(1.043)	97	1820730	50.0000	40.416	70.00- 130.00	100.00		
8.754	8.754	(1.043)	99	1164340			33.95- 93.95	63.95		
-----										
75	Carbon Tetrachloride					CAS #: 56-23-5				
9.003	9.003	(1.072)	119	1530499	50.0000	42.601	70.00- 130.00	100.00		
9.003	9.003	(1.072)	117	1709055			81.67- 141.67	111.67		
-----										
78	2,2,4-Trimethylpentane					CAS #: 540-84-1				
9.418	9.418	(1.122)	57	7210928	50.0000	42.003	70.00- 130.00	100.00		
9.446	9.446	(1.125)	56	2483917			4.45- 64.45	34.45		
9.418	9.418	(1.122)	41	2123088			0.00- 59.44	29.44		
-----										
79	Benzene					CAS #: 71-43-2				
9.418	9.418	(0.919)	78	3194794	50.0000	38.756	70.00- 130.00	100.00		
9.418	9.418	(0.919)	77	779333			0.00- 54.39	24.39		
-----										
81	1,2-Dichloroethane					CAS #: 107-06-2				
9.611	9.611	(0.938)	62	1566874	50.0000	40.725	70.00- 130.00	100.00		
9.611	9.611	(0.938)	64	492564			1.44- 61.44	31.44		
-----										
82	Heptane					CAS #: 142-82-5				
9.805	9.805	(0.957)	100	374296	50.0000	36.962	70.00- 130.00	100.00		
9.805	9.805	(0.957)	43	3067958			789.66- 849.66	819.66		
9.805	9.805	(0.957)	71	1213714			294.27- 354.27	324.27		
-----										
92	Trichloroethene					CAS #: 79-01-6				
10.662	10.662	(1.040)	95	1109731	50.0000	42.147	70.00- 130.00	100.00		
10.662	10.662	(1.040)	130	966953			57.13- 117.13	87.13		
10.662	10.662	(1.040)	97	693589			32.50- 92.50	62.50		
-----										
93	Methyl Cyclohexane					CAS #: 108-87-2				
10.883	10.883	(1.296)	83	1874222	50.0000	40.099	70.00- 130.00	100.00		
10.911	10.911	(1.300)	98	843453			15.00- 75.00	45.00		
10.883	10.883	(1.296)	55	2196126			87.18- 147.18	117.18		
-----										
95	1,2-Dichloropropane					CAS #: 78-87-5				
11.187	11.187	(1.092)	63	1234251	50.0000	40.899	70.00- 130.00	100.00		
11.187	11.187	(1.092)	62	926110			45.03- 105.03	75.03		
11.187	11.187	(1.092)	41	932094			45.52- 105.52	75.52		
-----										
96	1,4-Dioxane					CAS #: 123-91-1				
11.409	11.409	(1.113)	88	682915	50.0000	43.909	70.00- 130.00	100.00		
11.409	11.409	(1.113)	58	684769			70.27- 130.27	100.27		
11.409	11.409	(1.113)	57	208541			0.54- 60.54	30.54		
-----										

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
-----									
98 Bromodichloromethane						CAS #: 75-27-4			
11.740	11.740	(1.146)	83	1848691	50.0000	41.797	70.00- 130.00	100.00	
11.740	11.740	(1.146)	85	1153190			32.38- 92.38	62.38	
-----									
100 cis-1,3-Dichloropropene						CAS #: 10061-01-5			
12.625	12.625	(1.232)	75	1587750	50.0000	42.648	70.00- 130.00	100.00	
12.625	12.625	(1.232)	77	500399			1.52- 61.52	31.52	
12.625	12.625	(1.232)	39	1242642			48.26- 108.26	78.26	
-----									
101 4-Methyl-2-pentanone						CAS #: 108-10-1			
12.902	12.902	(1.259)	58	1264881	50.0000	41.312	70.00- 130.00	100.00	
12.902	12.902	(1.259)	43	3556128			251.14- 311.14	281.14	
12.902	12.902	(1.259)	85	408924			2.33- 62.33	32.33	
-----									
103 Toluene						CAS #: 108-88-3			
13.095	13.095	(1.278)	91	3201294	50.0000	44.044	70.00- 130.00	100.00	
13.095	13.095	(1.278)	92	1903249			29.45- 89.45	59.45	
-----									
106 trans-1,3-Dichloropropene						CAS #: 10061-02-6			
13.621	13.621	(0.895)	75	1684729	50.0000	42.825	70.00- 130.00	100.00	
13.621	13.621	(0.895)	77	529324			1.42- 61.42	31.42	
13.621	13.621	(0.895)	39	1239393			43.57- 103.57	73.57	
-----									
108 1,1,2-Trichloroethane						CAS #: 79-00-5			
13.897	13.897	(0.913)	97	977558	50.0000	42.307	70.00- 130.00	100.00	
13.897	13.897	(0.913)	99	600253			31.40- 91.40	61.40	
13.869	13.869	(0.911)	83	880417			60.06- 120.06	90.06	
-----									
109 Tetrachloroethene						CAS #: 127-18-4			
13.925	13.925	(0.915)	166	1071236	50.0000	44.520	70.00- 130.00	100.00	
13.925	13.925	(0.915)	129	817997			46.36- 106.36	76.36	
13.925	13.925	(0.915)	131	803418			45.00- 105.00	75.00	
-----									
112 2-Hexanone						CAS #: 591-78-6			
14.257	14.257	(0.936)	58	1676318	50.0000	47.923	70.00- 130.00	100.00	
14.257	14.257	(0.936)	43	3461935			176.52- 236.52	206.52	
14.257	14.257	(0.936)	100	262861			0.00- 45.68	15.68	
-----									
114 Dibromochloromethane						CAS #: 124-48-1			
14.422	14.422	(0.947)	129	1342940	50.0000	43.354	70.00- 130.00	100.00	
14.422	14.422	(0.947)	127	1033514			46.96- 106.96	76.96	
-----									
115 1,2-Dibromoethane						CAS #: 106-93-4			
14.588	14.588	(0.958)	107	1401072	50.0000	44.002	70.00- 130.00	100.00	
14.588	14.588	(0.958)	109	1301357			62.88- 122.88	92.88	
-----									



AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
124 Chlorobenzene						CAS #: 108-90-7			
15.252	15.252	(1.002)	112	2347440	50.0000	43.979	70.00- 130.00	100.00	
15.252	15.252	(1.002)	114	721344			0.73- 60.73	30.73	
15.252	15.252	(1.002)	77	1656838			40.58- 100.58	70.58	
-----									
127 Ethyl Benzene						CAS #: 100-41-4			
15.363	15.363	(1.009)	106	1291684	50.0000	43.455	70.00- 130.00	100.00	
15.363	15.363	(1.009)	91	4390490			309.90- 369.90	339.90	
-----									
128 m,p-Xylene						CAS #: 108-38-3			
15.528	15.528	(1.020)	106	1673125	50.0000	43.658	70.00- 130.00	100.00	
15.528	15.528	(1.020)	91	3469703			177.38- 237.38	207.38	
-----									
130 o-Xylene						CAS #: 95-47-6			
16.054	16.054	(1.054)	106	1592825	50.0000	42.841	70.00- 130.00	100.00	
16.054	16.054	(1.054)	91	3599628			195.99- 255.99	225.99	
-----									
131 Styrene						CAS #: 100-42-5			
16.081	16.081	(1.056)	104	2378760	50.0000	42.561	70.00- 130.00	100.00	
16.081	16.081	(1.056)	78	1398236			28.78- 88.78	58.78	
-----									
133 Bromoform						CAS #: 75-25-2			
16.358	16.358	(1.074)	173	1181942	50.0000	47.014	70.00- 130.00	100.00	
16.358	16.358	(1.074)	171	604836			21.17- 81.17	51.17	
-----									
135 Cumene						CAS #: 98-82-8			
16.524	16.524	(1.085)	105	4566296	50.0000	40.232	70.00- 130.00	100.00	
16.524	16.524	(1.085)	120	1056482			0.00- 53.14	23.14	
16.496	16.496	(1.084)	51	702949			0.00- 45.39	15.39	
-----									
142 1,1,2,2-Tetrachloroethane						CAS #: 79-34-5			
16.966	16.966	(1.114)	83	2276949	50.0000	43.238	70.00- 130.00	100.00	
16.966	16.966	(1.114)	85	1465188			34.35- 94.35	64.35	
-----									
143 Propylbenzene						CAS #: 103-65-1			
16.994	16.994	(1.116)	91	5640137	50.0000	44.037	70.00- 130.00	100.00	
16.994	16.994	(1.116)	120	1059941			0.00- 48.79	18.79	
16.994	16.994	(1.116)	105	188920			0.00- 33.35	3.35	
-----									
145 4-Ethyltoluene						CAS #: 622-96-8			
17.132	17.132	(1.125)	105	5072880	50.0000	45.392	70.00- 130.00	100.00	
17.132	17.132	(1.125)	120	1246244			0.00- 54.57	24.57	
-----									
146 1,3,5-Trimethylbenzene						CAS #: 108-67-8			
17.215	17.215	(1.131)	105	4377709	50.0000	41.288	70.00- 130.00	100.00	
17.215	17.215	(1.131)	120	1866622			12.64- 72.64	42.64	
-----									

AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
-----										
151	17.602	17.602	(1.156)	105	3993702	50.0000	41.912	70.00-	130.00	100.00
	17.602	17.602	(1.156)	120	1653770		11.41-	71.41		41.41
	CAS #: 95-63-6									
-----										
154	17.906	17.906	(1.176)	146	1880169	50.0000	46.679	70.00-	130.00	100.00
	17.906	17.906	(1.176)	148	1169490		32.20-	92.20		62.20
	17.879	17.879	(1.174)	111	813794		13.28-	73.28		43.28
	CAS #: 541-73-1									
-----										
155	17.989	17.989	(1.182)	146	2312899	50.0000	44.940	70.00-	130.00	100.00
	17.989	17.989	(1.182)	148	1483842		34.16-	94.16		64.16
	17.989	17.989	(1.182)	111	1073528		16.41-	76.41		46.41
	CAS #: 106-46-7									
-----										
156	18.127	18.127	(1.191)	91	3524875	50.0000	48.968	70.00-	130.00	100.00
	18.127	18.127	(1.191)	126	598031		0.00-	46.97		16.97
	CAS #: 100-44-7									
-----										
158	18.321	18.321	(1.203)	146	1968486	50.0000	43.435	70.00-	130.00	100.00
	18.321	18.321	(1.203)	148	1256428		33.83-	93.83		63.83
	18.321	18.321	(1.203)	111	1009554		21.29-	81.29		51.29
	CAS #: 95-50-1									
-----										
163	19.593	19.593	(1.287)	180	1246406	50.0000	49.000	70.00-	130.00	100.00
	19.593	19.593	(1.287)	182	1147386		62.06-	122.06		92.06
	CAS #: 120-82-1									
-----										
164	19.676	19.676	(1.292)	225	1473379	50.0000	39.242	70.00-	130.00	100.00
	19.676	19.676	(1.292)	223	971308		35.92-	95.92		65.92
	CAS #: 87-68-3									
-----										
165	19.786	19.786	(1.300)	128	3925545	50.0000	44.755	70.00-	130.00	100.00
	19.786	19.786	(1.300)	127	538768		0.00-	43.72		13.72
	CAS #: 91-20-3									
-----										

Report Date: 23-Mar-2007 08:09

Air Toxics Ltd.

INTERNAL STANDARD COMPOUNDS  
AREA AND RT SUMMARY

Instrument ID: msd8.i

Calibration Date: 22-MAR-2007

Lab File ID: 8032207.d

Calibration Time: 12:34

Lab Smp Id: ICAL

Client Smp ID: Level 5

Analysis Type: VOA

Level: LOW

Quant Type: ISTD

Sample Type: AIR

Operator: sjr

Method File: /chem/msd8.i/8-22mar.b/t14q322a.m

Misc Info: 200ppbv-&gt;50ppbv

COMPOUND	STANDARD	AREA LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	283735	170241	397229	283735	0.00
86 1,4-Difluorobenze	1370859	822515	1919203	1370859	0.00
123 Chlorobenzene-d5	1067063	640238	1493888	1067063	0.00


COMPOUND	STANDARD	RT LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	8.39	8.06	8.72	8.39	0.00
86 1,4-Difluorobenze	10.25	9.92	10.58	10.25	0.00
123 Chlorobenzene-d5	15.22	14.89	15.55	15.22	0.00

AREA UPPER LIMIT = + 40% of internal standard area.

AREA LOWER LIMIT = - 40% of internal standard area.

RT UPPER LIMIT = + 0.33 minutes of internal standard RT.

RT LOWER LIMIT = - 0.33 minutes of internal standard RT.



Report Date: 23-Mar-2007 08:09

## Air Toxics Ltd.

## AMBIENT AIR METHOD TO14A/TO15

Data file : /chem/msd8.i/8-22mar.b/8032208.d  
 Lab Smp Id: ICAL Client Smp ID: Level 6  
 Inj Date : 22-MAR-2007 13:02  
 Operator : sjr Inst ID: msd8.i  
 Smp Info : 100ml 1487-115  
 Misc Info : 200ppbv->100ppbv  
 Comment :  
 Method : /chem/msd8.i/8-22mar.b/t14q322a.m  
 Meth Date : 23-Mar-2007 08:09 sscott Quant Type: ISTD  
 Cal Date : 22-MAR-2007 13:02 Cal File: 8032208.d  
 Als bottle: 1 Calibration Sample, Level: 6  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: AT04+ENS.sub  
 Target Version: 3.50 Sample Matrix: AIR  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* CpndVariable

Cpnd Variable Local Compound Variable

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT	ON-COL	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
* 67 Bromochloromethane CAS #: 74-97-5									
8.395	8.395	(1.000)	130	274410	25.0000		70.00- 130.00	100.00	
8.367	8.367	(1.000)	128	219163			46.35- 106.35	79.87	
8.367	8.367	(1.000)	49	829492			255.78- 315.78	302.28	
-----									
* 86 1,4-Difluorobenzene CAS #: 540-36-3									
10.248	10.248	(1.000)	114	1375928	25.0000		70.00- 130.00	100.00	
10.248	10.248	(1.000)	88	254568			0.00- 48.88	18.50	
-----									
* 123 Chlorobenzene-d5 CAS #: 3114-55-4									
15.224	15.224	(1.000)	117	1061257	25.0000		70.00- 130.00	100.00	
15.197	15.197	(1.000)	82	723053			36.68- 96.68	68.13	
-----									
\$ 80 1,2-Dichloroethane-d4 CAS #: 17060-07-0									
9.473	9.473	(1.128)	65	598928	25.0000	25.655	70.00- 130.00	100.00	
9.473	9.473	(1.128)	67	388652			27.92- 87.92	64.89	
-----									
\$ 102 Toluene-d8 CAS #: 2037-26-5									
12.985	12.985	(1.267)	98	1363568	25.0000	25.312	70.00- 130.00	100.00	
12.985	12.985	(1.267)	70	193391			0.00- 42.61	14.18	

AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
\$ 102 Toluene-d8 (continued)										
12.985	12.985	(1.267)	100	1047544			40.27- 100.27	76.82		
-----										
\$ 138 Bromofluorobenzene										
						CAS #:	460-00-4			
16.745	16.745	(1.100)	174	558459	25.0000	25.787	70.00- 130.00	100.00		
16.745	16.745	(1.100)	95	992994			137.66- 197.66	177.81		
16.745	16.745	(1.100)	176	541545			63.61- 123.61	96.97		
-----										
3 Propylene						CAS #:	115-07-1			
2.395	2.395	(0.285)	41	2618213	100.000	87.920	70.00- 130.00	100.00		
2.395	2.395	(0.285)	42	1727704			35.20- 95.20	65.99		
2.395	2.395	(0.285)	39	1908330			42.80- 102.80	72.89		
-----										
4 Dichlorodifluoromethane/Fr12						CAS #:	75-71-8			
2.450	2.450	(0.292)	85	5165186	100.000	81.048	70.00- 130.00	100.00		
2.450	2.450	(0.292)	87	1658624			1.89- 61.89	32.11		
-----										
6 Freon 114						CAS #:	76-14-2			
2.616	2.616	(0.312)	135	3080858	100.000	85.614	70.00- 130.00	100.00		
2.616	2.616	(0.312)	137	966687			1.88- 61.88	31.38		
-----										
8 Chloromethane						CAS #:	74-87-3			
2.727	2.727	(0.325)	50	2730912	100.000	82.840	70.00- 130.00	100.00		
2.727	2.727	(0.325)	52	817475			0.00- 59.46	29.93		
-----										
9 Butane						CAS #:	106-97-8			
2.810	2.810	(0.335)	58	698080	100.000	81.599	70.00- 130.00	100.00		
2.810	2.810	(0.335)	43	5628034			772.78- 832.78	806.22		
-----										
10 Vinyl Chloride						CAS #:	75-01-4			
2.920	2.920	(0.348)	62	2663225	100.000	87.571	70.00- 130.00	100.00		
2.920	2.920	(0.348)	64	808264			0.33- 60.33	30.35		
-----										
11 1,3-Butadiene						CAS #:	106-99-0			
2.893	2.893	(0.345)	54	2662607	100.000	83.980	70.00- 130.00	100.00		
2.893	2.893	(0.345)	39	3839969			86.60- 146.60	144.22		
-----										
13 Bromomethane						CAS #:	74-83-9			
3.446	3.446	(0.410)	94	1439798	100.000	86.635	70.00- 130.00	100.00		
3.446	3.446	(0.410)	96	1295850			64.77- 124.77	90.00		
-----										
15 Isopentane						CAS #:	78-78-4			
3.556	3.556	(0.424)	43	4528672	100.000	86.303	70.00- 130.00	100.00		
3.556	3.556	(0.424)	57	2630973			27.51- 87.51	58.10		
3.556	3.556	(0.424)	72	266218			0.00- 35.95	5.88		
-----										

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	=====
-----									
16 Chloroethane						CAS #: 75-00-3			
3.584	3.584	(0.427)	64	1367985	100.000	86.372	70.00- 130.00	100.00	
3.584	3.584	(0.427)	49	426326			2.42- 62.42	31.16	
3.584	3.584	(0.427)	66	409887			0.00- 58.92	29.96	
-----									
18 Trichlorofluoromethane/Fr11						CAS #: 75-69-4			
3.888	3.888	(0.463)	101	5036405	100.000	89.717	70.00- 130.00	100.00	
3.888	3.888	(0.463)	103	3222575			34.71- 94.71	63.99	
-----									
21 Ethanol						CAS #: 64-17-5			
4.275	4.275	(0.509)	45	1124988	100.000	87.679	70.00- 130.00	100.00	
4.275	4.275	(0.509)	43	236899			0.00- 50.43	21.06	
4.275	4.275	(0.509)	46	475472			12.21- 72.21	42.26	
-----									
27 Freon 113						CAS #: 76-13-1			
4.745	4.745	(0.565)	151	2249177	100.000	83.467	70.00- 130.00	100.00	
4.745	4.745	(0.565)	153	1390357			32.67- 92.67	61.82	
4.745	4.745	(0.565)	101	3469458			124.83- 184.83	154.25	
-----									
29 1,1-Dichloroethene						CAS #: 75-35-4			
4.773	4.773	(0.569)	61	3818502	100.000	83.418	70.00- 130.00	100.00	
4.773	4.773	(0.569)	96	1663923			15.10- 75.10	43.58	
4.773	4.773	(0.569)	98	1077180			0.00- 58.17	28.21	
-----									
30 Acetone						CAS #: 67-64-1			
4.939	4.939	(0.588)	58	1396656	100.000	85.340	70.00- 130.00	100.00	
4.939	4.939	(0.588)	43	5093705			337.35- 397.35	364.71	
-----									
33 Carbon Disulfide						CAS #: 75-15-0			
5.132	5.132	(0.611)	76	5918817	100.000	84.172	70.00- 130.00	100.00	
-----									
34 2-Propanol						CAS #: 67-63-0			
5.160	5.160	(0.615)	45	5917346	100.000	93.820	70.00- 130.00	100.00	
5.160	5.160	(0.615)	43	1171046			0.00- 50.64	19.79	
5.160	5.160	(0.615)	59	195660			0.00- 33.21	3.31	
-----									
37 3-Chloropropene						CAS #: 107-05-1			
5.437	5.437	(0.648)	76	1017552	100.000	89.849	70.00- 130.00	100.00	
5.437	5.437	(0.648)	41	4591103			414.53- 474.53	451.19	
-----									
39 Methylene Chloride						CAS #: 75-09-2			
5.713	5.713	(0.681)	49	3236926	100.000	82.973	70.00- 130.00	100.00	
5.713	5.713	(0.681)	84	1550495			18.44- 78.44	47.90	
5.713	5.713	(0.681)	51	949361			0.00- 59.93	29.33	
-----									
42 MTBE						CAS #: 1634-04-4			
6.045	6.045	(0.720)	73	6577424	100.000	84.072	70.00- 130.00	100.00	

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
42 MTBE (continued)									
6.045	6.045	(0.720)	57	1977342			0.00- 59.33	30.06	
6.045	6.045	(0.720)	41	2097752			1.95- 61.95	31.89	
-----									
43 trans-1,2-Dichloroethene					CAS #: 156-60-5				
6.100	6.100	(0.727)	96	1890823	100.000	80.423	70.00- 130.00	100.00	
6.100	6.100	(0.727)	61	3716781			169.97- 229.97	196.57	
6.100	6.100	(0.727)	98	1203185			34.71- 94.71	63.63	
-----									
45 Hexane					CAS #: 110-54-3				
6.460	6.460	(0.769)	57	4810920	100.000	83.593	70.00- 130.00	100.00	
6.432	6.432	(0.766)	43	3611629			44.61- 104.61	75.07	
6.460	6.460	(0.769)	86	673414			0.00- 43.77	14.00	
-----									
52 1,1-Dichloroethane					CAS #: 75-34-3				
6.902	6.902	(0.822)	63	4203914	100.000	87.111	70.00- 130.00	100.00	
6.902	6.902	(0.822)	65	1256075			0.00- 59.58	29.88	
-----									
54 Vinyl Acetate					CAS #: 108-05-4				
6.957	6.957	(0.829)	86	561325	100.000	100.13	70.00- 130.00	100.00	
6.957	6.957	(0.829)	43	8762705			1469.62-1529.62	1561.08	
6.957	6.957	(0.829)	42	753113			100.08- 160.08	134.17	
-----									
63 cis-1,2-Dichloroethene					CAS #: 156-59-2				
7.953	7.953	(0.947)	61	3211133	100.000	87.457	70.00- 130.00	100.00	
7.953	7.953	(0.947)	96	1802223			27.36- 87.36	56.12	
7.953	7.953	(0.947)	98	1120011			5.83- 65.83	34.88	
-----									
64 2-Butanone					CAS #: 78-93-3				
7.980	7.980	(0.951)	72	1085412	100.000	81.509	70.00- 130.00	100.00	
7.980	7.980	(0.951)	43	6532458			574.29- 634.29	601.84	
7.980	7.980	(0.951)	57	434813			10.61- 70.61	40.06	
-----									
66 Tetrahydrofuran					CAS #: 109-99-9				
8.367	8.367	(0.997)	42	3913572	100.000	78.708	70.00- 130.00	100.00	
8.367	8.367	(0.997)	71	979340			0.00- 54.93	25.02	
8.367	8.367	(0.997)	72	1068465			0.00- 56.52	27.30	
-----									
69 Chloroform					CAS #: 67-66-3				
8.506	8.506	(1.013)	83	3587527	100.000	76.522	70.00- 130.00	100.00	
8.506	8.506	(1.013)	85	2300838			33.71- 93.71	64.13	
-----									
72 Cyclohexane					CAS #: 110-82-7				
8.727	8.727	(1.040)	84	2843440	100.000	81.577	70.00- 130.00	100.00	
8.727	8.727	(1.040)	56	4887392			138.68- 198.68	171.88	
8.727	8.727	(1.040)	41	2925031			73.81- 133.81	102.87	
-----									



AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
-----										
73	1,1,1-Trichloroethane					CAS #: 71-55-6				
8.754	8.754	(1.043)	97	3564210	100.000	84.896	70.00- 130.00	100.00		
8.754	8.754	(1.043)	99	2252502			33.95- 93.95	63.20		
-----										
75	Carbon Tetrachloride					CAS #: 56-23-5				
9.003	9.003	(1.072)	119	2964590	100.000	87.903	70.00- 130.00	100.00		
9.003	9.003	(1.072)	117	3302720			81.67- 141.67	111.41		
-----										
78	2,2,4-Trimethylpentane					CAS #: 540-84-1				
9.446	9.446	(1.125)	57	14024361	100.000	87.175	70.00- 130.00	100.00		
9.446	9.446	(1.125)	56	4907939			4.45- 64.45	35.00		
9.446	9.446	(1.125)	41	4092884			0.00- 59.44	29.18		
-----										
79	Benzene					CAS #: 71-43-2				
9.418	9.418	(0.919)	78	6216949	100.000	78.388	70.00- 130.00	100.00		
9.418	9.418	(0.919)	77	1485753			0.00- 54.39	23.90		
-----										
81	1,2-Dichloroethane					CAS #: 107-06-2				
9.612	9.612	(0.938)	62	3077011	100.000	83.056	70.00- 130.00	100.00		
9.612	9.612	(0.938)	64	957111			1.44- 61.44	31.11		
-----										
82	Heptane					CAS #: 142-82-5				
9.805	9.805	(0.957)	100	727731	100.000	75.911	70.00- 130.00	100.00		
9.805	9.805	(0.957)	43	6016039			789.66- 849.66	826.68		
9.805	9.805	(0.957)	71	2332656			294.27- 354.27	320.54		
-----										
92	Trichloroethene					CAS #: 79-01-6				
10.662	10.662	(1.040)	95	2140717	100.000	84.202	70.00- 130.00	100.00		
10.662	10.662	(1.040)	130	1842017			57.13- 117.13	86.05		
10.662	10.662	(1.040)	97	1345032			32.50- 92.50	62.83		
-----										
93	Methyl Cyclohexane					CAS #: 108-87-2				
10.911	10.911	(1.300)	83	3627158	100.000	83.541	70.00- 130.00	100.00		
10.911	10.911	(1.300)	98	1658014			15.00- 75.00	45.71		
10.883	10.883	(1.296)	55	4177340			87.18- 147.18	115.17		
-----										
95	1,2-Dichloropropane					CAS #: 78-87-5				
11.188	11.188	(1.092)	63	2399132	100.000	82.643	70.00- 130.00	100.00		
11.188	11.188	(1.092)	62	1815553			45.03- 105.03	75.68		
11.188	11.188	(1.092)	41	1814245			45.52- 105.52	75.62		
-----										
96	1,4-Dioxane					CAS #: 123-91-1				
11.409	11.409	(1.113)	88	1330284	100.000	88.488	70.00- 130.00	100.00		
11.409	11.409	(1.113)	58	1317282			70.27- 130.27	99.02		
11.409	11.409	(1.113)	57	412040			0.54- 60.54	30.97		
-----										

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
-----									
98 Bromodichloromethane						CAS #: 75-27-4			
11.741	11.741	(1.146)	83	3623259	100.000	84.732	70.00- 130.00	100.00	
11.741	11.741	(1.146)	85	2296301			32.38- 92.38	63.38	
-----									
100 cis-1,3-Dichloropropene						CAS #: 10061-01-5			
12.625	12.625	(1.232)	75	3193433	100.000	88.022	70.00- 130.00	100.00	
12.625	12.625	(1.232)	77	980903			1.52- 61.52	30.72	
12.598	12.598	(1.229)	39	2427968			48.26- 108.26	76.03	
-----									
101 4-Methyl-2-pentanone						CAS #: 108-10-1			
12.902	12.902	(1.259)	58	2452733	100.000	83.171	70.00- 130.00	100.00	
12.902	12.902	(1.259)	43	7163978			251.14- 311.14	292.08	
12.902	12.902	(1.259)	85	787723			2.33- 62.33	32.12	
-----									
103 Toluene						CAS #: 108-88-3			
13.095	13.095	(1.278)	91	6267602	100.000	88.404	70.00- 130.00	100.00	
13.095	13.095	(1.278)	92	3773465			29.45- 89.45	60.21	
-----									
106 trans-1,3-Dichloropropene						CAS #: 10061-02-6			
13.621	13.621	(0.895)	75	3305187	100.000	87.183	70.00- 130.00	100.00	
13.621	13.621	(0.895)	77	1023953			1.42- 61.42	30.98	
13.621	13.621	(0.895)	39	2407291			43.57- 103.57	72.83	
-----									
108 1,1,2-Trichloroethane						CAS #: 79-00-5			
13.897	13.897	(0.913)	97	1913808	100.000	86.161	70.00- 130.00	100.00	
13.897	13.897	(0.913)	99	1142609			31.40- 91.40	59.70	
13.870	13.870	(0.911)	83	1742327			60.06- 120.06	91.04	
-----									
109 Tetrachloroethene						CAS #: 127-18-4			
13.925	13.925	(0.915)	166	2121016	100.000	90.693	70.00- 130.00	100.00	
13.925	13.925	(0.915)	129	1630806			46.36- 106.36	76.89	
13.925	13.925	(0.915)	131	1573448			45.00- 105.00	74.18	
-----									
112 2-Hexanone						CAS #: 591-78-6			
14.257	14.257	(0.936)	58	3377651	100.000	97.801	70.00- 130.00	100.00	
14.257	14.257	(0.936)	43	6644301			176.52- 236.52	196.71	
14.257	14.257	(0.936)	100	491901			0.00- 45.68	14.56	
-----									
114 Dibromochloromethane						CAS #: 124-48-1			
14.423	14.423	(0.947)	129	2640038	100.000	88.219	70.00- 130.00	100.00	
14.423	14.423	(0.947)	127	2071838			46.96- 106.96	78.48	
-----									
115 1,2-Dibromoethane						CAS #: 106-93-4			
14.589	14.589	(0.958)	107	2779224	100.000	89.964	70.00- 130.00	100.00	
14.589	14.589	(0.958)	109	2584978			62.88- 122.88	93.01	
-----									

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
124 Chlorobenzene						CAS #: 108-90-7			
15.252	15.252	(1.002)	112	4418200	100.000	86.116	70.00- 130.00	100.00	
15.252	15.252	(1.002)	114	1396554			0.73- 60.73	31.61	
15.252	15.252	(1.002)	77	3350897			40.58- 100.58	75.84	
-----									
127 Ethyl Benzene						CAS #: 100-41-4			
15.363	15.363	(1.009)	106	2505821	100.000	87.427	70.00- 130.00	100.00	
15.363	15.363	(1.009)	91	8779160			309.90- 369.90	350.35	
-----									
128 m,p-Xylene						CAS #: 108-38-3			
15.529	15.529	(1.020)	106	3236516	100.000	87.557	70.00- 130.00	100.00	
15.529	15.529	(1.020)	91	6982531			177.38- 237.38	215.74	
-----									
130 o-Xylene						CAS #: 95-47-6			
16.054	16.054	(1.054)	106	3064463	100.000	85.813	70.00- 130.00	100.00	
16.054	16.054	(1.054)	91	7129975			195.99- 255.99	232.67	
-----									
131 Styrene						CAS #: 100-42-5			
16.082	16.082	(1.056)	104	4834510	100.000	88.904	70.00- 130.00	100.00	
16.082	16.082	(1.056)	78	2868142			28.78- 88.78	59.33	
-----									
133 Bromoform						CAS #: 75-25-2			
16.358	16.358	(1.074)	173	2259499	100.000	92.143	70.00- 130.00	100.00	
16.358	16.358	(1.074)	171	1149935			21.17- 81.17	50.89	
-----									
135 Cumene						CAS #: 98-82-8			
16.524	16.524	(1.085)	105	8919536	100.000	81.880	70.00- 130.00	100.00	
16.524	16.524	(1.085)	120	2075981			0.00- 53.14	23.27	
16.496	16.496	(1.084)	51	1377696			0.00- 45.39	15.45	
-----									
142 1,1,2,2-Tetrachloroethane						CAS #: 79-34-5			
16.966	16.966	(1.114)	83	4515433	100.000	88.659	70.00- 130.00	100.00	
16.966	16.966	(1.114)	85	2869533			34.35- 94.35	63.55	
-----									
143 Propylbenzene						CAS #: 103-65-1			
16.994	16.994	(1.116)	91	11424449	100.000	91.577	70.00- 130.00	100.00	
16.994	16.994	(1.116)	120	1995294			0.00- 48.79	17.47	
16.994	16.994	(1.116)	105	392922			0.00- 33.35	3.44	
-----									
145 4-Ethyltoluene						CAS #: 622-96-8			
17.132	17.132	(1.125)	105	10175388	100.000	93.121	70.00- 130.00	100.00	
17.132	17.132	(1.125)	120	2508882			0.00- 54.57	24.66	
-----									
146 1,3,5-Trimethylbenzene						CAS #: 108-67-8			
17.215	17.215	(1.131)	105	8697612	100.000	85.474	70.00- 130.00	100.00	
17.215	17.215	(1.131)	120	3667776			12.64- 72.64	42.17	
-----									

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
-----									
151	17.602	17.602	105	7984521	100.000	86.992	70.00- 130.00	100.00	
	17.602	(1.156)	120	3290133			11.41- 71.41	41.21	
								CAS #: 95-63-6	
-----									
154	17.906	17.906	146	3592491	100.000	91.569	70.00- 130.00	100.00	
	17.906	(1.176)	148	2270275			32.20- 92.20	63.20	
	17.879	(1.174)	111	1698157			13.28- 73.28	47.27	
								CAS #: 541-73-1	
-----									
155	17.989	17.989	146	4959897	100.000	97.504	70.00- 130.00	100.00	
	17.989	(1.182)	148	3107048			34.16- 94.16	62.64	
	17.989	(1.182)	111	2095150			16.41- 76.41	42.24	
								CAS #: 106-46-7	
-----									
156	18.128	18.128	91	7090361	100.000	99.229	70.00- 130.00	100.00	
	18.128	(1.191)	126	1205592			0.00- 46.97	17.00	
								CAS #: 100-44-7	
-----									
158	18.321	18.321	146	3753348	100.000	86.154	70.00- 130.00	100.00	
	18.321	(1.203)	148	2318870			33.83- 93.83	61.78	
	18.321	(1.203)	111	1975932			21.29- 81.29	52.64	
								CAS #: 95-50-1	
-----									
163	19.593	19.593	180	2760569	100.000	106.69	70.00- 130.00	100.00	
	19.593	(1.287)	182	2604598			62.06- 122.06	94.35	
								CAS #: 120-82-1	
-----									
164	19.676	19.676	225	2866710	100.000	81.503	70.00- 130.00	100.00	
	19.676	(1.292)	223	1823041			35.92- 95.92	63.59	
								CAS #: 87-68-3	
-----									
165	19.787	19.787	128	8609290	100.000	99.016	70.00- 130.00	100.00	
	19.787	(1.300)	127	1214285			0.00- 43.72	14.10	
								CAS #: 91-20-3	
-----									

Report Date: 23-Mar-2007 08:09

Air Toxics Ltd.

INTERNAL STANDARD COMPOUNDS  
AREA AND RT SUMMARY

Instrument ID: msd8.i

Calibration Date: 22-MAR-2007

Lab File ID: 8032208.d

Calibration Time: 12:34

Lab Smp Id: ICAL

Client Smp ID: Level 6

Analysis Type: VOA

Level: LOW

Quant Type: ISTD

Sample Type: AIR

Operator: sjr

Method File: /chem/msd8.i/8-22mar.b/t14q322a.m

Misc Info: 200ppbv-&gt;100ppbv

COMPOUND	STANDARD	AREA LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	283735	170241	397229	274410	-3.29
86 1,4-Difluorobenze	1370859	822515	1919203	1375928	0.37
123 Chlorobenzene-d5	1067063	640238	1493888	1061257	-0.54

COMPOUND	STANDARD	RT LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	8.39	8.06	8.72	8.40	0.00
86 1,4-Difluorobenze	10.25	9.92	10.58	10.25	0.00
123 Chlorobenzene-d5	15.22	14.89	15.55	15.22	0.00

AREA UPPER LIMIT = + 40% of internal standard area.

AREA LOWER LIMIT = - 40% of internal standard area.

RT UPPER LIMIT = + 0.33 minutes of internal standard RT.


RT LOWER LIMIT = - 0.33 minutes of internal standard RT.

Data File: /chem/msd8.1/8-22mar.b/8032208.d  
Date: 22-MAR-2007 13:02  
Client ID: Level 6  
Sample Info: 100ml 1487-115

Column phase: RTX-624

Instrument: msd8.1  
Operator: sjr  
Column diameter: 0.53

/chem/msd8.1/8-22mar.b/8032208.d



Report Date: 26-Mar-2007 13:58

## Air Toxics Ltd.

## AMBIENT AIR METHOD TO14A/TO15

Data file : /chem/msd8.i/8-26mar.b/8032608.d  
 Lab Smp Id: ICAL Client Smp ID: LEVEL 7  
 Inj Date : 26-MAR-2007 13:11  
 Operator : ea Inst ID: msd8.i  
 Smp Info : 200ml #1487-42  
 Misc Info : 200ppbv-200ppbv  
 Comment :  
 Method : /chem/msd8.i/8-26mar.b/t14q322b.m  
 Meth Date : 26-Mar-2007 13:58 ealcan Quant Type: ISTD  
 Cal Date : 26-MAR-2007 13:11 Cal File: 8032608.d  
 Als bottle: 1 Calibration Sample, Level: 7  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: sp5b.sub  
 Target Version: 3.50 Sample Matrix: AIR  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* CpndVariable

Cpnd Variable Local Compound Variable

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	( PPBV)	( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	====	=====	=====	=====	=====	=====	
* 67 Bromochloromethane CAS #: 74-97-5									
8.395	8.395	(1.000)	130	281494	25.0000		70.00- 130.00	100.00	
8.395	8.395	(1.000)	128	230544			46.88- 106.88	81.90	
8.367	8.367	(1.000)	49	801676			254.92- 314.92	284.79	
-----									
* 86 1,4-Difluorobenzene CAS #: 540-36-3									
10.248	10.248	(1.000)	114	1332533	25.0000		70.00- 130.00	100.00	
10.248	10.248	(1.000)	88	260063			0.00- 48.81	19.52	
-----									
* 123 Chlorobenzene-d5 CAS #: 3114-55-4									
15.224	15.224	(1.000)	117	1030363	25.0000		70.00- 130.00	100.00	
15.197	15.197	(1.000)	82	675300			35.49- 95.49	65.54	
-----									
194 2-Methylpentane CAS #: 107-83-5									
5.519	5.519	(0.657)	71	3347631	200.000	161.28	70.00- 130.00	100.00	
5.519	5.519	(0.657)	43	13000628			343.03- 403.03	388.35	
5.519	5.519	(0.657)	42	6719106			164.61- 224.61	200.71	
-----									
195 Thiopene CAS #: 110-02-1									
9.833	9.833	(0.960)	84	6737660	200.000	166.88	70.00- 130.00	100.00(T)	

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
195 Thiopene (continued)									
9.833	9.833	(0.960)	58	5424886			49.76- 109.76	80.52	
0.000	1.000	(0.000)	0	0			0.00- 30.00	0.00	
-----									
196 Indan									
						CAS #: 496-11-7			
18.183	18.183	(1.194)	117	14309681	200.000	170.54	70.00- 130.00	100.00	
18.183	18.183	(1.194)	118	8094903			26.81- 86.81	56.57	
18.183	18.183	(1.194)	91	2643183			0.00- 48.39	18.47	
-----									
197 Indene									
						CAS #: 95-13-6			
18.404	18.404	(1.209)	115	9753957	200.000	176.32	70.00- 130.00	100.00(T)	
0.000	1.000	(0.000)	16	0			0.00- 30.00	0.00	
-----									
83 2,3-Dimethylpentane									
						CAS #: 565-59-3			
8.782	8.782	(1.046)	71	2877196	200.000	168.21	70.00- 130.00	100.00	
8.782	8.782	(1.046)	56	10172829			314.83- 374.83	353.57	
8.782	8.782	(1.046)	43	8826150			271.50- 331.50	306.76	
-----									

QC Flag Legend

T - Target compound detected outside RT window.



Report Date: 26-Mar-2007 13:58

Air Toxics Ltd.

INTERNAL STANDARD COMPOUNDS  
AREA AND RT SUMMARY

Instrument ID: msd8.i

Calibration Date: 26-MAR-2007

Lab File ID: 8032608.d

Calibration Time: 12:43

Lab Smp Id: ICAL

Client Smp ID: LEVEL 7

Analysis Type: VOA

Level: LOW

Quant Type: ISTD

Sample Type: AIR

Operator: ea

Method File: /chem/msd8.i/8-26mar.b/t14q322b.m

Misc Info: 200ppbv-200ppbv

COMPOUND	STANDARD	AREA LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	291821	175093	408549	281494	-3.54
86 1,4-Difluorobenze	1354208	812525	1895891	1332533	-1.60
123 Chlorobenzene-d5	1071895	643137	1500653	1030363	-3.87

COMPOUND	STANDARD	RT LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	8.40	8.07	8.73	8.39	0.00
86 1,4-Difluorobenze	10.25	9.92	10.58	10.25	0.00
123 Chlorobenzene-d5	15.22	14.89	15.55	15.22	0.00

AREA UPPER LIMIT = + 40% of internal standard area.

AREA LOWER LIMIT = - 40% of internal standard area.

RT UPPER LIMIT = + 0.33 minutes of internal standard RT.

RT LOWER LIMIT = - 0.33 minutes of internal standard RT.

Data File: /chem/msd8.1/8-26mar.1b/8032608.d

Date: 26-MAR-2007 13:11

Client ID: LEVEL 7

Sample Info: 200ml #1487-42


Column phase: RTX-624

Instrument: msd8.1

Operator: ea

Column diameter: 0.53

/chem/msd8.1/8-26mar.1b/8032608.d



Report Date: 23-Mar-2007 08:09

## Air Toxics Ltd.

## AMBIENT AIR METHOD TO14A/TO15

Data file : /chem/msd8.i/8-22mar.b/8032209.d  
 Lab Smp Id: ICAL Client Smp ID: Level 7  
 Inj Date : 22-MAR-2007 13:30  
 Operator : sjr Inst ID: msd8.i  
 Smp Info : 200ml 1487-115  
 Misc Info : 200ppbv->200ppbv  
 Comment :  
 Method : /chem/msd8.i/8-22mar.b/t14q322a.m  
 Meth Date : 23-Mar-2007 08:09 sscott Quant Type: ISTD  
 Cal Date : 22-MAR-2007 13:30 Cal File: 8032209.d  
 Als bottle: 1 Calibration Sample, Level: 7  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: AT04+ENS.sub  
 Target Version: 3.50 Sample Matrix: AIR  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* CpndVariable

Cpnd Variable Local Compound Variable

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	( PPBV)	ON-COL	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
* 67 Bromochloromethane CAS #: 74-97-5									
8.395	8.395	(1.000)	130	279518	25.0000		70.00- 130.00	100.00	
8.395	8.395	(1.000)	128	222977			46.35- 106.35	79.77	
8.367	8.367	(1.000)	49	860256			255.78- 315.78	307.76	
-----									
* 86 1,4-Difluorobenzene CAS #: 540-36-3									
10.248	10.248	(1.000)	114	1399810	25.0000		70.00- 130.00	100.00	
10.248	10.248	(1.000)	88	256707			0.00- 48.88	18.34	
-----									
* 123 Chlorobenzene-d5 CAS #: 3114-55-4									
15.224	15.224	(1.000)	117	1085707	25.0000		70.00- 130.00	100.00	
15.197	15.197	(1.000)	82	735939			36.68- 96.68	67.78	
-----									
\$ 80 1,2-Dichloroethane-d4 CAS #: 17060-07-0									
9.473	9.473	(1.128)	65	661231	25.0000	27.367	70.00- 130.00	100.00	
9.446	9.446	(1.125)	67	465615			27.92- 87.92	70.42	
-----									
\$ 102 Toluene-d8 CAS #: 2037-26-5									
12.985	12.985	(1.267)	98	1412741	25.0000	25.663	70.00- 130.00	100.00	
12.985	12.985	(1.267)	70	173136			0.00- 42.61	12.26	

AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
\$ 102 Toluene-d8 (continued)										
12.985	12.985	(1.267)	100	954832			40.27- 100.27	67.59		
-----										
\$ 138 Bromofluorobenzene										
						CAS #:	460-00-4			
16.745	16.745	(1.100)	174	570442	25.0000	25.638	70.00- 130.00	100.00		
16.745	16.745	(1.100)	95	1011334			137.66- 197.66	177.29		
16.745	16.745	(1.100)	176	558711			63.61- 123.61	97.94		
-----										
3 Propylene						CAS #:	115-07-1			
2.395	2.395	(0.285)	41	5308101	200.000	179.48	70.00- 130.00	100.00		
2.395	2.395	(0.285)	42	3479256			35.20- 95.20	65.55		
2.395	2.395	(0.285)	39	3813826			42.80- 102.80	71.85		
-----										
4 Dichlorodifluoromethane/Fr12						CAS #:	75-71-8			
2.450	2.450	(0.292)	85	10101380	200.000	161.58	70.00- 130.00	100.00		
2.450	2.450	(0.292)	87	3235750			1.89- 61.89	32.03		
-----										
6 Freon 114						CAS #:	76-14-2			
2.616	2.616	(0.312)	135	5891518	200.000	166.16	70.00- 130.00	100.00		
2.616	2.616	(0.312)	137	1868154			1.88- 61.88	31.71		
-----										
8 Chloromethane						CAS #:	74-87-3			
2.727	2.727	(0.325)	50	5299226	200.000	164.76	70.00- 130.00	100.00		
2.727	2.727	(0.325)	52	1522402			0.00- 59.46	28.73		
-----										
9 Butane						CAS #:	106-97-8			
2.837	2.837	(0.338)	58	1398871	200.000	167.12	70.00- 130.00	100.00		
2.837	2.837	(0.338)	43	11305102			772.78- 832.78	808.16		
-----										
10 Vinyl Chloride						CAS #:	75-01-4			
2.893	2.893	(0.345)	62	5274740	200.000	174.60	70.00- 130.00	100.00		
2.893	2.893	(0.345)	64	1597095			0.33- 60.33	30.28		
-----										
11 1,3-Butadiene						CAS #:	106-99-0			
2.920	2.920	(0.348)	54	5298869	200.000	169.14	70.00- 130.00	100.00		
2.893	2.893	(0.345)	39	7534296			86.60- 146.60	142.19		
-----										
13 Bromomethane						CAS #:	74-83-9			
3.446	3.446	(0.410)	94	2889199	200.000	174.95	70.00- 130.00	100.00		
3.446	3.446	(0.410)	96	2663390			64.77- 124.77	92.18		
-----										
15 Isopentane						CAS #:	78-78-4			
3.556	3.556	(0.424)	43	8928696	200.000	172.74	70.00- 130.00	100.00		
3.556	3.556	(0.424)	57	5194820			27.51- 87.51	58.18		
3.584	3.584	(0.427)	72	519565			0.00- 35.95	5.82		
-----										

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	=====
-----									
16 Chloroethane						CAS #: 75-00-3			
3.612	3.612	(0.430)	64	2745566	200.000	174.52	70.00- 130.00	100.00	
3.584	3.584	(0.427)	49	861002			2.42- 62.42	31.36	
3.612	3.612	(0.430)	66	805195			0.00- 58.92	29.33	
-----									
18 Trichlorofluoromethane/Fr11						CAS #: 75-69-4			
3.916	3.916	(0.466)	101	10012696	200.000	178.81	70.00- 130.00	100.00	
3.916	3.916	(0.466)	103	6341136			34.71- 94.71	63.33	
-----									
21 Ethanol						CAS #: 64-17-5			
4.331	4.331	(0.516)	45	2255570	200.000	177.45	70.00- 130.00	100.00	
4.331	4.331	(0.516)	43	448642			0.00- 50.43	19.89	
4.331	4.331	(0.516)	46	957418			12.21- 72.21	42.45	
-----									
27 Freon 113						CAS #: 76-13-1			
4.745	4.745	(0.565)	151	4370646	200.000	164.83	70.00- 130.00	100.00	
4.745	4.745	(0.565)	153	2713210			32.67- 92.67	62.08	
4.745	4.745	(0.565)	101	6769526			124.83- 184.83	154.89	
-----									
29 1,1-Dichloroethene						CAS #: 75-35-4			
4.801	4.801	(0.572)	61	7533717	200.000	166.92	70.00- 130.00	100.00	
4.801	4.801	(0.572)	96	3316120			15.10- 75.10	44.02	
4.801	4.801	(0.572)	98	2080886			0.00- 58.17	27.62	
-----									
30 Acetone						CAS #: 67-64-1			
4.939	4.939	(0.588)	58	2805977	200.000	173.83	70.00- 130.00	100.00	
4.939	4.939	(0.588)	43	10159313			337.35- 397.35	362.06	
-----									
33 Carbon Disulfide						CAS #: 75-15-0			
5.160	5.160	(0.615)	76	11799377	200.000	169.72	70.00- 130.00	100.00	
-----									
34 2-Propanol						CAS #: 67-63-0			
5.160	5.160	(0.615)	45	11847068	200.000	187.32	70.00- 130.00	100.00	
5.160	5.160	(0.615)	43	2329975			0.00- 50.64	19.67	
5.160	5.160	(0.615)	59	395213			0.00- 33.21	3.34	
-----									
37 3-Chloropropene						CAS #: 107-05-1			
5.437	5.437	(0.648)	76	1998211	200.000	177.98	70.00- 130.00	100.00	
5.437	5.437	(0.648)	41	8970627			414.53- 474.53	448.93	
-----									
39 Methylene Chloride						CAS #: 75-09-2			
5.713	5.713	(0.681)	49	6321788	200.000	164.70	70.00- 130.00	100.00	
5.713	5.713	(0.681)	84	3055111			18.44- 78.44	48.33	
5.713	5.713	(0.681)	51	1846237			0.00- 59.93	29.20	
-----									
42 MTBE						CAS #: 1634-04-4			
6.045	6.045	(0.720)	73	13152510	200.000	169.99	70.00- 130.00	100.00	

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
42 MTBE (continued)									
6.045	6.045	(0.720)	57	3882461			0.00- 59.33	29.52	
6.045	6.045	(0.720)	41	4049097			1.95- 61.95	30.79	
-----									
43 trans-1,2-Dichloroethene					CAS #: 156-60-5				
6.100	6.100	(0.727)	96	3729950	200.000	161.71	70.00- 130.00	100.00	
6.100	6.100	(0.727)	61	7405681			169.97- 229.97	198.55	
6.100	6.100	(0.727)	98	2367667			34.71- 94.71	63.48	
-----									
45 Hexane					CAS #: 110-54-3				
6.432	6.432	(0.766)	57	9508794	200.000	167.48	70.00- 130.00	100.00	
6.432	6.432	(0.766)	43	7106283			44.61- 104.61	74.73	
6.460	6.460	(0.769)	86	1305143			0.00- 43.77	13.73	
-----									
52 1,1-Dichloroethane					CAS #: 75-34-3				
6.902	6.902	(0.822)	63	8387780	200.000	174.91	70.00- 130.00	100.00	
6.902	6.902	(0.822)	65	2490463			0.00- 59.58	29.69	
-----									
54 Vinyl Acetate					CAS #: 108-05-4				
6.957	6.957	(0.829)	86	1135815	200.000	199.12	70.00- 130.00	100.00	
6.957	6.957	(0.829)	43	17639765			1469.62-1529.62	1553.05	
6.957	6.957	(0.829)	42	1475594			100.08- 160.08	129.91	
-----									
63 cis-1,2-Dichloroethene					CAS #: 156-59-2				
7.953	7.953	(0.947)	61	6381943	200.000	174.92	70.00- 130.00	100.00	
7.953	7.953	(0.947)	96	3500554			27.36- 87.36	54.85	
7.953	7.953	(0.947)	98	2250958			5.83- 65.83	35.27	
-----									
64 2-Butanone					CAS #: 78-93-3				
7.980	7.980	(0.951)	72	2143034	200.000	163.72	70.00- 130.00	100.00	
7.980	7.980	(0.951)	43	13228944			574.29- 634.29	617.30	
7.980	7.980	(0.951)	57	881095			10.61- 70.61	41.11	
-----									
66 Tetrahydrofuran					CAS #: 109-99-9				
8.367	8.367	(0.997)	42	7835677	200.000	160.78	70.00- 130.00	100.00	
8.367	8.367	(0.997)	71	1964568			0.00- 54.93	25.07	
8.367	8.367	(0.997)	72	2111107			0.00- 56.52	26.94	
-----									
69 Chloroform					CAS #: 67-66-3				
8.506	8.506	(1.013)	83	7043808	200.000	153.24	70.00- 130.00	100.00	
8.506	8.506	(1.013)	85	4530865			33.71- 93.71	64.32	
-----									
72 Cyclohexane					CAS #: 110-82-7				
8.727	8.727	(1.040)	84	5609830	200.000	163.73	70.00- 130.00	100.00	
8.727	8.727	(1.040)	56	9584320			138.68- 198.68	170.85	
8.727	8.727	(1.040)	41	5766105			73.81- 133.81	102.79	
-----									

AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
-----										
73	1,1,1-Trichloroethane					CAS #: 71-55-6				
8.754	8.754	(1.043)	97	7033900	200.000	169.50	70.00- 130.00	100.00		
8.754	8.754	(1.043)	99	4470247			33.95- 93.95	63.55		
-----										
75	Carbon Tetrachloride					CAS #: 56-23-5				
9.003	9.003	(1.072)	119	5870008	200.000	175.12	70.00- 130.00	100.00		
9.003	9.003	(1.072)	117	6520132			81.67- 141.67	111.08		
-----										
78	2,2,4-Trimethylpentane					CAS #: 540-84-1				
9.446	9.446	(1.125)	57	27970218	200.000	174.96	70.00- 130.00	100.00		
9.446	9.446	(1.125)	56	9810126			4.45- 64.45	35.07		
9.446	9.446	(1.125)	41	8184579			0.00- 59.44	29.26		
-----										
79	Benzene					CAS #: 71-43-2				
9.418	9.418	(0.919)	78	12323899	200.000	158.07	70.00- 130.00	100.00		
9.418	9.418	(0.919)	77	2904375			0.00- 54.39	23.57		
-----										
81	1,2-Dichloroethane					CAS #: 107-06-2				
9.612	9.612	(0.938)	62	6122562	200.000	167.69	70.00- 130.00	100.00		
9.612	9.612	(0.938)	64	1881649			1.44- 61.44	30.73		
-----										
82	Heptane					CAS #: 142-82-5				
9.805	9.805	(0.957)	100	1422278	200.000	152.72	70.00- 130.00	100.00		
9.805	9.805	(0.957)	43	11924797			789.66- 849.66	838.43		
9.805	9.805	(0.957)	71	4647758			294.27- 354.27	326.78		
-----										
92	Trichloroethene					CAS #: 79-01-6				
10.662	10.662	(1.040)	95	4227724	200.000	168.59	70.00- 130.00	100.00		
10.662	10.662	(1.040)	130	3693207			57.13- 117.13	87.36		
10.662	10.662	(1.040)	97	2711041			32.50- 92.50	64.13		
-----										
93	Methyl Cyclohexane					CAS #: 108-87-2				
10.883	10.883	(1.296)	83	7201337	200.000	168.04	70.00- 130.00	100.00		
10.911	10.911	(1.300)	98	3309603			15.00- 75.00	45.96		
10.883	10.883	(1.296)	55	8341830			87.18- 147.18	115.84		
-----										
95	1,2-Dichloropropane					CAS #: 78-87-5				
11.188	11.188	(1.092)	63	4749296	200.000	166.24	70.00- 130.00	100.00		
11.188	11.188	(1.092)	62	3604653			45.03- 105.03	75.90		
11.188	11.188	(1.092)	41	3676935			45.52- 105.52	77.42		
-----										
96	1,4-Dioxane					CAS #: 123-91-1				
11.409	11.409	(1.113)	88	2658131	200.000	178.47	70.00- 130.00	100.00		
11.409	11.409	(1.113)	58	2649215			70.27- 130.27	99.66		
11.409	11.409	(1.113)	57	807873			0.54- 60.54	30.39		
-----										

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
-----									
98 Bromodichloromethane						CAS #: 75-27-4			
11.741	11.741	(1.146)	83	7190559	200.000	170.21	70.00- 130.00	100.00	
11.741	11.741	(1.146)	85	4538542			32.38- 92.38	63.12	
-----									
100 cis-1,3-Dichloropropene						CAS #: 10061-01-5			
12.625	12.625	(1.232)	75	6325832	200.000	175.57	70.00- 130.00	100.00	
12.625	12.625	(1.232)	77	1954764			1.52- 61.52	30.90	
12.598	12.598	(1.229)	39	4867305			48.26- 108.26	76.94	
-----									
101 4-Methyl-2-pentanone						CAS #: 108-10-1			
12.902	12.902	(1.259)	58	4857563	200.000	167.22	70.00- 130.00	100.00	
12.902	12.902	(1.259)	43	14289544			251.14- 311.14	294.17	
12.902	12.902	(1.259)	85	1557025			2.33- 62.33	32.05	
-----									
103 Toluene						CAS #: 108-88-3			
13.095	13.095	(1.278)	91	12583853	200.000	178.26	70.00- 130.00	100.00	
13.095	13.095	(1.278)	92	7516713			29.45- 89.45	59.73	
-----									
106 trans-1,3-Dichloropropene						CAS #: 10061-02-6			
13.621	13.621	(0.895)	75	6649054	200.000	175.62	70.00- 130.00	100.00	
13.621	13.621	(0.895)	77	2070772			1.42- 61.42	31.14	
13.621	13.621	(0.895)	39	4783510			43.57- 103.57	71.94	
-----									
108 1,1,2-Trichloroethane						CAS #: 79-00-5			
13.870	13.870	(0.911)	97	3726387	200.000	169.06	70.00- 130.00	100.00	
13.870	13.870	(0.911)	99	2290121			31.40- 91.40	61.46	
13.870	13.870	(0.911)	83	3451808			60.06- 120.06	92.63	
-----									
109 Tetrachloroethene						CAS #: 127-18-4			
13.925	13.925	(0.915)	166	4167743	200.000	178.02	70.00- 130.00	100.00	
13.925	13.925	(0.915)	129	3212119			46.36- 106.36	77.07	
13.925	13.925	(0.915)	131	3048828			45.00- 105.00	73.15	
-----									
112 2-Hexanone						CAS #: 591-78-6			
14.257	14.257	(0.936)	58	6522486	200.000	187.49	70.00- 130.00	100.00	
14.257	14.257	(0.936)	43	13577863			176.52- 236.52	208.17	
14.257	14.257	(0.936)	100	1038835			0.00- 45.68	15.93	
-----									
114 Dibromochloromethane						CAS #: 124-48-1			
14.423	14.423	(0.947)	129	5277849	200.000	176.45	70.00- 130.00	100.00	
14.423	14.423	(0.947)	127	4101573			46.96- 106.96	77.71	
-----									
115 1,2-Dibromoethane						CAS #: 106-93-4			
14.588	14.588	(0.958)	107	5509731	200.000	178.14	70.00- 130.00	100.00	
14.588	14.588	(0.958)	109	5152228			62.88- 122.88	93.51	
-----									



AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
124 Chlorobenzene						CAS #: 108-90-7			
15.252	15.252	(1.002)	112	8721021	200.000	170.98	70.00- 130.00	100.00	
15.252	15.252	(1.002)	114	2743027			0.73- 60.73	31.45	
15.252	15.252	(1.002)	77	6567675			40.58- 100.58	75.31	
-----									
127 Ethyl Benzene						CAS #: 100-41-4			
15.363	15.363	(1.009)	106	5041657	200.000	176.06	70.00- 130.00	100.00	
15.363	15.363	(1.009)	91	17579589			309.90- 369.90	348.69	
-----									
128 m,p-Xylene						CAS #: 108-38-3			
15.529	15.529	(1.020)	106	6372305	200.000	173.05	70.00- 130.00	100.00	
15.529	15.529	(1.020)	91	13893912			177.38- 237.38	218.04	
-----									
130 o-Xylene						CAS #: 95-47-6			
16.054	16.054	(1.054)	106	6076266	200.000	171.12	70.00- 130.00	100.00	
16.054	16.054	(1.054)	91	14006948			195.99- 255.99	230.52	
-----									
131 Styrene						CAS #: 100-42-5			
16.082	16.082	(1.056)	104	9762261	200.000	178.61	70.00- 130.00	100.00	
16.082	16.082	(1.056)	78	5787443			28.78- 88.78	59.28	
-----									
133 Bromoform						CAS #: 75-25-2			
16.358	16.358	(1.074)	173	4574387	200.000	185.07	70.00- 130.00	100.00	
16.358	16.358	(1.074)	171	2341683			21.17- 81.17	51.19	
-----									
135 Cumene						CAS #: 98-82-8			
16.524	16.524	(1.085)	105	17970251	200.000	165.84	70.00- 130.00	100.00	
16.524	16.524	(1.085)	120	3974756			0.00- 53.14	22.12	
16.496	16.496	(1.084)	51	2706656			0.00- 45.39	15.06	
-----									
142 1,1,2,2-Tetrachloroethane						CAS #: 79-34-5			
16.966	16.966	(1.114)	83	8927989	200.000	175.54	70.00- 130.00	100.00	
16.966	16.966	(1.114)	85	5676275			34.35- 94.35	63.58	
-----									
143 Propylbenzene						CAS #: 103-65-1			
16.994	16.994	(1.116)	91	19404297	200.000	158.37	70.00- 130.00	100.00	
16.994	16.994	(1.116)	120	3967403			0.00- 48.79	20.45	
16.994	16.994	(1.116)	105	753290			0.00- 33.35	3.88	
-----									
145 4-Ethyltoluene						CAS #: 622-96-8			
17.132	17.132	(1.125)	105	15990495	200.000	150.17	70.00- 130.00	100.00	
17.132	17.132	(1.125)	120	4984716			0.00- 54.57	31.17	
-----									
146 1,3,5-Trimethylbenzene						CAS #: 108-67-8			
17.215	17.215	(1.131)	105	16835111	200.000	167.05	70.00- 130.00	100.00	
17.215	17.215	(1.131)	120	7155884			12.64- 72.64	42.51	
-----									

AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====	=====	
-----										
151	17.602	17.602	(1.156)	105	15608537	200.000	171.04	70.00-	130.00	100.00
	17.602	17.602	(1.156)	120	6302882		11.41-	71.41		40.38
	CAS #: 95-63-6									
-----										
154	17.906	17.906	(1.176)	146	7427639	200.000	187.39	70.00-	130.00	100.00
	17.906	17.906	(1.176)	148	4681195		32.20-	92.20		63.02
	17.879	17.879	(1.174)	111	3455393		13.28-	73.28		46.52
	CAS #: 541-73-1									
-----										
155	17.989	17.989	(1.182)	146	8613025	200.000	170.40	70.00-	130.00	100.00
	17.989	17.989	(1.182)	148	5485543		34.16-	94.16		63.69
	17.989	17.989	(1.182)	111	3763221		16.41-	76.41		43.69
	CAS #: 106-46-7									
-----										
156	18.128	18.128	(1.191)	91	14714814	200.000	201.08	70.00-	130.00	100.00(A)
	18.128	18.128	(1.191)	126	2503464		0.00-	46.97		17.01
	CAS #: 100-44-7									
-----										
158	18.321	18.321	(1.203)	146	7632385	200.000	175.45	70.00-	130.00	100.00
	18.321	18.321	(1.203)	148	4754513		33.83-	93.83		62.29
	18.321	18.321	(1.203)	111	3985203		21.29-	81.29		52.21
	CAS #: 95-50-1									
-----										
163	19.593	19.593	(1.287)	180	5912056	200.000	218.24	70.00-	130.00	100.00(A)
	19.593	19.593	(1.287)	182	5405242		62.06-	122.06		91.43
	CAS #: 120-82-1									
-----										
164	19.676	19.676	(1.292)	225	5815816	200.000	168.08	70.00-	130.00	100.00
	19.676	19.676	(1.292)	223	3716506		35.92-	95.92		63.90
	CAS #: 87-68-3									
-----										
165	19.787	19.787	(1.300)	128	15395818	200.000	177.87	70.00-	130.00	100.00(A)
	19.787	19.787	(1.300)	127	2483525		0.00-	43.72		16.13
	CAS #: 91-20-3									
-----										

QC Flag Legend

A - Target compound detected but, quantitated amount exceeded maximum amount.

Report Date: 23-Mar-2007 08:09

Air Toxics Ltd.

INTERNAL STANDARD COMPOUNDS  
AREA AND RT SUMMARY

Instrument ID: msd8.i

Calibration Date: 22-MAR-2007

Lab File ID: 8032209.d

Calibration Time: 12:34

Lab Smp Id: ICAL

Client Smp ID: Level 7

Analysis Type: VOA

Level: LOW

Quant Type: ISTD

Sample Type: AIR

Operator: sjr

Method File: /chem/msd8.i/8-22mar.b/t14q322a.m

Misc Info: 200ppbv-&gt;200ppbv

COMPOUND	STANDARD	AREA LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	283735	170241	397229	279518	-1.49
86 1,4-Difluorobenze	1370859	822515	1919203	1399810	2.11
123 Chlorobenzene-d5	1067063	640238	1493888	1085707	1.75

COMPOUND	STANDARD	RT LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	8.39	8.06	8.72	8.40	0.00
86 1,4-Difluorobenze	10.25	9.92	10.58	10.25	0.00
123 Chlorobenzene-d5	15.22	14.89	15.55	15.22	0.00

AREA UPPER LIMIT = + 40% of internal standard area.

AREA LOWER LIMIT = - 40% of internal standard area.


RT UPPER LIMIT = + 0.33 minutes of internal standard RT.

RT LOWER LIMIT = - 0.33 minutes of internal standard RT.

Data File: /chem/msd8.1/8-22mar.1b/8032209.d  
Date: 22-MAR-2007 13:30  
Client ID: Level 7  
Sample Info: 200ml 1487-115

Column phase: RTX-624

Instrument: msd8.1  
Operator: sjr  
Column diameter: 0.53





AN ENVIRONMENTAL ANALYTICAL LABORATORY

Client Sample ID: CCV

Lab ID#: 0704143-04A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	8041202	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 4/12/07 09:53 AM

Compound	%Recovery
Freon 12	80
Freon 114	95
Vinyl Chloride	100
Bromomethane	89
Chloroethane	86
Freon 11	95
1,1-Dichloroethene	90
Freon 113	91
Methylene Chloride	90
1,1-Dichloroethane	88
cis-1,2-Dichloroethene	88
Chloroform	75
1,1,1-Trichloroethane	82
Carbon Tetrachloride	88
Benzene	82
1,2-Dichloroethane	93
Trichloroethene	88
1,2-Dichloropropane	83
cis-1,3-Dichloropropene	85
Toluene	86
trans-1,3-Dichloropropene	89
1,1,2-Trichloroethane	87
Tetrachloroethene	94
1,2-Dibromoethane (EDB)	92
Chlorobenzene	87
Ethyl Benzene	85
m,p-Xylene	85
o-Xylene	86
Styrene	82
1,1,2,2-Tetrachloroethane	82
1,3,5-Trimethylbenzene	88
1,2,4-Trimethylbenzene	86
1,3-Dichlorobenzene	95
1,4-Dichlorobenzene	96
alpha-Chlorotoluene	84
1,2-Dichlorobenzene	91
1,3-Butadiene	95
Hexane	84
Cyclohexane	79



AN ENVIRONMENTAL ANALYTICAL LABORATORY

Client Sample ID: CCV

Lab ID#: 0704143-04A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	8041202	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 4/12/07 09:53 AM

Compound	%Recovery
Heptane	75
Bromodichloromethane	86
Dibromochloromethane	93
Cumene	81
Propylbenzene	85
Chloromethane	93
1,2,4-Trichlorobenzene	81
Hexachlorobutadiene	101
Acetone	89
Carbon Disulfide	90
2-Propanol	93
trans-1,2-Dichloroethene	85
2-Butanone (Methyl Ethyl Ketone)	77
Tetrahydrofuran	72
1,4-Dioxane	77
4-Methyl-2-pentanone	74
2-Hexanone	83
Bromoform	96
4-Ethyltoluene	88
Ethanol	98
Methyl tert-butyl ether	84
3-Chloropropene	95
2,2,4-Trimethylpentane	82
Naphthalene	85

Container Type: NA - Not Applicable

Surrogates	%Recovery	Method Limits
Toluene-d8	99	70-130
1,2-Dichloroethane-d4	95	70-130
4-Bromofluorobenzene	102	70-130

Report Date: 12-Apr-2007 11:42

## Air Toxics Ltd.

## CONTINUING CALIBRATION COMPOUNDS

Instrument ID: msd8.i                      Injection Date: 12-APR-2007 09:53  
 Lab File ID: 8041202.d                    Init. Cal. Date(s): 22-MAR-2007 26-MAR-2007  
 Analysis Type: AIR                        Init. Cal. Times: 10:41                    13:11  
 Lab Sample ID: CCV-1                      Quant Type: ISTD  
 Method: /var/chem/msd8.i/8-12apr.b/t14q322b.m

COMPOUND	RRF / AMOUNT	RF50	MIN	MAX	CURVE TYPE	
			RRF	%D / %DRIFT	%D / %DRIFT	
\$ 80 1,2-Dichloroethane-d4	2.16099	2.06361	0.010	4.50641	30.00000	Averaged
\$ 102 Toluene-d8	0.98315	0.96956	0.010	1.38260	30.00000	Averaged
\$ 138 Bromofluorobenzene	0.51234	0.52331	0.010	-2.14293	30.00000	Averaged
3 Propylene	2.64519	2.68608	0.010	-1.54565	30.00000	Averaged
4 Dichlorodifluoromethane/Fr1	5.59128	4.45601	0.010	20.30434	30.00000	Averaged
6 Freon 114	3.17116	3.02677	0.010	4.55296	30.00000	Averaged
8 Chloromethane	2.87665	2.67358	0.010	7.05912	30.00000	Averaged
10 Vinyl Chloride	2.70205	2.70821	0.010	-0.22785	30.00000	Averaged
11 1,3-Butadiene	2.80201	2.67008	0.010	4.70843	30.00000	Averaged
13 Bromomethane	1.47708	1.31079	0.010	11.25756	30.00000	Averaged
16 Chloroethane	1.40708	1.20698	0.010	14.22091	30.00000	Averaged
18 Trichlorofluoromethane/Fr11	5.00821	4.73880	0.010	5.37940	30.00000	Averaged
21 Ethanol	1.13689	1.11619	0.010	1.82109	30.00000	Averaged
27 Freon 113	2.37158	2.15427	0.010	9.16314	30.00000	Averaged
29 1,1-Dichloroethene	4.03682	3.62202	0.010	10.27547	30.00000	Averaged
30 Acetone	1.44376	1.27945	0.010	11.38037	30.00000	Averaged
34 2-Propanol	5.65645	5.27107	0.010	6.81307	30.00000	Averaged
33 Carbon Disulfide	6.21805	5.62212	0.010	9.58382	30.00000	Averaged
37 3-Chloropropene	1.00414	0.95300	0.010	5.09274	30.00000	Averaged
39 Methylene Chloride	3.43297	3.08557	0.010	10.11937	30.00000	Averaged
42 MTBE	6.91999	5.82586	0.010	15.81112	30.00000	Averaged
43 trans-1,2-Dichloroethene	2.06297	1.76323	0.010	14.52952	30.00000	Averaged
45 Hexane	5.07809	4.29420	0.010	15.43672	30.00000	Averaged
54 Vinyl Acetate	0.51018	0.49178	0.010	3.60676	30.00000	Averaged
52 1,1-Dichloroethane	4.28903	3.78004	0.010	11.86719	30.00000	Averaged
64 2-Butanone	1.17072	0.90608	0.010	22.60486	30.00000	Averaged
63 cis-1,2-Dichloroethene	3.26321	2.85736	0.010	12.43740	30.00000	Averaged
66 Tetrahydrofuran	4.35898	3.12440	0.010	28.32258	30.00000	Averaged
69 Chloroform	4.11103	3.08145	0.010	25.04440	30.00000	Averaged
73 1,1,1-Trichloroethane	3.71165	3.05974	0.010	17.56386	30.00000	Averaged
72 Cyclohexane	3.06440	2.42350	0.010	20.91436	30.00000	Averaged
75 Carbon Tetrachloride	2.99797	2.62591	0.010	12.41039	30.00000	Averaged
78 2,2,4-Trimethylpentane	14.29843	11.66494	0.010	18.41806	30.00000	Averaged
79 Benzene	1.39238	1.13490	0.010	18.49225	30.00000	Averaged
81 1,2-Dichloroethane	0.65206	0.60843	0.010	6.69103	30.00000	Averaged

Air Toxics Ltd.

CONTINUING CALIBRATION COMPOUNDS

Instrument ID: msd8.i                    Injection Date: 12-APR-2007 09:53  
 Lab File ID: 8041202.d                Init. Cal. Date(s): 22-MAR-2007 26-MAR-2007  
 Analysis Type: AIR                    Init. Cal. Times: 10:41 13:11  
 Lab Sample ID: CCV-1                 Quant Type: ISTD  
 Method: /var/chem/msd8.i/8-12apr.b/t14q322b.m

COMPOUND	RRF / AMOUNT	RF50	MIN	MAX	CURVE TYPE
			RRF   %D / %DRIFT	%D / %DRIFT	
82 Heptane	0.16632	0.12450	0.010   25.14338	30.00000	Averaged
92 Trichloroethene	0.44787	0.39585	0.010   11.61321	30.00000	Averaged
95 1,2-Dichloropropane	0.51024	0.42177	0.010   17.33923	30.00000	Averaged
96 1,4-Dioxane	0.26600	0.20432	0.010   23.18580	30.00000	Averaged
98 Bromodichloromethane	0.75448	0.64670	0.010   14.28560	30.00000	Averaged
100 cis-1,3-Dichloropropene	0.64347	0.54716	0.010   14.96847	30.00000	Averaged
101 4-Methyl-2-pentanone	0.51882	0.38626	0.010   25.55014	30.00000	Averaged
103 Toluene	1.26076	1.08224	0.010   14.15995	30.00000	Averaged
106 trans-1,3-Dichloropropene	0.87181	0.77491	0.010   11.11420	30.00000	Averaged
108 1,1,2-Trichloroethane	0.50754	0.44013	0.010   13.28331	30.00000	Averaged
109 Tetrachloroethene	0.53907	0.50752	0.010   5.85300	30.00000	Averaged
112 2-Hexanone	0.80104	0.66393	0.010   17.11628	30.00000	Averaged
114 Dibromochloromethane	0.68875	0.63938	0.010   7.16722	30.00000	Averaged
115 1,2-Dibromoethane	0.71217	0.65669	0.010   7.78980	30.00000	Averaged
124 Chlorobenzene	1.17450	1.02668	0.010   12.58595	30.00000	Averaged
127 Ethyl Benzene	0.65940	0.56116	0.010   14.89818	30.00000	Averaged
128 m,p-Xylene	0.84792	0.72100	0.010   14.96910	30.00000	Averaged
130 o-Xylene	0.81763	0.70004	0.010   14.38145	30.00000	Averaged
131 Styrene	1.25857	1.03417	0.010   17.82982	30.00000	Averaged
133 Bromoform	0.56916	0.54398	0.010   4.42341	30.00000	Averaged
142 1,1,2,2-Tetrachloroethane	1.17112	0.96223	0.010   17.83668	30.00000	Averaged
145 4-Ethyltoluene	2.45190	2.16435	0.010   11.72775	30.00000	Averaged
146 1,3,5-Trimethylbenzene	2.32061	2.03438	0.010   12.33429	30.00000	Averaged
151 1,2,4-Trimethylbenzene	2.10131	1.81941	0.010   13.41546	30.00000	Averaged
154 1,3-Dichlorobenzene	0.91270	0.87129	0.010   4.53626	30.00000	Averaged
155 1,4-Dichlorobenzene	1.16387	1.11620	0.010   4.09533	30.00000	Averaged
156 alpha-Chlorotoluene	1.68507	1.41580	0.010   15.97961	30.00000	Averaged
158 1,2-Dichlorobenzene	1.00169	0.91615	0.010   8.53963	30.00000	Averaged
163 1,2,4-Trichlorobenzene	0.62377	0.50751	0.010   18.63806	30.00000	Averaged
164 Hexachlorobutadiene	0.79677	0.80648	0.010   -1.21795	30.00000	Averaged
143 Propylbenzene	2.82133	2.40942	0.010   14.59987	30.00000	Averaged
135 Cumene	2.49512	2.03265	0.010   18.53483	30.00000	Averaged
165 Naphthalene	1.99311	1.69111	0.010   15.15198	30.00000	Averaged



Report Date: 12-Apr-2007 11:42

## Air Toxics Ltd.

## AMBIENT AIR METHOD TO14A/TO15

Data file : /chem/msd8.i/8-12apr.b/8041202.d  
 Lab Smp Id: CCV-1 Client Smp ID: CCV-1  
 Inj Date : 12-APR-2007 09:53  
 Operator : JG Inst ID: msd8.i  
 Smp Info : 100ml #1408-387A  
 Misc Info : 100ppbv-50ppbv  
 Comment :  
 Method : /var/chem/msd8.i/8-12apr.b/t14q322b.m  
 Meth Date : 12-Apr-2007 11:41 jgray Quant Type: ISTD  
 Cal Date : 26-MAR-2007 13:11 Cal File: 8032608.d  
 Als bottle: 1 Continuing Calibration Sample  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: AT04+Na.sub  
 Target Version: 3.50 Sample Matrix: AIR  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* CpndVariable

Cpnd Variable Local Compound Variable

AMOUNTS								
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT	ON-COL	TARGET RANGE	RATIO
==	=====	=====	=====	=====	=====	=====	=====	=====
-----								
* 67	Bromochloromethane						CAS #: 74-97-5	
8.395	8.395	(1.000)	130	269481	25.0000		80.00- 120.00	100.00
8.395	8.395	(1.000)	128	203942			45.68- 105.68	75.68
8.395	8.395	(1.000)	49	804339			268.48- 328.48	298.48
-----								
* 86	1,4-Difluorobenzene						CAS #: 540-36-3	
10.275	10.275	(1.000)	114	1204367	25.0000		80.00- 120.00	100.00
10.248	10.248	(1.000)	88	216932			0.00- 48.01	18.01
-----								
* 123	Chlorobenzene-d5						CAS #: 3114-55-4	
15.225	15.225	(1.000)	117	906226	25.0000		80.00- 120.00	100.00
15.225	15.225	(1.000)	82	619586			35.49- 95.49	68.37
-----								
\$ 80	1,2-Dichloroethane-d4						CAS #: 17060-07-0	
9.474	9.474	(1.128)	65	556103	25.0000	23.873	80.00- 120.00	100.00
9.474	9.474	(1.128)	67	306280			27.92- 87.92	55.08
-----								
\$ 102	Toluene-d8						CAS #: 2037-26-5	
12.985	12.985	(1.264)	98	1167702	25.0000	24.654	80.00- 120.00	100.00
12.985	12.985	(1.264)	70	141822			0.00- 42.61	12.15

AMOUNTS										
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO		
==	=====	=====	=====	=====	=====	=====	=====	=====		
\$ 102 Toluene-d8 (continued)										
12.985	12.985	(1.264)	100	820302			40.27- 100.27	70.25		
-----										
\$ 138 Bromofluorobenzene										
						CAS #: 460-00-4				
16.773	16.773	(1.102)	174	474241	25.0000	25.536	80.00- 120.00	100.00		
16.745	16.745	(1.100)	95	797167			138.09- 198.09	168.09		
16.773	16.773	(1.102)	176	472028			69.53- 129.53	99.53		
-----										
3 Propylene						CAS #: 115-07-1				
2.451	2.451	(0.292)	41	1447693	50.0000	50.773	80.00- 120.00	100.00		
2.451	2.451	(0.292)	42	966284			35.20- 95.20	66.75		
2.451	2.451	(0.292)	39	1045601			42.80- 102.80	72.23		
-----										
4 Dichlorodifluoromethane/Fr12						CAS #: 75-71-8				
2.506	2.506	(0.298)	85	2401619	50.0000	39.848	80.00- 120.00	100.00		
2.506	2.506	(0.298)	87	767232			1.89- 61.89	31.95		
-----										
6 Freon 114						CAS #: 76-14-2				
2.644	2.644	(0.315)	135	1631316	50.0000	47.724	80.00- 120.00	100.00		
2.644	2.644	(0.315)	137	501306			0.73- 60.73	30.73		
-----										
8 Chloromethane						CAS #: 74-87-3				
2.782	2.782	(0.331)	50	1440958	50.0000	46.470	80.00- 120.00	100.00		
2.782	2.782	(0.331)	52	432814			0.00- 59.46	30.04		
-----										
10 Vinyl Chloride						CAS #: 75-01-4				
2.948	2.948	(0.351)	62	1459623	50.0000	50.114	80.00- 120.00	100.00		
2.948	2.948	(0.351)	64	430203			0.33- 60.33	29.47		
-----										
11 1,3-Butadiene						CAS #: 106-99-0				
2.921	2.921	(0.348)	54	1439069	50.0000	47.646	80.00- 120.00	100.00		
2.921	2.921	(0.348)	39	2032206			86.60- 146.60	141.22		
-----										
13 Bromomethane						CAS #: 74-83-9				
3.474	3.474	(0.414)	94	706468	50.0000	44.371	80.00- 120.00	100.00		
3.474	3.474	(0.414)	96	637728			60.27- 120.27	90.27		
-----										
16 Chloroethane						CAS #: 75-00-3				
3.640	3.640	(0.434)	64	650517	50.0000	42.890	80.00- 120.00	100.00		
3.612	3.612	(0.430)	49	210392			2.42- 62.42	32.34		
3.640	3.640	(0.434)	66	196969			0.00- 58.92	30.28		
-----										
18 Trichlorofluoromethane/Fr11						CAS #: 75-69-4				
3.916	3.916	(0.466)	101	2554032	50.0000	47.310	80.00- 120.00	100.00		
3.916	3.916	(0.466)	103	1606507			32.90- 92.90	62.90		
-----										

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
21 Ethanol						CAS #: 64-17-5			
4.331	4.331	(0.516)	45	601582	50.0000	49.089	80.00- 120.00	100.00	
4.331	4.331	(0.516)	43	111463			0.00- 50.43	18.53	
4.303	4.303	(0.513)	46	250703			12.21- 72.21	41.67	
-----									
27 Freon 113						CAS #: 76-13-1			
4.773	4.773	(0.569)	151	1161069	50.0000	45.418	80.00- 120.00	100.00	
4.773	4.773	(0.569)	153	746129			34.26- 94.26	64.26	
4.773	4.773	(0.569)	101	1806663			125.60- 185.60	155.60	
-----									
29 1,1-Dichloroethene						CAS #: 75-35-4			
4.801	4.801	(0.572)	61	1952130	50.0000	44.862	80.00- 120.00	100.00	
4.801	4.801	(0.572)	96	897244			15.96- 75.96	45.96	
4.801	4.801	(0.572)	98	549425			0.00- 58.14	28.14	
-----									
30 Acetone						CAS #: 67-64-1			
4.967	4.967	(0.592)	58	689576	50.0000	44.310	80.00- 120.00	100.00	
4.967	4.967	(0.592)	43	2502029			337.35- 397.35	362.84	
-----									
34 2-Propanol						CAS #: 67-63-0			
5.188	5.188	(0.618)	45	2840907	50.0000	46.593	80.00- 120.00	100.00	
5.188	5.188	(0.618)	43	577708			0.00- 50.64	20.34	
5.188	5.188	(0.618)	59	87500			0.00- 33.21	3.08	
-----									
33 Carbon Disulfide						CAS #: 75-15-0			
5.160	5.160	(0.615)	76	3030110	50.0000	45.208	80.00- 120.00	100.00	
-----									
37 3-Chloropropene						CAS #: 107-05-1			
5.464	5.464	(0.651)	76	513630	50.0000	47.454	80.00- 120.00	100.00	
5.464	5.464	(0.651)	41	2288906			414.53- 474.53	445.63	
-----									
39 Methylene Chloride						CAS #: 75-09-2			
5.741	5.741	(0.684)	49	1663006	50.0000	44.940	80.00- 120.00	100.00	
5.741	5.741	(0.684)	84	790473			17.53- 77.53	47.53	
5.741	5.741	(0.684)	51	486663			0.00- 59.93	29.26	
-----									
42 MTBE						CAS #: 1634-04-4			
6.073	6.073	(0.723)	73	3139917	50.0000	42.094	80.00- 120.00	100.00	
6.073	6.073	(0.723)	57	924237			0.00- 59.44	29.44	
6.073	6.073	(0.723)	41	1017827			1.95- 61.95	32.42	
-----									
43 trans-1,2-Dichloroethene						CAS #: 156-60-5			
6.128	6.128	(0.730)	96	950313	50.0000	42.735	80.00- 120.00	100.00	
6.128	6.128	(0.730)	61	1853202			165.01- 225.01	195.01	
6.128	6.128	(0.730)	98	602613			34.71- 94.71	63.41	
-----									

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
-----									
45 Hexane						CAS #: 110-54-3			
6.460	6.460	(0.769)	57	2314411	50.0000	42.282	80.00- 120.00	100.00	
6.460	6.460	(0.769)	43	1747963			44.61- 104.61	75.53	
6.460	6.460	(0.769)	86	323171			0.00- 43.77	13.96	
-----									
54 Vinyl Acetate						CAS #: 108-05-4			
6.985	6.985	(0.832)	86	265051	50.0000	48.197	80.00- 120.00	100.00	
6.985	6.985	(0.832)	43	4117795			1469.62-1529.62	1553.59	
6.985	6.985	(0.832)	42	348978			100.08- 160.08	131.66	
-----									
52 1,1-Dichloroethane						CAS #: 75-34-3			
6.930	6.930	(0.825)	63	2037300	50.0000	44.066	80.00- 120.00	100.00	
6.902	6.902	(0.822)	65	634842			1.16- 61.16	31.16	
-----									
64 2-Butanone						CAS #: 78-93-3			
8.008	8.008	(0.954)	72	488343	50.0000	38.698	80.00- 120.00	100.00	
8.008	8.008	(0.954)	43	2943365			572.72- 632.72	602.72	
8.008	8.008	(0.954)	57	195653			10.61- 70.61	40.06	
-----									
63 cis-1,2-Dichloroethene						CAS #: 156-59-2			
7.953	7.953	(0.947)	61	1540006	50.0000	43.781	80.00- 120.00	100.00	
7.953	7.953	(0.947)	96	839253			24.50- 84.50	54.50	
7.953	7.953	(0.947)	98	521411			3.86- 63.86	33.86	
-----									
66 Tetrahydrofuran						CAS #: 109-99-9			
8.368	8.368	(0.997)	42	1683934	50.0000	35.839	80.00- 120.00	100.00	
8.368	8.368	(0.997)	71	429513			0.00- 55.51	25.51	
8.368	8.368	(0.997)	72	435565			0.00- 56.52	25.87	
-----									
69 Chloroform						CAS #: 67-66-3			
8.533	8.533	(1.016)	83	1660785	50.0000	37.478	80.00- 120.00	100.00	
8.533	8.533	(1.016)	85	1049589			33.20- 93.20	63.20	
-----									
73 1,1,1-Trichloroethane						CAS #: 71-55-6			
8.782	8.782	(1.046)	97	1649084	50.0000	41.218	80.00- 120.00	100.00	
8.782	8.782	(1.046)	99	1078808			35.42- 95.42	65.42	
-----									
72 Cyclohexane						CAS #: 110-82-7			
8.755	8.755	(1.043)	84	1306174	50.0000	39.543	80.00- 120.00	100.00	
8.755	8.755	(1.043)	56	2191970			137.82- 197.82	167.82	
8.755	8.755	(1.043)	41	1306608			70.03- 130.03	100.03	
-----									
75 Carbon Tetrachloride						CAS #: 56-23-5			
9.004	9.004	(1.072)	119	1415264	50.0000	43.795	80.00- 120.00	100.00	
9.004	9.004	(1.072)	117	1568713			80.84- 140.84	110.84	
-----									

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
-----									
78	2,2,4-Trimethylpentane					CAS #: 540-84-1			
9.446	9.446	(1.125)	57	6286959	50.0000	40.791	80.00- 120.00	100.00	
9.446	9.446	(1.125)	56	2180713			4.45- 64.45	34.69	
9.446	9.446	(1.125)	41	1812142			0.00- 59.44	28.82	
-----									
79	Benzene					CAS #: 71-43-2			
9.446	9.446	(0.919)	78	2733672	50.0000	40.754	80.00- 120.00	100.00	
9.446	9.446	(0.919)	77	655120			0.00- 54.39	23.96	
-----									
81	1,2-Dichloroethane					CAS #: 107-06-2			
9.612	9.612	(0.935)	62	1465556	50.0000	46.654	80.00- 120.00	100.00	
9.612	9.612	(0.935)	64	447256			1.44- 61.44	30.52	
-----									
82	Heptane					CAS #: 142-82-5			
9.833	9.833	(0.957)	100	299895	50.0000	37.428	80.00- 120.00	100.00	
9.833	9.833	(0.957)	43	2615242			789.66- 849.66	872.05	
9.833	9.833	(0.957)	71	1012656			294.27- 354.27	337.67	
-----									
92	Trichloroethene					CAS #: 79-01-6			
10.690	10.690	(1.040)	95	953506	50.0000	44.193	80.00- 120.00	100.00	
10.690	10.690	(1.040)	130	844049			58.52- 118.52	88.52	
10.690	10.690	(1.040)	97	594482			32.35- 92.35	62.35	
-----									
95	1,2-Dichloropropane					CAS #: 78-87-5			
11.188	11.188	(1.089)	63	1015923	50.0000	41.330	80.00- 120.00	100.00	
11.188	11.188	(1.089)	62	766641			45.46- 105.46	75.46	
11.188	11.188	(1.089)	41	800850			48.83- 108.83	78.83	
-----									
96	1,4-Dioxane					CAS #: 123-91-1			
11.409	11.409	(1.110)	88	492158	50.0000	38.407	80.00- 120.00	100.00	
11.409	11.409	(1.110)	58	490538			69.67- 129.67	99.67	
11.409	11.409	(1.110)	57	156888			0.54- 60.54	31.88	
-----									
98	Bromodichloromethane					CAS #: 75-27-4			
11.741	11.741	(1.143)	83	1557725	50.0000	42.857	80.00- 120.00	100.00	
11.741	11.741	(1.143)	85	993840			33.80- 93.80	63.80	
-----									
100	cis-1,3-Dichloropropene					CAS #: 10061-01-5			
12.626	12.626	(1.229)	75	1317951	50.0000	42.516	80.00- 120.00	100.00	
12.626	12.626	(1.229)	77	413228			1.35- 61.35	31.35	
12.626	12.626	(1.229)	39	1036219			48.62- 108.62	78.62	
-----									
101	4-Methyl-2-pentanone					CAS #: 108-10-1			
12.902	12.902	(1.256)	58	930393	50.0000	37.225	80.00- 120.00	100.00	
12.902	12.902	(1.256)	43	2728228			251.14- 311.14	293.23	
12.902	12.902	(1.256)	85	307371			2.33- 62.33	33.04	
-----									

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
-----									
103	Toluene					CAS #:	108-88-3		
13.096	13.096	(1.274)	91	2606823	50.0000	42.920	80.00-	120.00	100.00
13.096	13.096	(1.274)	92	1574935			30.42-	90.42	60.42
-----									
106	trans-1,3-Dichloropropene					CAS #:	10061-02-6		
13.621	13.621	(0.895)	75	1404489	50.0000	44.443	80.00-	120.00	100.00
13.621	13.621	(0.895)	77	451623			2.16-	62.16	32.16
13.621	13.621	(0.895)	39	1045806			44.46-	104.46	74.46
-----									
108	1,1,2-Trichloroethane					CAS #:	79-00-5		
13.897	13.897	(0.913)	97	797706	50.0000	43.358	80.00-	120.00	100.00
13.897	13.897	(0.913)	99	497998			32.43-	92.43	62.43
13.897	13.897	(0.913)	83	707971			58.75-	118.75	88.75
-----									
109	Tetrachloroethene					CAS #:	127-18-4		
13.925	13.925	(0.915)	166	919857	50.0000	47.073	80.00-	120.00	100.00
13.925	13.925	(0.915)	129	725435			48.86-	108.86	78.86
13.925	13.925	(0.915)	131	683438			44.30-	104.30	74.30
-----									
112	2-Hexanone					CAS #:	591-78-6		
14.257	14.257	(0.936)	58	1203341	50.0000	41.442	80.00-	120.00	100.00
14.257	14.257	(0.936)	43	2609991			186.90-	246.90	216.90
14.257	14.257	(0.936)	100	197485			0.00-	45.68	16.41
-----									
114	Dibromochloromethane					CAS #:	124-48-1		
14.423	14.423	(0.947)	129	1158849	50.0000	46.416	80.00-	120.00	100.00
14.423	14.423	(0.947)	127	914575			46.96-	106.96	78.92
-----									
115	1,2-Dibromoethane					CAS #:	106-93-4		
14.589	14.589	(0.958)	107	1190222	50.0000	46.105	80.00-	120.00	100.00
14.589	14.589	(0.958)	109	1101156			62.52-	122.52	92.52
-----									
124	Chlorobenzene					CAS #:	108-90-7		
15.252	15.252	(1.002)	112	1860808	50.0000	43.707	80.00-	120.00	100.00
15.252	15.252	(1.002)	114	586218			1.50-	61.50	31.50
15.252	15.252	(1.002)	77	1344145			42.23-	102.23	72.23
-----									
127	Ethyl Benzene					CAS #:	100-41-4		
15.363	15.363	(1.009)	106	1017073	50.0000	42.551	80.00-	120.00	100.00
15.363	15.363	(1.009)	91	3468101			309.90-	369.90	340.99
-----									
128	m,p-Xylene					CAS #:	108-38-3		
15.529	15.529	(1.020)	106	1306774	50.0000	42.515	80.00-	120.00	100.00
15.529	15.529	(1.020)	91	2816721			177.38-	237.38	215.55
-----									
130	o-Xylene					CAS #:	95-47-6		
16.054	16.054	(1.054)	106	1268791	50.0000	42.809	80.00-	120.00	100.00

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
130 o-Xylene (continued)									
16.054	16.054	(1.054)	91	2828593			192.94- 252.94	222.94	
-----									
131 Styrene CAS #: 100-42-5									
16.109	16.109	(1.058)	104	1874387	50.0000	41.085	80.00- 120.00	100.00	
16.082	16.082	(1.056)	78	1125651			30.05- 90.05	60.05	
-----									
133 Bromoform CAS #: 75-25-2									
16.358	16.358	(1.074)	173	985942	50.0000	47.788	80.00- 120.00	100.00	
16.358	16.358	(1.074)	171	500988			20.81- 80.81	50.81	
-----									
142 1,1,2,2-Tetrachloroethane CAS #: 79-34-5									
16.967	16.967	(1.114)	83	1744004	50.0000	41.082	80.00- 120.00	100.00	
16.967	16.967	(1.114)	85	1119648			34.20- 94.20	64.20	
-----									
145 4-Ethyltoluene CAS #: 622-96-8									
17.132	17.132	(1.125)	105	3922775	50.0000	44.136	80.00- 120.00	100.00	
17.132	17.132	(1.125)	120	976533			0.00- 54.89	24.89	
-----									
146 1,3,5-Trimethylbenzene CAS #: 108-67-8									
17.215	17.215	(1.131)	105	3687216	50.0000	43.833	80.00- 120.00	100.00	
17.215	17.215	(1.131)	120	1527275			12.64- 72.64	41.42	
-----									
151 1,2,4-Trimethylbenzene CAS #: 95-63-6									
17.602	17.602	(1.156)	105	3297592	50.0000	43.292	80.00- 120.00	100.00	
17.602	17.602	(1.156)	120	1310668			9.75- 69.75	39.75	
-----									
154 1,3-Dichlorobenzene CAS #: 541-73-1									
17.907	17.907	(1.176)	146	1579178	50.0000	47.732	80.00- 120.00	100.00	
17.907	17.907	(1.176)	148	1013361			32.20- 92.20	64.17	
17.907	17.907	(1.176)	111	682316			13.28- 73.28	43.21	
-----									
155 1,4-Dichlorobenzene CAS #: 106-46-7									
17.990	17.990	(1.182)	146	2023062	50.0000	47.952	80.00- 120.00	100.00	
17.990	17.990	(1.182)	148	1262465			34.16- 94.16	62.40	
17.990	17.990	(1.182)	111	935359			16.41- 76.41	46.23	
-----									
156 alpha-Chlorotoluene CAS #: 100-44-7									
18.128	18.128	(1.191)	91	2566068	50.0000	42.010	80.00- 120.00	100.00	
18.128	18.128	(1.191)	126	434938			0.00- 46.97	16.95	
-----									
158 1,2-Dichlorobenzene CAS #: 95-50-1									
18.321	18.321	(1.203)	146	1660471	50.0000	45.730	80.00- 120.00	100.00	
18.321	18.321	(1.203)	148	1037318			32.47- 92.47	62.47	
18.321	18.321	(1.203)	111	845360			20.91- 80.91	50.91	
-----									

AMOUNTS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	CAL-AMT ( PPEV)	ON-COL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
-----									
163	1,2,4-Trichlorobenzene					CAS #: 120-82-1			
19.593	19.593	(1.287)	180	919840	50.0000	40.681	80.00- 120.00	100.00	
19.593	19.593	(1.287)	182	880119			65.68- 125.68	95.68	
-----									
164	Hexachlorobutadiene					CAS #: 87-68-3			
19.676	19.676	(1.292)	225	1461701	50.0000	50.609	80.00- 120.00	100.00	
19.676	19.676	(1.292)	223	928590			33.53- 93.53	63.53	
-----									
143	Propylbenzene					CAS #: 103-65-1			
16.994	16.994	(1.116)	91	4366955	50.0000	42.700	80.00- 120.00	100.00	
16.994	16.994	(1.116)	120	820542			0.00- 48.79	18.79	
16.994	16.994	(1.116)	105	157247			0.00- 33.35	3.60	
-----									
135	Cumene					CAS #: 98-82-8			
16.524	16.524	(1.085)	105	3684084	50.0000	40.732	80.00- 120.00	100.00	
16.524	16.524	(1.085)	120	825219			0.00- 53.14	22.40	
16.497	16.497	(1.084)	51	570181			0.00- 45.39	15.48	
-----									
165	Naphthalene					CAS #: 91-20-3			
19.787	19.787	(1.300)	128	1532529	25.0000	21.212	80.00- 120.00	100.00	
19.787	19.787	(1.300)	127	220121			0.00- 43.72	14.36	
-----									



Report Date: 12-Apr-2007 11:42

Air Toxics Ltd.

INTERNAL STANDARD COMPOUNDS  
AREA AND RT SUMMARY

Instrument ID: msd8.i

Calibration Date: 12-APR-2007

Lab File ID: 8041202.d

Calibration Time: 09:53

Lab Smp Id: CCV-1

Client Smp ID: CCV-1

Analysis Type: VOA

Level: LOW

Quant Type: ISTD

Sample Type: AIR

Operator: JG

Method File: /var/chem/msd8.i/8-12apr.b/t14q322b.m

Misc Info: 100ppbv-50ppbv

COMPOUND	STANDARD	AREA LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	269481	161689	377273	269481	0.00
86 1,4-Difluorobenze	1204367	722620	1686114	1204367	0.00
123 Chlorobenzene-d5	906226	543736	1268716	906226	0.00

COMPOUND	STANDARD	RT LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	8.40	8.07	8.73	8.40	0.00
86 1,4-Difluorobenze	10.28	9.95	10.61	10.28	0.00
123 Chlorobenzene-d5	15.22	14.89	15.55	15.22	0.00

AREA UPPER LIMIT = + 40% of internal standard area.

AREA LOWER LIMIT = - 40% of internal standard area.

RT UPPER LIMIT = + 0.33 minutes of internal standard RT.


RT LOWER LIMIT = - 0.33 minutes of internal standard RT.

Data File: /chem/msd8.1/8-12apr.1b/8041202.d  
Date: 12-APR-2007 09:53  
Client ID: CCV-1  
Sample Info: 100ml #1408-387A

Column phase: RTX-624

Instrument: msd8.1  
Operator: JG  
Column diameter: 0.53

/chem/msd8.1/8-12apr.1b/8041202.d





AN ENVIRONMENTAL ANALYTICAL LABORATORY

Client Sample ID: LCS

Lab ID#: 0704143-05A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	8041203	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 4/12/07 10:21 AM

Compound	%Recovery
Freon 12	90
Freon 114	95
Vinyl Chloride	101
Bromomethane	99
Chloroethane	98
Freon 11	99
1,1-Dichloroethene	109
Freon 113	110
Methylene Chloride	108
1,1-Dichloroethane	102
cis-1,2-Dichloroethene	100
Chloroform	85
1,1,1-Trichloroethane	95
Carbon Tetrachloride	101
Benzene	86
1,2-Dichloroethane	101
Trichloroethene	94
1,2-Dichloropropane	87
cis-1,3-Dichloropropene	91
Toluene	94
trans-1,3-Dichloropropene	94
1,1,2-Trichloroethane	93
Tetrachloroethene	103
1,2-Dibromoethane (EDB)	93
Chlorobenzene	94
Ethyl Benzene	94
m,p-Xylene	90
o-Xylene	92
Styrene	88
1,1,2,2-Tetrachloroethane	88
1,3,5-Trimethylbenzene	94
1,2,4-Trimethylbenzene	91
1,3-Dichlorobenzene	110
1,4-Dichlorobenzene	104
alpha-Chlorotoluene	109
1,2-Dichlorobenzene	104
1,3-Butadiene	92
Hexane	96
Cyclohexane	88



AN ENVIRONMENTAL ANALYTICAL LABORATORY

Client Sample ID: LCS

Lab ID#: 0704143-05A

MODIFIED EPA METHOD TO-15 GC/MS FULL SCAN

File Name:	8041203	Date of Collection: NA
Dil. Factor:	1.00	Date of Analysis: 4/12/07 10:21 AM

Compound	%Recovery
Heptane	82
Bromodichloromethane	93
Dibromochloromethane	102
Cumene	91
Propylbenzene	99
Chloromethane	102
1,2,4-Trichlorobenzene	116
Hexachlorobutadiene	100
Acetone	106
Carbon Disulfide	99
2-Propanol	100
trans-1,2-Dichloroethene	96
2-Butanone (Methyl Ethyl Ketone)	87
Tetrahydrofuran	83
1,4-Dioxane	84
4-Methyl-2-pentanone	81
2-Hexanone	87
Bromoform	107
4-Ethyltoluene	101
Ethanol	102
Methyl tert-butyl ether	96
3-Chloropropene	103
2,2,4-Trimethylpentane	90
Naphthalene	131

Container Type: NA - Not Applicable

Surrogates	%Recovery	Method Limits
Toluene-d8	96	70-130
1,2-Dichloroethane-d4	105	70-130
4-Bromofluorobenzene	106	70-130

Air Toxics Ltd.

RECOVERY REPORT

Client Name: Client SDG: 8-12apr  
 Sample Matrix: GAS Fraction: VOA  
 Lab Smp Id: LCS-1 Client Smp ID: LCS-1  
 Level: LOW Operator: JG  
 Data Type: MS DATA SampleType: LCS  
 SpikeList File: Spectra.spk Quant Type: ISTD  
 Sublist File: AT04+Na.sub  
 Method File: /chem/msd8.i/8-12apr.b/t14q322b.m  
 Misc Info: 100ppbv-50ppbv

SPIKE COMPOUND	CONC ADDED PPBV	CONC RECOVERED PPBV	% RECOVERED	LIMITS
131 Styrene	50.000	43.939	87.88	70-130
106 trans-1,3-Dichloro	50.000	46.966	93.93	70-130
4 Dichlorodifluorome	50.000	44.869	89.74	70-130
6 Freon 114	50.000	47.709	95.42	70-130
8 Chloromethane	50.000	51.124	102.25	70-130
10 Vinyl Chloride	50.000	50.463	100.93	70-130
11 1,3-Butadiene	50.000	46.259	92.52	60-140
13 Bromomethane	50.000	49.393	98.79	70-130
16 Chloroethane	50.000	49.287	98.57	70-130
18 Trichlorofluoromet	50.000	49.343	98.69	70-130
21 Ethanol	50.000	51.022	102.04	60-140
27 Freon 113	50.000	55.282	110.56	70-130
29 1,1-Dichloroethene	50.000	54.699	109.40	70-130
30 Acetone	50.000	53.204	106.41	60-140
33 Carbon Disulfide	50.000	49.341	98.68	60-140
34 2-Propanol	50.000	50.202	100.40	60-140
39 Methylene Chloride	50.000	54.051	108.10	70-130
42 MTBE	50.000	47.912	95.82	60-140
43 trans-1,2-Dichloro	50.000	47.782	95.56	60-140
45 Hexane	50.000	47.838	95.68	60-140
52 1,1-Dichloroethane	50.000	51.240	102.48	70-130
63 cis-1,2-Dichloroet	50.000	49.929	99.86	70-130
64 2-Butanone	50.000	43.737	87.47	60-140
66 Tetrahydrofuran	50.000	41.389	82.78	60-140
69 Chloroform	50.000	42.452	84.90	70-130
72 Cyclohexane	50.000	43.968	87.94	60-140
73 1,1,1-Trichloroeth	50.000	47.425	94.85	70-130
75 Carbon Tetrachlori	50.000	50.643	101.29	70-130
79 Benzene	50.000	42.832	85.66	70-130
81 1,2-Dichloroethane	50.000	50.355	100.71	70-130
82 Heptane	50.000	41.265	82.53	60-140
92 Trichloroethene	50.000	47.159	94.32	70-130
95 1,2-Dichloropropan	50.000	43.310	86.62	70-130

SPIKE COMPOUND	CONC ADDED PPBV	CONC RECOVERED PPBV	% RECOVERED	LIMITS
96 1,4-Dioxane	50.000	42.176	84.35	60-140
98 Bromodichlorometha	50.000	46.392	92.78	60-140
100 cis-1,3-Dichloropr	50.000	45.395	90.79	70-130
101 4-Methyl-2-pentano	50.000	40.591	81.18	60-140
103 Toluene	50.000	47.160	94.32	70-130
108 1,1,2-Trichloroeth	50.000	46.460	92.92	70-130
109 Tetrachloroethene	50.000	51.549	103.10	70-130
112 2-Hexanone	50.000	43.338	86.68	60-140
114 Dibromochlorometha	50.000	50.842	101.68	60-140
115 1,2-Dibromoethane	50.000	46.380	92.76	70-130
124 Chlorobenzene	50.000	47.107	94.21	70-130
127 Ethyl Benzene	50.000	46.952	93.90	70-130
128 m,p-Xylene	50.000	44.887	89.77	70-130
130 o-Xylene	50.000	46.078	92.16	70-130
133 Bromoform	50.000	53.446	106.89	60-140
142 1,1,2,2-Tetrachlor	50.000	43.788	87.58	70-130
145 4-Ethyltoluene	50.000	50.715	101.43	60-140
146 1,3,5-Trimethylben	50.000	46.758	93.52	70-130
151 1,2,4-Trimethylben	50.000	45.718	91.44	70-130
154 1,3-Dichlorobenzen	50.000	54.925	109.85	70-130
155 1,4-Dichlorobenzen	50.000	52.101	104.20	70-130
156 alpha-Chlorotoluen	50.000	54.441	108.88	70-130
158 1,2-Dichlorobenzen	50.000	52.112	104.22	70-130
163 1,2,4-Trichloroben	50.000	57.884	115.77	70-130
164 Hexachlorobutadien	50.000	49.913	99.83	70-130
135 Cumene	50.000	45.667	91.33	60-140
143 Propylbenzene	50.000	49.495	98.99	60-140
37 3-Chloropropene	50.000	51.579	103.16	60-140
78 2,2,4-Trimethylpen	50.000	45.148	90.30	60-140
165 Naphthalene	50.000	65.407	130.81	60-140

SURROGATE COMPOUND	CONC ADDED PPBV	CONC RECOVERED PPBV	% RECOVERED	LIMITS
\$ 80 1,2-Dichloroethane	25.000	26.334	105.34	70-130
\$ 102 Toluene-d8	25.000	24.000	96.00	70-130
\$ 138 Bromofluorobenzene	25.000	26.419	105.67	70-130

Report Date: 12-Apr-2007 13:05

## Air Toxics Ltd.

## AMBIENT AIR METHOD TO14A/TO15

Data file : /chem/msd8.i/8-12apr.b/8041203.d  
 Lab Smp Id: LCS-1 Client Smp ID: LCS-1  
 Inj Date : 12-APR-2007 10:21  
 Operator : JG Inst ID: msd8.i  
 Smp Info : 100mL #1408-408A  
 Misc Info : 100ppbv-50ppbv  
 Comment :  
 Method : /chem/msd8.i/8-12apr.b/t14q322b.m  
 Meth Date : 12-Apr-2007 13:04 jgray Quant Type: ISTD  
 Cal Date : 26-MAR-2007 13:11 Cal File: 8032608.d  
 Als bottle: 1 QC Sample: LCS  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: AT04+Na.sub  
 Target Version: 3.50 Sample Matrix: AIR  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* CpndVariable

Cpnd Variable Local Compound Variable

CONCENTRATIONS								
RT	EXP RT	(REL RT)	MASS	RESPONSE		TARGET RANGE	RATIO	
				( PPBV)	( PPBV)			
==	=====	=====	====	=====	=====	=====	=====	=====
-----								
* 67	Bromochloromethane					CAS #: 74-97-5		
8.395	8.395	(1.000)	130	189537	25.0000	80.00- 120.00	100.00	
8.395	8.395	(1.000)	128	155056		45.68- 105.68	81.81	
8.367	8.395	(1.000)	49	572041		268.48- 328.48	301.81	
-----								
* 86	1,4-Difluorobenzene					CAS #: 540-36-3		
10.248	10.275	(1.000)	114	907795	25.0000	80.00- 120.00	100.00	
10.248	10.275	(1.000)	88	177690		0.00- 48.01	19.57	
-----								
* 123	Chlorobenzene-d5					CAS #: 3114-55-4		
15.225	15.225	(1.000)	117	694716	25.0000	80.00- 120.00	100.00	
15.197	15.225	(1.000)	82	451855		35.49- 95.49	65.04	
-----								
\$ 80	1,2-Dichloroethane-d4					CAS #: 17060-07-0		
9.473	9.474	(1.128)	65	431451	26.3345	26.334 80.00- 120.00	100.00	
9.473	9.474	(1.128)	67	235905		27.92- 87.92	54.68	
-----								
\$ 102	Toluene-d8					CAS #: 2037-26-5		
12.985	12.985	(1.267)	98	856817	24.0005	24.000 80.00- 120.00	100.00	
12.985	12.985	(1.267)	70	115276		0.00- 42.61	13.45	

CONCENTRATIONS

ON-COL FINAL

RT	EXP RT	(REL RT)	MASS	RESPONSE	( PPEV)	( PPBV)	TARGET RANGE	RATIO
==	=====	=====	=====	=====	=====	=====	=====	=====

\$ 102 Toluene-d8 (continued)

12.985	12.985	(1.267)	100	609950			40.27- 100.27	71.19
--------	--------	---------	-----	--------	--	--	---------------	-------

\$ 138 Bromofluorobenzene

CAS #: 460-00-4

16.745	16.773	(1.100)	174	376126	26.4187	26.419	80.00- 120.00	100.00
16.745	16.773	(1.100)	95	630189			138.09- 198.09	167.55
16.745	16.773	(1.100)	176	387754			69.53- 129.53	103.09

3 Propylene

CAS #: 115-07-1

2.395	2.451	(0.285)	41	1048926	52.3039	52.304	80.00- 120.00	100.00
2.395	2.451	(0.285)	42	688682			35.20- 95.20	65.66
2.395	2.451	(0.285)	39	751613			42.80- 102.80	71.66

4 Dichlorodifluoromethane/Fr12

CAS #: 75-71-8

2.450	2.506	(0.292)	85	1902021	44.8694	44.869	80.00- 120.00	100.00
2.450	2.506	(0.292)	87	603028			1.89- 61.89	31.70

6 Freon 114

CAS #: 76-14-2

2.589	2.644	(0.308)	135	1147019	47.7089	47.709	80.00- 120.00	100.00
2.589	2.644	(0.308)	137	361920			0.73- 60.73	31.55

8 Chloromethane

CAS #: 74-87-3

2.727	2.782	(0.325)	50	1114969	51.1237	51.124	80.00- 120.00	100.00
2.727	2.782	(0.325)	52	326982			0.00- 59.46	29.33

10 Vinyl Chloride

CAS #: 75-01-4

2.893	2.948	(0.345)	62	1033759	50.4628	50.463	80.00- 120.00	100.00
2.893	2.948	(0.345)	64	302442			0.33- 60.33	29.26

11 1,3-Butadiene

CAS #: 106-99-0

2.893	2.921	(0.345)	54	982696	46.2590	46.259	80.00- 120.00	100.00
2.893	2.921	(0.345)	39	1117903			86.60- 146.60	113.76

13 Bromomethane

CAS #: 74-83-9

3.418	3.474	(0.407)	94	553123	49.3929	49.393	80.00- 120.00	100.00
3.418	3.474	(0.407)	96	501704			60.27- 120.27	90.70

16 Chloroethane

CAS #: 75-00-3

3.584	3.640	(0.427)	64	525778	49.2867	49.287	80.00- 120.00	100.00
3.584	3.640	(0.427)	49	172748			2.42- 62.42	32.86
3.584	3.640	(0.427)	66	162195			0.00- 58.92	30.85

18 Trichlorofluoromethane/Fr11

CAS #: 75-69-4

3.888	3.916	(0.463)	101	1873534	49.3429	49.343	80.00- 120.00	100.00
3.888	3.916	(0.463)	103	1223021			32.90- 92.90	65.28



CONCENTRATIONS

ON-COL FINAL

RT EXP RT (REL RT) MASS RESPONSE ( PPEV) ( PPBV) TARGET RANGE RATIO  
 == == ===== == ===== ===== ===== =====

21 Ethanol CAS #: 64-17-5  
 4.303 4.331 (0.513) 45 439772 51.0217 51.022 80.00- 120.00 100.00  
 4.303 4.331 (0.513) 43 89690 0.00- 50.43 20.39  
 4.275 4.331 (0.509) 46 181736 12.21- 72.21 41.33

27 Freon 113 CAS #: 76-13-1  
 4.745 4.773 (0.565) 151 993983 55.2824 55.282 80.00- 120.00 100.00  
 4.745 4.773 (0.565) 153 638835 34.26- 94.26 64.27  
 4.745 4.773 (0.565) 101 1520996 125.60- 185.60 153.02

29 1,1-Dichloroethene CAS #: 75-35-4  
 4.773 4.801 (0.569) 61 1674057 54.6987 54.699 80.00- 120.00 100.00  
 4.773 4.801 (0.569) 96 747283 15.96- 75.96 44.64  
 4.773 4.801 (0.569) 98 461816 0.00- 58.14 27.59

30 Acetone CAS #: 67-64-1  
 4.939 4.967 (0.588) 58 582362 53.2041 53.204 80.00- 120.00 100.00  
 4.939 4.967 (0.588) 43 2107650 337.35- 397.35 361.91

34 2-Propanol CAS #: 67-63-0  
 5.160 5.188 (0.615) 45 2152892 50.2024 50.202 80.00- 120.00 100.00  
 5.160 5.188 (0.615) 43 435288 0.00- 50.64 20.22  
 5.160 5.188 (0.615) 59 73647 0.00- 33.21 3.42

33 Carbon Disulfide CAS #: 75-15-0  
 5.132 5.160 (0.611) 76 2326055 49.3414 49.341 80.00- 120.00 100.00

37 3-Chloropropene CAS #: 107-05-1  
 5.437 5.464 (0.648) 76 392663 51.5790 51.579 80.00- 120.00 100.00  
 5.437 5.464 (0.648) 41 1758156 414.53- 474.53 447.75

39 Methylene Chloride CAS #: 75-09-2  
 5.713 5.741 (0.681) 49 1406775 54.0507 54.051 80.00- 120.00 100.00  
 5.713 5.741 (0.681) 84 658646 17.53- 77.53 46.82  
 5.713 5.741 (0.681) 51 401544 0.00- 59.93 28.54

42 MTBE CAS #: 1634-04-4  
 6.045 6.073 (0.720) 73 2513635 47.9118 47.912 80.00- 120.00 100.00  
 6.045 6.073 (0.720) 57 729075 0.00- 59.44 29.00  
 6.045 6.073 (0.720) 41 792049 1.95- 61.95 31.51

43 trans-1,2-Dichloroethene CAS #: 156-60-5  
 6.100 6.128 (0.727) 96 747326 47.7819 47.782 80.00- 120.00 100.00  
 6.100 6.128 (0.727) 61 1440981 165.01- 225.01 192.82  
 6.100 6.128 (0.727) 98 471333 34.71- 94.71 63.07

CONCENTRATIONS

ON-COL FINAL

RT EXP RT (REL RT) MASS RESPONSE ( PPEV) ( PPBV) TARGET RANGE RATIO  
 == == ===== == ===== ===== =====

45 Hexane CAS #: 110-54-3  
 6.432 6.460 (0.766) 57 1841736 47.8380 47.838 80.00- 120.00 100.00  
 6.432 6.460 (0.766) 43 1353085 44.61- 104.61 73.47  
 6.432 6.460 (0.766) 86 254183 0.00- 43.77 13.80

54 Vinyl Acetate CAS #: 108-05-4  
 6.957 6.985 (0.829) 86 206581 53.4087 53.409 80.00- 120.00 100.00  
 6.957 6.985 (0.829) 43 3172834 1469.62-1529.62 1535.88  
 6.957 6.985 (0.829) 42 273742 100.08- 160.08 132.51

52 1,1-Dichloroethane CAS #: 75-34-3  
 6.902 6.930 (0.822) 63 1666191 51.2403 51.240 80.00- 120.00 100.00  
 6.902 6.930 (0.822) 65 503972 1.16- 61.16 30.25

64 2-Butanone CAS #: 78-93-3  
 7.980 8.008 (0.951) 72 388200 43.7369 43.737 80.00- 120.00 100.00  
 7.980 8.008 (0.951) 43 2366073 572.72- 632.72 609.50  
 7.980 8.008 (0.951) 57 157316 10.61- 70.61 40.52

63 cis-1,2-Dichloroethene CAS #: 156-59-2  
 7.953 7.953 (0.947) 61 1235255 49.9295 49.929 80.00- 120.00 100.00  
 7.953 7.953 (0.947) 96 661986 24.50- 84.50 53.59  
 7.953 7.953 (0.947) 98 434420 3.86- 63.86 35.17

66 Tetrahydrofuran CAS #: 109-99-9  
 8.367 8.368 (0.997) 42 1367813 41.3893 41.389 80.00- 120.00 100.00  
 8.367 8.368 (0.997) 71 336418 0.00- 55.51 24.60  
 8.367 8.368 (0.997) 72 358476 0.00- 56.52 26.21

69 Chloroform CAS #: 67-66-3  
 8.506 8.533 (1.013) 83 1323123 42.4517 42.452 80.00- 120.00 100.00  
 8.506 8.533 (1.013) 85 852584 33.20- 93.20 64.44

73 1,1,1-Trichloroethane CAS #: 71-55-6  
 8.755 8.782 (1.043) 97 1334533 47.4251 47.425 80.00- 120.00 100.00  
 8.755 8.782 (1.043) 99 880162 35.42- 95.42 65.95

72 Cyclohexane CAS #: 110-82-7  
 8.727 8.755 (1.040) 84 1021505 43.9685 43.968 80.00- 120.00 100.00  
 8.727 8.755 (1.040) 56 1684831 137.82- 197.82 164.94  
 8.727 8.755 (1.040) 41 1048732 70.03- 130.03 102.67

75 Carbon Tetrachloride CAS #: 56-23-5  
 9.003 9.004 (1.072) 119 1151065 50.6430 50.643 80.00- 120.00 100.00  
 9.003 9.004 (1.072) 117 1283856 80.84- 140.84 111.54

CONCENTRATIONS									
RT	EXP RT	(REL RT)	MASS	RESPONSE		ON-COL	FINAL	TARGET RANGE	RATIO
				( PPEV)	( PPEV)				
==	=====	=====	=====	=====	=====	=====	=====	=====	=====
-----									
78	2,2,4-Trimethylpentane					CAS #: 540-84-1			
9.418	9.446	(1.122)	57	4894153	45.1476	45.148	80.00-	120.00	100.00
9.418	9.446	(1.122)	56	1718308			4.45-	64.45	35.11
9.418	9.446	(1.122)	41	1452430			0.00-	59.44	29.68
-----									
79	Benzene					CAS #: 71-43-2			
9.418	9.446	(0.919)	78	2165586	42.8321	42.832	80.00-	120.00	100.00
9.418	9.446	(0.919)	77	539807			0.00-	54.39	24.93
-----									
81	1,2-Dichloroethane					CAS #: 107-06-2			
9.612	9.612	(0.938)	62	1192281	50.3548	50.355	80.00-	120.00	100.00
9.612	9.612	(0.938)	64	376817			1.44-	61.44	31.60
-----									
82	Heptane					CAS #: 142-82-5			
9.805	9.833	(0.957)	100	249221	41.2655	41.265	80.00-	120.00	100.00
9.805	9.833	(0.957)	43	2051882			789.66-	849.66	823.32
9.805	9.833	(0.957)	71	803601			294.27-	354.27	322.45
-----									
92	Trichloroethene					CAS #: 79-01-6			
10.662	10.690	(1.040)	95	766935	47.1589	47.159	80.00-	120.00	100.00
10.662	10.690	(1.040)	130	673192			58.52-	118.52	87.78
10.662	10.690	(1.040)	97	473544			32.35-	92.35	61.74
-----									
95	1,2-Dichloropropane					CAS #: 78-87-5			
11.188	11.188	(1.092)	63	802440	43.3104	43.310	80.00-	120.00	100.00
11.188	11.188	(1.092)	62	586158			45.46-	105.46	73.05
11.188	11.188	(1.092)	41	654856			48.83-	108.83	81.61
-----									
96	1,4-Dioxane					CAS #: 123-91-1			
11.409	11.409	(1.113)	88	407374	42.1766	42.176	80.00-	120.00	100.00
11.409	11.409	(1.113)	58	390090			69.67-	129.67	95.76
11.409	11.409	(1.113)	57	142758			0.54-	60.54	35.04
-----									
98	Bromodichloromethane					CAS #: 75-27-4			
11.741	11.741	(1.146)	83	1270993	46.3924	46.392	80.00-	120.00	100.00
11.741	11.741	(1.146)	85	813985			33.80-	93.80	64.04
-----									
100	cis-1,3-Dichloropropene					CAS #: 10061-01-5			
12.625	12.626	(1.232)	75	1060684	45.3950	45.395	80.00-	120.00	100.00
12.625	12.626	(1.232)	77	324037			1.35-	61.35	30.55
12.598	12.626	(1.229)	39	862157			48.62-	108.62	81.28
-----									
101	4-Methyl-2-pentanone					CAS #: 108-10-1			
12.902	12.902	(1.259)	58	764709	40.5915	40.591	80.00-	120.00	100.00
12.902	12.902	(1.259)	43	2201230			251.14-	311.14	287.85
12.902	12.902	(1.259)	85	252376			2.33-	62.33	33.00
-----									

CONCENTRATIONS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	ON-COL ( PPEV)	FINAL ( PPBV)	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	
103 Toluene						CAS #:	108-88-3		
13.095	13.096	(1.278)	91	2159015	47.1601	47.160	80.00-	120.00	100.00
13.095	13.096	(1.278)	92	1311933			30.42-	90.42	60.77
-----									
106 trans-1,3-Dichloropropene						CAS #:	10061-02-6		
13.621	13.621	(0.895)	75	1137818	46.9663	46.966	80.00-	120.00	100.00
13.621	13.621	(0.895)	77	354909			2.16-	62.16	31.19
13.621	13.621	(0.895)	39	880130			44.46-	104.46	77.35
-----									
108 1,1,2-Trichloroethane						CAS #:	79-00-5		
13.870	13.897	(0.911)	97	655266	46.4597	46.460	80.00-	120.00	100.00
13.870	13.897	(0.911)	99	379466			32.43-	92.43	57.91
13.870	13.897	(0.911)	83	574913			58.75-	118.75	87.74
-----									
109 Tetrachloroethene						CAS #:	127-18-4		
13.925	13.925	(0.915)	166	772206	51.5488	51.549	80.00-	120.00	100.00
13.925	13.925	(0.915)	129	600360			48.86-	108.86	77.75
13.925	13.925	(0.915)	131	549221			44.30-	104.30	71.12
-----									
112 2-Hexanone						CAS #:	591-78-6		
14.257	14.257	(0.936)	58	964706	43.3386	43.338	80.00-	120.00	100.00
14.257	14.257	(0.936)	43	1927849			186.90-	246.90	199.84
14.257	14.257	(0.936)	100	155770			0.00-	45.68	16.15
-----									
114 Dibromochloromethane						CAS #:	124-48-1		
14.423	14.423	(0.947)	129	973082	50.8421	50.842	80.00-	120.00	100.00
14.423	14.423	(0.947)	127	753065			46.96-	106.96	77.39
-----									
115 1,2-Dibromoethane						CAS #:	106-93-4		
14.589	14.589	(0.958)	107	917861	46.3796	46.380	80.00-	120.00	100.00
14.589	14.589	(0.958)	109	873115			62.52-	122.52	95.12
-----									
124 Chlorobenzene						CAS #:	108-90-7		
15.252	15.252	(1.002)	112	1537462	47.1068	47.107	80.00-	120.00	100.00
15.252	15.252	(1.002)	114	479806			1.50-	61.50	31.21
15.252	15.252	(1.002)	77	1152336			42.23-	102.23	74.95
-----									
127 Ethyl Benzene						CAS #:	100-41-4		
15.363	15.363	(1.009)	106	860331	46.9517	46.952	80.00-	120.00	100.00
15.363	15.363	(1.009)	91	2800986			309.90-	369.90	325.57
-----									
128 m,p-Xylene						CAS #:	108-38-3		
15.529	15.529	(1.020)	106	1057652	44.8868	44.887	80.00-	120.00	100.00
15.529	15.529	(1.020)	91	2260788			177.38-	237.38	213.76
-----									
130 o-Xylene						CAS #:	95-47-6		
16.054	16.054	(1.054)	106	1046931	46.0782	46.078	80.00-	120.00	100.00

CONCENTRATIONS

RT	EXP RT	(REL RT)	MASS	RESPONSE	ON-COL (PPEV)	FINAL (PPBV)	TARGET RANGE	RATIO
==	=====	=====	=====	=====	=====	=====	=====	=====
130 o-Xylene (continued)								
16.054	16.054	(1.054)	91	2371912			192.94- 252.94	226.56
-----								
131 Styrene						CAS #:	100-42-5	
16.082	16.109	(1.056)	104	1536720	43.9389	43.939	80.00- 120.00	100.00
16.082	16.109	(1.056)	78	958143			30.05- 90.05	62.35
-----								
133 Bromoform						CAS #:	75-25-2	
16.358	16.358	(1.074)	173	845311	53.4461	53.446	80.00- 120.00	100.00
16.358	16.358	(1.074)	171	432748			20.81- 80.81	51.19
-----								
142 1,1,2,2-Tetrachloroethane						CAS #:	79-34-5	
16.966	16.967	(1.114)	83	1425054	43.7886	43.788	80.00- 120.00	100.00
16.966	16.967	(1.114)	85	903957			34.20- 94.20	63.43
-----								
145 4-Ethyltoluene						CAS #:	622-96-8	
17.132	17.132	(1.125)	105	3455497	50.7155	50.715	80.00- 120.00	100.00
17.132	17.132	(1.125)	120	844659			0.00- 54.89	24.44
-----								
146 1,3,5-Trimethylbenzene						CAS #:	108-67-8	
17.215	17.215	(1.131)	105	3015303	46.7586	46.758	80.00- 120.00	100.00
17.215	17.215	(1.131)	120	1246154			12.64- 72.64	41.33
-----								
151 1,2,4-Trimethylbenzene						CAS #:	95-63-6	
17.602	17.602	(1.156)	105	2669566	45.7176	45.718	80.00- 120.00	100.00
17.602	17.602	(1.156)	120	1096850			9.75- 69.75	41.09
-----								
154 1,3-Dichlorobenzene						CAS #:	541-73-1	
17.906	17.907	(1.176)	146	1393046	54.9252	54.925	80.00- 120.00	100.00
17.906	17.907	(1.176)	148	855649			32.20- 92.20	61.42
17.879	17.907	(1.174)	111	590556			13.28- 73.28	42.39
-----								
155 1,4-Dichlorobenzene						CAS #:	106-46-7	
17.989	17.990	(1.182)	146	1685072	52.1013	52.101	80.00- 120.00	100.00
17.989	17.990	(1.182)	148	1058710			34.16- 94.16	62.83
17.989	17.990	(1.182)	111	771690			16.41- 76.41	45.80
-----								
156 alpha-Chlorotoluene						CAS #:	100-44-7	
18.128	18.128	(1.191)	91	2549241	54.4411	54.441	80.00- 120.00	100.00
18.128	18.128	(1.191)	126	433771			0.00- 46.97	17.02
-----								
158 1,2-Dichlorobenzene						CAS #:	95-50-1	
18.321	18.321	(1.203)	146	1450552	52.1116	52.112	80.00- 120.00	100.00
18.321	18.321	(1.203)	148	897648			32.47- 92.47	61.88
18.321	18.321	(1.203)	111	740575			20.91- 80.91	51.05
-----								

CONCENTRATIONS									
RT	EXP RT	(REL RT)	MASS	RESPONSE	( PPEV)	FINAL	TARGET RANGE	RATIO	
==	=====	=====	=====	=====	=====	=====	=====	=====	=====
-----									
163	1,2,4-Trichlorobenzene			CAS #: 120-82-1					
19.593	19.593	(1.287)	180	1003346	57.8841	57.884	80.00-	120.00	100.00
19.593	19.593	(1.287)	182	962355			65.68-	125.68	95.91
-----									
164	Hexachlorobutadiene			CAS #: 87-68-3					
19.676	19.676	(1.292)	225	1105134	49.9129	49.913	80.00-	120.00	100.00
19.676	19.676	(1.292)	223	708369			33.53-	93.53	64.10
-----									
143	Propylbenzene			CAS #: 103-65-1					
16.994	16.994	(1.116)	91	3880478	49.4953	49.495	80.00-	120.00	100.00
16.994	16.994	(1.116)	120	710168			0.00-	48.79	18.30
16.994	16.994	(1.116)	105	148600			0.00-	33.35	3.83
-----									
135	Cumene			CAS #: 98-82-8					
16.524	16.524	(1.085)	105	3166375	45.6672	45.667	80.00-	120.00	100.00
16.524	16.524	(1.085)	120	718396			0.00-	53.14	22.69
16.496	16.524	(1.084)	51	490947			0.00-	45.39	15.51
-----									
165	Naphthalene			CAS #: 91-20-3					
19.787	19.787	(1.300)	128	3622625	65.4072	65.407	80.00-	120.00	100.00
19.787	19.787	(1.300)	127	514084			0.00-	43.72	14.19
-----									

Report Date: 12-Apr-2007 13:05

Air Toxics Ltd.

INTERNAL STANDARD COMPOUNDS  
AREA AND RT SUMMARY

Instrument ID: msd8.i

Calibration Date: 12-APR-2007

Lab File ID: 8041203.d

Calibration Time: 09:53

Lab Smp Id: LCS-1

Client Smp ID: LCS-1

Analysis Type: VOA

Level: LOW

Quant Type: ISTD

Sample Type: AIR

Operator: JG

Method File: /chem/msd8.i/8-12apr.b/t14q322b.m

Misc Info: 100ppbv-50ppbv

COMPOUND	STANDARD	AREA LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	269481	161689	377273	189537	-29.67
86 1,4-Difluorobenze	1204367	722620	1686114	907795	-24.62
123 Chlorobenzene-d5	906226	543736	1268716	694716	-23.34


COMPOUND	STANDARD	RT LIMIT		SAMPLE	%DIFF
		LOWER	UPPER		
67 Bromochloromethan	8.40	8.07	8.73	8.40	0.00
86 1,4-Difluorobenze	10.28	9.95	10.61	10.25	-0.27
123 Chlorobenzene-d5	15.22	14.89	15.55	15.22	0.00

AREA UPPER LIMIT = + 40% of internal standard area.

AREA LOWER LIMIT = - 40% of internal standard area.

RT UPPER LIMIT = + 0.33 minutes of internal standard RT.

RT LOWER LIMIT = - 0.33 minutes of internal standard RT.





m/z	ION ABUNDANCE CRITERIA	% REL. ABUNDANCE
50	15.0 - 40.0% of mass 95	26.09
75	30.0 - 60.0% of mass 95	52.45
95	Base peak, 100.00% relative abundance	100.00
96	5.0 - 9.0% of mass 95	6.49
173	Less than 2.0% of mass 174	(0.00) <sup>1</sup>
174	Greater than 50.0% of mass 95	80.40
175	5.0 - 9.0% of mass 174	(8.17) <sup>1</sup>
176	Greater than 95.0% but less than 101.0% of mass 174	(26.41) <sup>1</sup>
177	5.0 - 9.0% of mass 176	(6.22) <sup>2</sup>

- value in parenthesis is % mass 174  
 - value in parenthesis is % mass 176

Verify 176/174 m/z Ratio:  $\frac{1334784}{1334784} = 1.00$

1334784 / 1334784 = 0.0011

NOAH Cart #: 5/8  
 File #: F041102 / F041108

File ID:	F041102
Compound:	T81-d8
Initials:	DR

$$\frac{\text{Area}_{\text{sample}}}{\text{Areas}} \times \frac{\text{Conc.}_{\text{is}}}{\text{RRF}} = \frac{(1167702)}{(1204367)} \times \frac{(25)}{(0.98315)} = 24.454$$

Reported Result: 24.654

Calculation Check:

ppbv of compound =

Use	File #	Sample / Client Name	Can #	Pressure	Amt Loaded	DF	Loader Init.	Date Analyzed	Time Analyzed	Review Init.	Comments
✓	6041201	BFB Tune Check	2423	SDwg	2µL	1.00	Q-	4/12/07	0922	Q-	good +1
✓	02	CEV # 1408-387A	100µL	50µPa	100µL	1.00	Q-		0953	Q-	
✓	03	LCC # 1408-408A							1621	Q-	
X	04	System Blank	13473	Turned	200µL	1.00	Q-		1344	Q-	
✓	05	Lab Blank							1412	Q-	
✓	06	0704143 - OVA	24231	50µPa	250µL	1.00	Q-		1440	Q-	
✓	07	07	1207	50µPa	1µL	1.00	Q-		1723	Q-	
✓	08	0704145	1488	50µPa	95µL	6.38	Q-		1801	Q-	
✓	09	02AA	1488	50µPa	1µL	1.00	Q-		1839	Q-	

Signature: [Signature]  
 Date: 4/12/07

Report Date: 22-Mar-2007 09:23

## Air Toxics Ltd.

Data file : /var/chem/msd8.i/8-22mar.b/8032201.d  
 Lab Smp Id: BFB Client Smp ID: BFB  
 Inj Date : 22-MAR-2007 09:29  
 Operator : srs Inst ID: msd8.i  
 Smp Info : BFB Tune Check  
 Misc Info : 50ng 2uL #843-2786  
 Comment :  
 Method : /var/chem/msd8.i/8-22mar.b/bfb30.m  
 Meth Date : 22-Mar-2007 09:23 Quant Type: ESTD  
 Cal Date : Cal File:  
 Als bottle: 1 QC Sample: BFB  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: all.sub  
 Target Version: 3.50 Sample Matrix: WATER  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* Uf \* Vf \* Vi \* CpndVariable

Name	Value	Description
DF	1.00000	Dilution Factor
Uf	1.00000	ng unit correction factor
Vf	1.00000	Volumetric correction factor
Vi	2.00000	Injection Volume

Cpnd Variable Local Compound Variable

## CONCENTRATIONS

ON-COL FINAL

RT	EXP RT	DLT RT	MASS	RESPONSE	( ug/L)	( ug/L)	TARGET RANGE	RATIO
==	=====	=====	=====	=====	=====	=====	=====	=====

CAS #: 460-00-4

1 bfb								
3.803	3.748	0.055	95	1175749			100.00- 100.00	100.00
3.803	3.748	0.055	50	367505			15.00- 40.00	31.26
3.803	3.748	0.055	75	659620			30.00- 60.00	56.10
3.803	3.748	0.055	96	74114			5.00- 9.00	6.30
3.803	3.748	0.055	173	176			0.00- 2.00	0.03
3.803	3.748	0.055	174	701792			50.00- 100.00	59.69
3.803	3.748	0.055	175	44998			5.00- 9.00	6.41
3.803	3.748	0.055	176	675517			95.00- 101.00	96.26
3.803	3.748	0.055	177	43845			5.00- 9.00	6.49

Date : 22-MAR-2007 09:29

Client ID: BFB

Instrument: msd8.i


Sample Info: BFB Tune Check

Volume Injected (uL): 2.0

Operator: srs

Column phase:

Column diameter: 0.53



Date : 22-MAR-2007 09:29

Client ID: BFB

Instrument: msd8,i

Sample Info: BFB Tune Check


Volume Injected (uL): 2.0

Operator: srs

Column phase:

Column diameter: 0.53

1 bfb



m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
95	Base Peak, 100% relative abundance	100.00
50	15.00 - 40.00% of mass 95	31.26
75	30.00 - 60.00% of mass 95	56.10
96	5.00 - 9.00% of mass 95	6.30
173	Less than 2.00% of mass 174	0.01 ( 0.03)
174	50.00 - 100.00% of mass 95	59.69
175	5.00 - 9.00% of mass 174	3.83 ( 6.41)
176	95.00 - 101.00% of mass 174	57.45 ( 96.26)
177	5.00 - 9.00% of mass 176	3.73 ( 6.49)

Date : 22-MAR-2007 09:29

Client ID: BFB

Instrument: msd8.i

Sample Info: BFB Tune Check

Volume Injected (uL): 2.0

Operator: srs

Column phase:

Column diameter: 0.53

Data File: 8032201.d

Spectrum: Avg. Scans 26-28 ( 3.80), Background Scan 24

Location of Maximum: 95.00

Number of points: 216

m/z	Y	m/z	Y	m/z	Y	m/z	Y
36.00	17368	95.00	1175552	156.00	547	230.00	147
37.00	95296	96.00	74112	157.00	1221	231.00	120
38.00	79344	97.00	1966	159.00	1076	238.00	137
39.00	30888	102.00	374	160.00	412	242.00	192
40.00	1004	103.00	713	161.00	978	243.00	158
42.00	747	104.00	3265	162.00	76	245.00	103
44.00	8653	105.00	1211	163.00	156	247.00	145
45.00	14978	106.00	3258	164.00	185	248.00	153
46.00	1377	107.00	1067	165.00	212	249.00	152
47.00	24584	108.00	211	166.00	512	250.00	70
48.00	10631	109.00	364	167.00	233	253.00	677
49.00	81272	110.00	373	168.00	477	255.00	60
50.00	367488	111.00	793	169.00	450	257.00	111
51.00	109000	112.00	548	170.00	961	258.00	314
52.00	4533	113.00	829	171.00	617	259.00	138
53.00	295	115.00	620	172.00	200	260.00	411
54.00	656	116.00	2850	173.00	176	261.00	201
55.00	4277	117.00	4491	174.00	701760	262.00	199
56.00	23336	118.00	3301	175.00	44992	263.00	144
57.00	39272	119.00	3137	176.00	675456	267.00	241
58.00	1897	120.00	285	177.00	43840	273.00	69
59.00	523	121.00	749	178.00	1250	284.00	162
60.00	14828	122.00	199	179.00	221	289.00	71
61.00	68712	123.00	280	180.00	99	290.00	167
62.00	64208	124.00	489	181.00	96	293.00	90
63.00	46984	125.00	96	182.00	48	296.00	181
64.00	4051	126.00	136	184.00	158	297.00	152
65.00	326	127.00	288	185.00	75	298.00	89
66.00	502	128.00	2354	187.00	233	300.00	159
67.00	1549	129.00	1080	191.00	383	303.00	180
68.00	136640	130.00	2707	192.00	28	304.00	82
69.00	131392	131.00	1090	194.00	245	306.00	188
70.00	8582	132.00	77	195.00	83	307.00	71
71.00	1100	133.00	759	197.00	112	309.00	220
72.00	7269	134.00	381	201.00	76	310.00	73

Date : 22-MAR-2007 09:29

Client ID: BFB

Instrument: msd8.i

Sample Info: BFB Tune Check

Volume Injected (uL): 2.0

Operator: srs

Column phase:

Column diameter: 0.53

Data File: 8032201.d

Spectrum: Avg. Scans 26-28 ( 3.80), Background Scan 24

Location of Maximum: 95.00

Number of points: 216

m/z	Y	m/z	Y	m/z	Y	m/z	Y
73.00	57272	135.00	1208	202.00	314	311.00	185
74.00	219264	136.00	521	203.00	77	312.00	76
75.00	659584	137.00	1129	205.00	105	313.00	40
76.00	55072	138.00	58	207.00	644	314.00	151
77.00	6210	139.00	602	208.00	178	316.00	68
78.00	4088	141.00	8415	210.00	259	317.00	69
79.00	25056	142.00	737	211.00	74	319.00	72
80.00	8805	143.00	9048	212.00	88	322.00	40
81.00	25072	144.00	821	214.00	182	326.00	78
82.00	5117	145.00	1133	215.00	73	331.00	79
83.00	380	146.00	882	216.00	224	336.00	74
84.00	413	147.00	301	217.00	176	337.00	101
86.00	1542	148.00	1337	219.00	301	341.00	103
87.00	50160	149.00	391	220.00	204	343.00	481
88.00	47656	150.00	1148	221.00	118	344.00	292
89.00	267	151.00	708	223.00	202	346.00	47
91.00	3315	152.00	499	225.00	170		
92.00	30656	153.00	823	226.00	164		
93.00	48528	154.00	585	228.00	128		
94.00	142784	155.00	1853	229.00	71		

Report Date: 26-Mar-2007 08:57

Air Toxics Ltd.

Data file : /var/chem/msd8.i/8-26mar.b/8032601.d  
 Lab Smp Id: BFB Client Smp ID: BFB  
 Inj Date : 26-MAR-2007 09:03  
 Operator : ea Inst ID: msd8.i  
 Smp Info : BFB Tune Check  
 Misc Info : 50ng 2uL #843-2912  
 Comment :  
 Method : /var/chem/msd8.i/8-26mar.b/bfb30.m  
 Meth Date : 26-Mar-2007 08:57 Quant Type: ESTD  
 Cal Date : Cal File:  
 Als bottle: 1 QC Sample: BFB  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: all.sub  
 Target Version: 3.50 Sample Matrix: WATER  
 Processing Host: eeyore

Concentration Formula: Amt \* DF \* Uf \* Vf \* Vi \* CpndVariable

Name	Value	Description
DF	1.00000	Dilution Factor
Uf	1.00000	ng unit correction factor
Vf	1.00000	Volumetric correction factor
Vi	2.00000	Injection Volume

Cpnd Variable Local Compound Variable

CONCENTRATIONS

ON-COL FINAL

RT EXP RT DLT RT MASS RESPONSE ( ug/L) ( ug/L) TARGET RANGE RATIO  
 == =====

RT	EXP RT	DLT RT	MASS	RESPONSE	( ug/L)	( ug/L)	TARGET RANGE	RATIO
1	bfb					CAS #: 460-00-4		
3.803	3.748	0.055	95	2384521			100.00- 100.00	100.00
3.803	3.748	0.055	50	692921			15.00- 40.00	29.06
3.803	3.748	0.055	75	1291079			30.00- 60.00	54.14
3.803	3.748	0.055	96	153259			5.00- 9.00	6.43
3.803	3.748	0.055	173	0			0.00- 2.00	0.00
3.803	3.748	0.055	174	1255899			50.00- 100.00	52.67
3.803	3.748	0.055	175	99889			5.00- 9.00	7.95
3.803	3.748	0.055	176	1198128			95.00- 101.00	95.40
3.803	3.748	0.055	177	78262			5.00- 9.00	6.53

Date : 26-MAR-2007 09:03

Client ID: BFB

Instrument: msd8.i


Sample Info: BFB Tune Check

Volume Injected (uL): 2.0

Operator: ea

Column phase:

Column diameter: 0.53





Date : 26-MAR-2007 09:03

Client ID: BFB

Instrument: msd8,i

Sample Info: BFB Tune Check


Volume Injected (uL): 2.0

Operator: ea

Column phase:

Column diameter: 0.53

1 bfb



m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
95	Base Peak, 100% relative abundance	100.00
50	15.00 - 40.00% of mass 95	29.06
75	30.00 - 60.00% of mass 95	54.14
96	5.00 - 9.00% of mass 95	6.43
173	Less than 2.00% of mass 174	0.00 ( 0.00)
174	50.00 - 100.00% of mass 95	52.67
175	5.00 - 9.00% of mass 174	4.19 ( 7.95)
176	95.00 - 101.00% of mass 174	50.25 ( 95.40)
177	5.00 - 9.00% of mass 176	3.28 ( 6.53)

Date : 26-MAR-2007 09:03

Client ID: BFB

Instrument: msd8.i

Sample Info: BFB Tune Check

Volume Injected (uL): 2.0

Operator: ea

Column phase:

Column diameter: 0.53

Data File: 8032601.d

Spectrum: Avg. Scans 26-28 ( 3.80), Background Scan 24

Location of Maximum: 95.00

Number of points: 203

m/z	Y	m/z	Y	m/z	Y	m/z	Y
35.00	195	91.00	6069	149.00	856	232.00	975
36.00	29232	92.00	62096	150.00	1457	236.00	199
37.00	163072	93.00	96864	151.00	920	244.00	235
38.00	139136	94.00	261312	152.00	276	246.00	291
39.00	54504	95.00	2384384	153.00	1211	250.00	294
40.00	476	96.00	153216	154.00	515	252.00	248
44.00	9853	97.00	3416	155.00	3267	254.00	26
45.00	26728	98.00	440	157.00	2464	256.00	5
47.00	47912	100.00	128	158.00	425	257.00	317
48.00	20328	101.00	987	159.00	1563	259.00	100
49.00	140992	104.00	6092	160.00	134	263.00	388
50.00	692864	105.00	1471	161.00	1495	267.00	53
51.00	200768	106.00	5907	162.00	248	268.00	17
52.00	7992	107.00	1526	163.00	126	269.00	224
54.00	568	108.00	19	166.00	463	270.00	186
55.00	6269	110.00	730	168.00	783	271.00	64
56.00	42136	111.00	1093	169.00	1180	273.00	57
57.00	74808	112.00	933	170.00	701	274.00	358
58.00	2453	113.00	853	171.00	1361	276.00	71
59.00	1045	114.00	317	174.00	1255424	277.00	179
60.00	25104	115.00	1261	175.00	99888	279.00	160
61.00	129520	116.00	5132	176.00	1198080	285.00	234
62.00	129464	117.00	8915	177.00	78256	286.00	70
63.00	90936	118.00	4419	178.00	2052	287.00	97
64.00	7300	119.00	7927	179.00	52	288.00	166
65.00	952	121.00	53	181.00	208	290.00	14
66.00	177	124.00	1048	182.00	25	292.00	113
67.00	7131	125.00	179	183.00	103	293.00	88
68.00	255232	126.00	635	184.00	24	297.00	55
69.00	261568	127.00	482	186.00	45	301.00	237
70.00	16728	128.00	5195	188.00	320	305.00	165
71.00	946	129.00	2340	189.00	192	307.00	191
72.00	12813	130.00	4681	191.00	669	309.00	173
73.00	105472	131.00	1755	192.00	550	311.00	511
74.00	409664	132.00	97	193.00	171	312.00	167

Date : 26-MAR-2007 09:03

Client ID: BFB

Instrument: msd8.i

Sample Info: BFB Tune Check

Volume Injected (uL): 2.0

Operator: ea

Column phase:

Column diameter: 0.53

Data File: 8032601.d  
Spectrum: Avg. Scans 26-28 ( 3.80), Background Scan 24  
Location of Maximum: 95.00  
Number of points: 203

m/z	Y	m/z	Y	m/z	Y	m/z	Y
75.00	1290752	133.00	1720	197.00	21	313.00	174
76.00	108680	134.00	1145	198.00	104	315.00	130
77.00	12010	135.00	1028	199.00	132	323.00	77
78.00	7804	136.00	442	201.00	86	324.00	295
79.00	44592	137.00	1916	203.00	239	327.00	37
80.00	13959	138.00	222	204.00	112	328.00	138
81.00	46472	139.00	458	206.00	279	329.00	193
82.00	8716	140.00	409	207.00	1386	332.00	130
83.00	1106	141.00	14444	211.00	262	334.00	233
84.00	86	142.00	2541	215.00	164	335.00	129
85.00	637	143.00	15084	218.00	320	340.00	104
86.00	843	144.00	623	223.00	362	341.00	351
87.00	96208	145.00	495	224.00	160	344.00	386
88.00	88216	146.00	1934	226.00	114	346.00	405
89.00	256	147.00	1293	228.00	159	349.00	73
90.00	728	148.00	3617	231.00	208		

Report Date: 12-Apr-2007 11:19

Air Toxics Ltd.

Data file : /chem/msd8.i/8-12apr.b/8041201.d  
 Lab Smp Id: BFB Client Smp ID: BFB  
 Inj Date : 12-APR-2007 09:22  
 Operator : JG Inst ID: msd8.i  
 Smp Info : BFB Tune Check  
 Misc Info : 50ng 2uL #843-2912  
 Comment :  
 Method : /var/chem/msd8.i/8-12apr.b/bfb30.m  
 Meth Date : 12-Apr-2007 09:16 Quant Type: ESTD  
 Cal Date : Cal File:  
 Als bottle: 1 QC Sample: BFB  
 Dil Factor: 1.00000  
 Integrator: HP RTE Compound Sublist: all.sub  
 Target Version: 3.50 Sample Matrix: WATER

Concentration Formula: Amt \* DF \* Uf \* Vf \* Vi \* CpndVariable

Name	Value	Description
DF	1.00000	Dilution Factor
Uf	1.00000	ng unit correction factor
Vf	1.00000	Volumetric correction factor
Vi	2.00000	Injection Volume

Cpnd Variable Local Compound Variable

CONCENTRATIONS

ON-COL FINAL

RT	EXP RT	DLT RT	MASS	RESPONSE ( ug/L)	( ug/L)	TARGET RANGE	RATIO
==	=====	=====	====	=====	=====	=====	=====

1 bfb

CAS #: 460-00-4

3.804	3.748	0.056	95	1721856		100.00- 100.00	100.00
3.804	3.748	0.056	50	449216		15.00- 40.00	26.09
3.804	3.748	0.056	75	903104		30.00- 60.00	52.45
3.804	3.748	0.056	96	111776		5.00- 9.00	6.49
3.804	3.748	0.056	173	0		0.00- 2.00	0.00
3.804	3.748	0.056	174	1384448		50.00- 100.00	80.40
3.804	3.748	0.056	175	113120		5.00- 9.00	8.17
3.804	3.748	0.056	176	1334784		95.00- 101.00	96.41
3.804	3.748	0.056	177	83504		5.00- 9.00	6.26

Date : 12-APR-2007 09:22

Client ID: BFB

Instrument: msd8.i


Sample Info: BFB Tune Check

Volume Injected (uL): 2.0

Operator: JG

Column phase:

Column diameter: 0.53



Date : 12-APR-2007 09:22

Client ID: BFB

Instrument: msd8.i

Sample Info: BFB Tune Check


Volume Injected (uL): 2.0

Operator: JG

Column phase:

Column diameter: 0.53

1 bfb



m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
95	Base Peak, 100% relative abundance	100.00
50	15.00 - 40.00% of mass 95	26.09
75	30.00 - 60.00% of mass 95	52.45
96	5.00 - 9.00% of mass 95	6.49
173	Less than 2.00% of mass 174	0.00 ( 0.00)
174	50.00 - 100.00% of mass 95	80.40
175	5.00 - 9.00% of mass 174	6.57 ( 8.17)
176	95.00 - 101.00% of mass 174	77.52 ( 96.41)
177	5.00 - 9.00% of mass 176	4.85 ( 6.26)

Date : 12-APR-2007 09:22

Client ID: BFB

Instrument: msd8.i

Sample Info: BFB Tune Check

Volume Injected (uL): 2.0

Operator: JG

Column phase:

Column diameter: 0.53

Data File: 8041201.d

Spectrum: HP MS 8041201.d, Scan 28: 3.831 min.

Location of Maximum: 95.00

Number of points: 191

m/z	Y	m/z	Y	m/z	Y	m/z	Y
35.90	19312	86.90	58136	138.90	884	197.40	248
36.90	102576	87.90	60752	139.90	1149	198.80	255
37.90	86352	90.00	863	140.90	14953	205.90	219
39.00	34152	90.90	6941	141.80	2350	206.90	3106
40.00	3683	91.90	51112	142.80	14091	207.80	724
40.90	1623	92.90	73168	143.90	1024	208.80	438
42.00	813	94.00	189888	144.80	2168	211.00	218
42.90	2137	95.00	1721856	145.90	2398	219.00	332
43.90	26712	95.90	111776	146.80	1088	231.90	659
44.90	19640	96.90	3636	147.80	3740	233.10	616
46.00	1382	99.20	313	148.80	1226	239.20	252
47.00	30224	100.40	298	149.90	1957	240.90	282
47.90	14128	100.90	281	151.10	514	241.30	228
48.90	90152	102.70	942	151.90	849	247.00	268
49.90	449216	103.80	5709	152.70	1100	248.50	441
50.90	136640	104.80	2334	153.90	1045	249.60	248
52.00	6284	105.80	5291	154.90	3807	252.80	1518
52.80	807	106.90	1470	155.80	565	254.10	594
54.90	5353	108.30	308	156.80	2388	255.00	406
55.90	27744	108.70	416	158.00	409	259.90	709
56.90	51792	109.60	511	158.60	1713	260.70	439
57.80	2040	110.10	396	159.90	478	265.40	452
58.90	816	110.90	1022	160.90	1757	267.10	335
59.90	16082	111.80	897	162.40	244	267.90	327
60.90	87200	112.70	714	163.00	479	268.90	4537
61.90	85608	113.90	208	164.70	281	270.00	829
62.90	66040	114.90	1860	165.10	432	271.20	595
64.00	5047	115.90	4635	165.60	298	271.90	245
64.90	1126	116.90	7499	166.10	381	280.80	771
66.80	4718	117.80	4982	167.40	383	283.00	341
67.90	172032	118.90	6607	168.10	488	283.60	212
68.90	173056	120.00	323	168.90	536	285.30	240
69.90	11014	120.90	296	170.00	1028	294.80	324
71.00	914	121.80	465	170.70	1510	309.30	234
72.00	8317	122.70	453	173.90	1384448	310.40	225

Date : 12-APR-2007 09:22

Client ID: BFB

Instrument: msd8.i

Sample Info: BFB Tune Check

Volume Injected (uL): 2.0

Operator: JG

Column phase:

Column diameter: 0.53

Data File: 8041201.d

Spectrum: HP MS 8041201.d, Scan 28: 3.831 min.

Location of Maximum: 95.00

Number of points: 191

m/z	Y	m/z	Y	m/z	Y	m/z	Y
72.90	74672	123.70	885	174.90	113120	311.10	251
73.90	288512	124.80	542	175.90	1334784	319.50	238
74.90	903104	125.90	391	176.90	83504	321.10	203
75.90	74584	126.90	896	177.80	1955	325.10	300
76.90	7236	127.80	5566	179.00	678	327.10	267
78.00	5946	128.90	3002	186.00	212	327.70	227
78.80	34128	129.80	5473	190.70	801	334.10	380
79.90	9471	130.90	2221	191.70	515	334.90	279
80.80	35808	132.80	1565	193.00	479	340.90	223
81.90	7334	134.80	2728	193.60	236	343.10	1042
82.90	1390	136.00	396	194.80	290	344.10	479
84.20	283	136.80	2117	195.10	324	344.90	547
84.80	450	137.90	227	196.20	214		



## **Shipping/ Receiving Documents**



AN ENVIRONMENTAL ANALYTICAL LABORATORY

**180 Blue Ravine Road, Suite B  
Folsom, CA 95630**

**Phone (916) 985-1000 FAX (916) 985-1020  
Hours 8:00 A.M. to 6:00 P.M. Pacific**

COMPANY: \_\_\_\_\_ GEI Consultants, Inc.  
ATTENTION: \_\_\_\_\_ Ms. Sarah Aldridge  
FAX #: \_\_\_\_\_ 860-368-5307  
FROM: \_\_\_\_\_ Sample Receiving  
Workorder #: \_\_\_\_\_ 0704143  
# of pages (Including Cover): \_\_\_\_\_ 1

4/24/2007

Thank you for selecting Air Toxics Ltd. We have received your samples and have found discrepancies. In order to expedite analysis and reporting, please review the attached information for accuracy. Corrections can be faxed to **Kelly Buettner at 916-985-1020**. ATL will proceed with the analysis as specified on the Chain of Custody and Sample Login page.

The following discrepancy has been observed:

Sample identifications for samples 000001016 Downwind and 12007 Upwind were not provided on the sample tags. The samples were identified by matching the canister number on each sample to the canister numbers on the Chain of Custody (COC). The Sample ID information on the COC will be used to process and report the samples.

*Your prompt response is appreciated.*

Receipt VLR  
4/7/07

# AIR TOXICS LTD.

AN ENVIRONMENTAL ANALYTICAL LABORATORY

## CHAIN-OF-CUSTODY RECORD

### Sample Transportation Notice

Requisitioning signature on this document indicates that sample is being shipped in compliance with all applicable local, State, Federal, national, and international laws, regulations and ordinances of any kind. Air Toxics Limited assumes no liability with respect to the collection, handling, or shipping of these samples. Requisitioning signature also indicates agreement to hold harmless, defend, and indemnify Air Toxics Limited against any claim, demand, or action of any kind, related to the collection, handling, or shipping of samples. D.O.T. Hotline (800) 497-4922

180 BLUE RAVINE ROAD, SUITE B  
FOLSOM, CA 95630-4719

(916) 985-1000 FAX: (916) 985-1020

Contact: *Karen Swartz*  
 Company: GEI Consultants, Inc.  
 Address: 455 Whiting Brook Glastonbury CT 06033  
 Phone: 860-368-8300 Cell: \_\_\_\_\_  
 Collected By: Signature: \_\_\_\_\_

Project Info: *Baysshore S. Cell OU-1*  
 P.O. #: \_\_\_\_\_  
 Project #: 061140-8-1703  
 Project Name: Baysshore OU1 Southern cell  
 Air Monitoring

Turn Around Time:  
 Normal  
 Rush \_\_\_\_\_  
 Specify \_\_\_\_\_

Lab I.D.	Field Sample I.D.	Date & Time	Analyses Requested	Canister Pressure/Vacuum Inlet Final Receipt
01A	000001016 <i>Dawson</i>	4/5/07 6:50-15:30	TO-15 + Naphthalene	-30 <sup>+</sup> 5.004g
02A	12007 <i>upwind</i>	4/5/07 15:30	TO-15 + Naphthalene	-30 <sup>+</sup> 5.004g

CUSTODY SEAL INTACT?  
 IN NONE TEMP. A/C

Requisitioned By: (Signature) *[Signature]* Date/Time: 4/5/07 15:30 pm  
 Received By: (Signature) *[Signature]* Date/Time: 4/6/07-0900  
 Requisitioned By: (Signature) \_\_\_\_\_ Date/Time: \_\_\_\_\_  
 Received By: (Signature) \_\_\_\_\_ Date/Time: \_\_\_\_\_

Notes: used flow controllers included  
 Initial and final can pressures in inches Hg: \_\_\_\_\_  
 Send Data Pack to Lisa McDonough and EDD to  
 datagroup@geiconsultants.com

Lab: \_\_\_\_\_ Shipper Name: \_\_\_\_\_ Air Bill #: \_\_\_\_\_  
 Use: \_\_\_\_\_ FedEx: 8595 2431 9034  
 Only: \_\_\_\_\_  
 Opened By: *TR* Temp. (S): *NA*  
 Condition: *good*  
 Quality Seal Intact:  Yes  No  None  
 Work Order #: 070414



AN ENVIRONMENTAL ANALYTICAL LABORATORY

## SAMPLE RECEIPT SUMMARY

### WORKORDER 0704143

**Client**

Ms. Sarah Aldridge  
GEI Consultants, Inc.  
455 Winding Brook Dr. Suite 201  
Glastonbury, CT 06033

**Phone**

860-368-5300

**Fax**

860-368-5307

**Date Promised:** 04/20/07

**Date Completed:** 4/19/07

**Date Received:** 4/6/07

**PO#:**

**Project#:** 061140-8-1703 Bayshore OU1 Southern cell  
Air Monitor

**Total \$:** \$ 624.00

**Logged By:** ANC

**Sales Rep:** ANS

<u>Fraction</u>	<u>Sample #</u>	<u>Analysis</u>	<u>Collected</u>	<u>Receipt Vac./Pres.</u>	<u>Amount\$</u>
01A	000001016 Downwind	Modified TO-15	4/5/2007	5.0 "Hg	\$225.00
02A	12007 Upwind	Modified TO-15	4/5/2007	5.0 "Hg	\$225.00
03A	Lab Blank	Modified TO-15	NA	NA	\$0.00
04A	CCV	Modified TO-15	NA	NA	\$0.00
05A	LCS	Modified TO-15	NA	NA	\$0.00
Misc. Charges 6 Liter Summa Canister (2) @ \$50.00 each.					\$100.00
Blue Body Flow Controller (2) @ \$35.00 each.					\$70.00
Fuel Surcharge (2) @ \$2.00 each.					\$4.00

**Note:** Samples received after 3 P.M. PST are considered to be received on the following work day.  
Atlas Project Name/Profile#: Keyspan -Bayshore Southern Cell IRM/9699

**BILL TO:** Ms. Sarah Aldridge  
GEI Consultants, Inc.  
455 Winding Brook Dr. Suite 201  
Glastonbury, CT 06033

Analysis Code: TO-14A

Reporting Method: Modified TO-15 + Naph

180 BLUE RAVINE ROAD, SUITE B FOLSOM, CA - 95630  
(916) 985-1000 . (800) 985-5955 . FAX (916) 985-1020

# Sample Discrepancy Report

If Section III or IV is filled out CSR must be notified within 24 hrs of initiation

Initiated By: ANC

Date: 4/7/07 Given To: \_\_\_\_\_

File to folder

Sections I – II/III/IV must be filled out by person initiating this Sample Discrepancy Report

I. Workorder(s) affected: 0704143  
Sample(s) affected: 01A-02A (All)

II. Sample Receipt Discrepancies (Document on Cover Page of Sample Receipt Confirmation and in Receiving Notes of Lab Narrative)

- |   |   |
|---|---|
| <input type="checkbox"/> COC improperly relinquished / received.  | <input type="checkbox"/> Flow controller used - canister samples received at ambient or under pressure.               |
| <input type="checkbox"/> COC was not filled out in ink.   | <input type="checkbox"/> No brass cap on canister ( <i>do not narrate</i> ).  |
| <input checked="" type="checkbox"/> Sample tags / labels do not match the COC.  | <input type="checkbox"/> VOA vial for RSK-175 analysis received with headspace bubble <5mm ( <i>do not narrate</i> ). |
| <input type="checkbox"/> Samples received at wrong temperature ( $\neq 4\pm 2^\circ\text{C}$ ); ice / blue ice (circle one) was present. A temp. blank was / was <i>not</i> present (circle one). | <input type="checkbox"/> Other (describe below).  |
| <input type="checkbox"/> Sample container (Tube/VOA vial) was received broken, <i>however</i> sample was intact.  |   |

Describe the Discrepancy: No IDs on tags. Matching by can #.

Initials: \_\_\_\_\_ Date: \_\_\_\_\_

III. Sample Receipt Discrepancies requiring CSR notification (document on Cover Page of Sample Receipt Confirmation and in Receiving Notes of Lab Narrative)

- |   |  |
|---|--|
| <input type="checkbox"/> COC was not received with samples.   | <input type="checkbox"/> Canister leaked to ambient during pressurization.   |
| <input type="checkbox"/> Analysis method(s) is not specified / incorrectly specified (circle one) on the COC.           | <input type="checkbox"/> Tedlar bag / canister received emitting a strong odor; sample can / cannot (circle one) be analyzed.  |
| <input type="checkbox"/> Number of samples on the COC does not match the number of samples that were received.          | <input type="checkbox"/> Canister sample received with a vacuum difference >7.0"Hg between the receipt vac. and the final vac. reported on the COC, indicating loss of vacuum. |
| <input type="checkbox"/> Samples were received expired.   | <input type="checkbox"/> Canister sample received at >15"Hg ( <i>not</i> identified as a Trip/Field Blank).  |
| <input type="checkbox"/> Sampling date / time is not documented for <u>some</u> / <u>any</u> samples (circle one).      | <input type="checkbox"/> Trip Blank received at low vacuum (< 25"Hg).  |
| <input type="checkbox"/> Sample received with discernable volume of H <sub>2</sub> O in the Tedlar Bag.                 | <input type="checkbox"/> Tedlar Bag for Sulfur analysis has metal fitting.   |
| <input type="checkbox"/> Sample container (Tube/VOA vial/DNPH Bottle, etc.) was received broken / leaking (circle one). | <input type="checkbox"/> Incorrect sampling media / container for analysis requested.  |
| <input type="checkbox"/> VOA vial for RSK-175 analysis received with headspace bubble >5mm.                             | <input type="checkbox"/> Custody Seal on the outside of the container was broken / improperly placed (circle one).   |
| <input type="checkbox"/> Samples for RSK-175 CO <sub>2</sub> analysis received preserved with HCl.                      | <input type="checkbox"/> Other (describe below).   |
| <input type="checkbox"/> Tedlar Bag received leaking / flat (circle one). Sample can / cannot (circle one) be analyzed. |  |

Describe the Discrepancy: \_\_\_\_\_


Initials: \_\_\_\_\_ Date: \_\_\_\_\_

## **Other Records**

Team A

Date / Initial	4-19-07 CP
Poor Integration	
Split Peak	
Peak Tailing	
Background Subtraction	X
Zoom In	
Missed Peak	
Merged Peaks	

-01A 0704143



## DILUTION FACTORS

$$\text{Dilution Factor} = \frac{\text{Final Pressure}}{\text{Initial Vacuum}} = \frac{14.7 \text{ psi} + \text{Final Pressure (psi)}}{14.7 \text{ psi} - [(\text{Initial Pressure ("Hg)}) (14.7 \text{ psi} / 30 \text{ "Hg})]}$$

$$\text{Dilution Factor} = \frac{\text{Final Pressure}}{\text{Initial Pressure}} = \frac{14.7 \text{ psi} + \text{Final Pressure (psi)}}{14.7 \text{ psi} + \text{Initial Pressure (psi)}}$$

Initial Vacuum ("Hg)	5 psi Final Press. Dil. Factor	10 psi Final Press. Dil. Factor	15 psi Final Press. Dil. Factor
0.0	1.34	1.68	2.02
0.5	1.36	1.71	2.05
1.0	1.39	1.74	2.09
1.5	1.41	1.77	2.13
2.0	1.44	1.80	2.16
2.5	1.46	1.83	2.20
3.0	1.49	1.87	2.24
3.5	1.52	1.90	2.29
4.0	1.55	1.94	2.33
4.5	1.58	1.98	2.38
5.0	1.61	2.02	2.42
5.5	1.64	2.06	2.47
6.0	1.68	2.10	2.53
6.5	1.71	2.15	2.58
7.0	1.75	2.19	2.64
7.5	1.79	2.24	2.69
8.0	1.83	2.29	2.76
8.5	1.87	2.34	2.82
9.0	1.91	2.40	2.89
9.5	1.96	2.46	2.96
10.0	2.01	2.52	3.03
10.5	2.06	2.59	3.11
11.0	2.12	2.65	3.19
11.5	2.17	2.72	3.28
12.0	2.23	2.80	3.37
12.5	2.30	2.88	3.46
13.0	2.36	2.97	3.57
13.5	2.44	3.06	3.67
14.0	2.51	3.15	3.79
14.5	2.59	3.25	3.91
15.0	2.68	3.36	4.04
15.5	2.77	3.48	4.18
16.0	2.87	3.60	4.33
16.5	2.98	3.73	4.49
17.0	3.09	3.88	4.66
17.5	3.22	4.03	4.85
18.0	3.35	4.20	5.05
18.5	3.50	4.38	5.27
19.0	3.65	4.58	5.51
19.5	3.83	4.80	5.77
20.0	4.02	5.04	6.06
20.5	4.23	5.31	6.38

Initial Vacuum ("Hg)	5 psi Final Press. Dil. Factor	10 psi Final Press. Dil. Factor	15 psi Final Press. Dil. Factor
21.0	4.47	5.60	6.73
21.5	4.73	5.93	7.13
22.0	5.03	6.30	7.58
22.5	5.36	6.72	8.08
23.0	5.74	7.20	8.66
23.5	6.19	7.76	9.32
24.0	6.70	8.40	10.10
24.5	7.31	9.17	11.02
25.0	8.04	10.08	12.12
25.5	8.93	11.20	13.47
26.0	10.05	12.60	15.15
26.5	11.49	14.40	17.32
27.0	13.40	16.80	20.20
27.5	16.08	20.16	24.24
28.0	20.10	25.20	30.31
28.5	26.80	33.61	40.41
29.0	40.20	50.41	60.61

Initial Pressure (psi)	5 psi Final Press. Dil. Factor	10 psi Final Press. Dil. Factor	15 psi Final Press. Dil. Factor
0.0	1.34	1.68	2.02
0.2	1.32	1.66	1.99
0.4	1.30	1.64	1.97
0.6	1.29	1.61	1.94
0.8	1.27	1.59	1.92
1.0	1.25	1.57	1.89
1.2	1.24	1.55	1.87
1.4	1.22	1.53	1.84
1.6	1.21	1.52	1.82
1.8	1.19	1.50	1.80
2.0	1.18	1.48	1.78
2.2	1.17	1.46	1.76
2.4	1.15	1.44	1.74
2.6	1.14	1.43	1.72
2.8	1.13	1.41	1.70
3.0	1.11	1.40	1.68
3.2	1.10	1.38	1.66
3.4	1.09	1.36	1.64
3.6	1.08	1.35	1.62
3.8	1.06	1.34	1.61
4.0	1.05	1.32	1.59



## DILUTION FACTORS

$$\text{Dilution Factor} = \frac{\text{Final Pressure}}{\text{Initial Pressure}} = \frac{14.7 \text{ psi} + \text{Final Pressure (psi)}}{14.7 \text{ psi} + \text{Initial Pressure (psi)}}$$

Initial Pressure (psi)	5 psi Final Press. Dil. Factor	10 psi Final Press. Dil. Factor	15 psi Final Press. Dil. Factor
0.0	1.34	1.68	2.02
0.2	1.32	1.66	1.99
0.4	1.30	1.64	1.97
0.6	1.29	1.61	1.94
0.8	1.27	1.59	1.92
1.0	1.25	1.57	1.89
1.2	1.24	1.55	1.87
1.4	1.22	1.53	1.84
1.6	1.21	1.52	1.82
1.8	1.19	1.50	1.80
2.0	1.18	1.48	1.78
2.2	1.17	1.46	1.76
2.4	1.15	1.44	1.74
2.6	1.14	1.43	1.72
2.8	1.13	1.41	1.70
3.0	1.11	1.40	1.68
3.2	1.10	1.38	1.66
3.4	1.09	1.36	1.64
3.6	1.08	1.35	1.62
3.8	1.06	1.34	1.61
4.0	1.05	1.32	1.59
4.2	1.04	1.31	1.57
4.4	1.03	1.29	1.55
4.6	1.02	1.28	1.54
4.8	1.01	1.27	1.52
5.0	1.00	1.25	1.51
5.2	NA	1.24	1.49
5.4	NA	1.23	1.48
5.6	NA	1.22	1.46
5.8	NA	1.20	1.45
6.0	NA	1.19	1.43
6.2	NA	1.18	1.42
6.4	NA	1.17	1.41
6.6	NA	1.16	1.39
6.8	NA	1.15	1.38
7.0	NA	1.14	1.37
7.2	NA	1.13	1.36
7.4	NA	1.12	1.34

Initial Pressure (psi)	5 psi Final Press. Dil. Factor	10 psi Final Press. Dil. Factor	15 psi Final Press. Dil. Factor
7.6	NA	1.11	1.33
7.8	NA	1.10	1.32
8.0	NA	1.09	1.31
8.2	NA	1.08	1.30
8.4	NA	1.07	1.29
8.6	NA	1.06	1.27
8.8	NA	1.05	1.26
9.0	NA	1.04	1.25
9.2	NA	1.03	1.24
9.4	NA	1.02	1.23
9.6	NA	1.02	1.22
9.8	NA	1.01	1.21
10.0	NA	1.00	1.20
10.2	NA	NA	1.19
10.4	NA	NA	1.18
10.6	NA	NA	1.17
10.8	NA	NA	1.16
11.0	NA	NA	1.16
11.2	NA	NA	1.15
11.4	NA	NA	1.14
11.6	NA	NA	1.13
11.8	NA	NA	1.12
12.0	NA	NA	1.11
12.2	NA	NA	1.10
12.4	NA	NA	1.10
12.6	NA	NA	1.09
12.8	NA	NA	1.08
13.0	NA	NA	1.07
13.2	NA	NA	1.06
13.4	NA	NA	1.06
13.6	NA	NA	1.05
13.8	NA	NA	1.04
14.0	NA	NA	1.03
14.2	NA	NA	1.03
14.4	NA	NA	1.02
14.6	NA	NA	1.01
14.8	NA	NA	1.01

# Compound Listing

## Modified TO-15 + Naph

CAS Number	Compound	Detection Limit	Type
		ppbv	
75-71-8	Freon 12	0.50	
76-14-2	Freon 114	0.50	
108-38-3	m,p-Xylene	0.50	
95-47-6	o-Xylene	0.50	
100-42-5	Styrene	0.50	
79-34-5	1,1,2,2-Tetrachloroethane	0.50	
108-67-8	1,3,5-Trimethylbenzene	0.50	
95-63-6	1,2,4-Trimethylbenzene	0.50	
541-73-1	1,3-Dichlorobenzene	0.50	
106-46-7	1,4-Dichlorobenzene	0.50	
100-44-7	alpha-Chlorotoluene	0.50	
95-50-1	1,2-Dichlorobenzene	0.50	
106-99-0	1,3-Butadiene	0.50	
110-54-3	Hexane	0.50	
110-82-7	Cyclohexane	0.50	
142-82-5	Heptane	0.50	
75-27-4	Bromodichloromethane	0.50	
124-48-1	Dibromochloromethane	0.50	
98-82-8	Cumene	0.50	
103-65-1	Propylbenzene	0.50	
74-87-3	Chloromethane	2.0	
120-82-1	1,2,4-Trichlorobenzene	2.0	
87-68-3	Hexachlorobutadiene	2.0	
67-64-1	Acetone	2.0	
75-15-0	Carbon Disulfide	0.50	
67-63-0	2-Propanol	2.0	
156-60-5	trans-1,2-Dichloroethene	0.50	
78-93-3	2-Butanone (Methyl Ethyl Ketone)	0.50	
109-99-9	Tetrahydrofuran	0.50	
123-91-1	1,4-Dioxane	2.0	
108-10-1	4-Methyl-2-pentanone	0.50	
591-78-6	2-Hexanone	2.0	
75-25-2	Bromoform	0.50	
622-96-8	4-Ethyltoluene	0.50	
64-17-5	Ethanol	2.0	
1634-04-4	Methyl tert-butyl ether	0.50	
91-20-3	Naphthalene	2.0	
107-05-1	3-Chloropropene	2.0	
540-84-1	2,2,4-Trimethylpentane	0.50	
2037-26-5	Toluene-d8		
17060-07-0	1,2-Dichloroethane-d4		
460-00-4	4-Bromofluorobenzene		
75-01-4	Vinyl Chloride	0.50	
74-83-9	Bromomethane	0.50	
75-00-3	Chloroethane	0.50	
75-69-4	Freon 11	0.50	

# Compound Listing

## Modified TO-15 + Naph

CAS Number	Compound	Detection Limit	Type
		ppbv	
75-35-4	1,1-Dichloroethene	0.50	
76-13-1	Freon 113	0.50	
75-09-2	Methylene Chloride	0.50	
75-34-3	1,1-Dichloroethane	0.50	
156-59-2	cis-1,2-Dichloroethene	0.50	
67-66-3	Chloroform	0.50	
71-55-6	1,1,1-Trichloroethane	0.50	
56-23-5	Carbon Tetrachloride	0.50	
71-43-2	Benzene	0.50	
107-06-2	1,2-Dichloroethane	0.50	
79-01-6	Trichloroethene	0.50	
78-87-5	1,2-Dichloropropane	0.50	
10061-01-5	cis-1,3-Dichloropropene	0.50	
108-88-3	Toluene	0.50	
10061-02-6	trans-1,3-Dichloropropene	0.50	
79-00-5	1,1,2-Trichloroethane	0.50	
127-18-4	Tetrachloroethene	0.50	
106-93-4	1,2-Dibromoethane (EDB)	0.50	
108-90-7	Chlorobenzene	0.50	
100-41-4	Ethyl Benzene	0.50	

DATA REVIEW CHECKLIST

Work Order #:

0704143

A R T M Q

- Analysis/Reporting vs. Project Profile/SOP requirements checked (i.e. 100% Dups, J-Flag to MDL, etc)
The final report has the correct reporting list, special units, and header info.
Lab Narrative is correct (proper method & description/Receiving & Analytical notes correct)
Corrective Action issued - #
Unusual circumstances have been documented in the notes section below

LUMEN validation report present and initialed

CIRCLE (YES) (NO)

- Lab Blank, CCV, LCS and DUP met QC criteria
Hold time is met for all samples
Appropriate data qualifier flags are applied
Manual integrations for samples and QC are properly documented
Samples analyzed within the project or method specific clock
Retention times have been verified
Appropriate ICAL(s) included
At least one result per sample is verified against the target quant sheets/raw data

- Dilution factor correctly calculated (sample load volume, syringe and bag dilutions, can pressurization(s))
Correct amount of sample analyzed (i.e. sample not over-diluted)
Spectra verified - documentation of spectral defense included (Section 5A of eCVP pkg)
TICs resemble reference spectra
TICs between duplicate samples are consistent
Checked samples for trends (i.e. Influent>Effluent, Landfill or Ambient etc)
Special units for all samples in the final report are correctly calculated
Manually entered results checked (i.e. special CCV compounds)
TPH/NMOC (verify calculations and correct reference compound used)
Chain of Custody scanned correctly
Verify sample id's vs. chain of custody
Samples pressurized w/ appropriate gas (N2 or He)
Final pressure consistent with canister size (6L vs. 1L)
Verify receipt pressures against logbook and Target
Verify canister ID #'s
Extra printed copies are provided per client profile
Final invoice amount correct (adjusted for TAT, Penalties, Re-issue Charges etc.)
Client LUMEN report reviewed for accuracy and completeness

Notes: (to include: noting samples with QA/QC problems, Blanks with positive hits, narratives, etc.)

A/R:
M/Q:

A

R/T

M

Q

(Analytical Review/Date)

(Reporting Review/Date)

(Management Review/Date)

(QA Review/Date)

4/13/07

R: Taylor 4-19-07

M: 4/19/07

T:

**Not Applicable**